A-A-53586
29 July 1988

# COMMERCIAL ITEM DESCRIPTION

## TOURNIQUET, NON-PNEUMATIC

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers an adult non-pneumatic tourniquet suitable for use in taking blood during phlebotomies.

Salient characteristics:

The adult non pneumatic tourniquet shall consist of a gray rubber (natural or synthetic or combination) strip 14 ± 1/4 inches long and 1 ± 1/8 inch wide.

Each end of the tourniquet strip shall be rendered inelastic by the inclusion of a cloth scrim within the rubber. The scrim shall extend the full width of the strip. The elastic center portion of the strip shall be 1.6 ± 0.4 mm thick. Each end of the tourniquet shall contain a reservoir of rubber (i.e., thickened portion between the elastic portion and the inelastic scrim portion. The reservoir shall extend the full width of the strip and shall be 3 ± 0.4 mm thick. The reservoir shall be approximately 1 inch long.

One end of the strip shall have sewn to it a length of pile fastener material 3-1/4 ± 1/4 inch long and 15/16 ± 1/16 inch wide. On the opposite face, same end, a label shall be sewn. The label shall indicate size 'adult' and the name of manufacturer. The other end of the strip shall have sewn to it a length of hook material 4-1/2 ± 1/4 inch long and 15/16 ± 1/16 inch wide. The hook material shall be sewn to the opposite face and opposite end of the pile material. The hook and pile fastener shall have a minimum shear strength of 7.5 pounds.

There shall be 6 to 10 stitches per inch in the sewing operation. The stitch type shall be a 301 lock-stitch. The thread shall be the same as used for standard commercial practice.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Workmanship. The non-pneumatic tourniquets shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One tourniquet, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one tourniquet as specified in this description.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Defects.

Any cut, hole, or tear.

Any pinhole.

Reservoir missing or not specified.

Any blister, tunnel or delamination of coating.

Label, hook and pile fastener, or scrim missing or otherwise not usable.

Identification marking not complete, permanent, and legible.

Label, hook and pile fastener not properly sewn or not sewn with required stitches per inch.

Wrong color.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 percent defective and an inspection level of S-1. The unit of product for test purposes shall be one non-pneumatic tourniquet

Tests

Performance of tourniquet. The adult non-pneumatic tourniquet shall withstand at least 1800 full stretch-relax cycles without breaking or showing any visible damage when tested. The testing device shall be such that at one extreme of its motion the tourniquet shall be fully relaxed, and at the other extreme, the length of the elastic center portion shall be stretched to 300 percent of its relaxed length.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirement of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirement of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Number is covered by this document:

6515-01-146-7794

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and or order.

MILITARY INTERESTS.

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPIAS

Project No. 6515-4256

FSC 5944

Location. CID/TOURNIQUET/H15-32

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (DO NOT STAPLE), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed. If you would like a specific reply, you must provide your mailing address.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEPARTMENT OF DEFENSE

Commander
Defense Personnel Support Center
ATTN: DPSC-RCMB
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

*See Instructions – Reverse Side*

| 1. DOCUMENT NUMBER | 2. DOCUMENT TITLE |
|---|---|
| A-A-53586 | TOURNIQUET, NON-PNEUMATIC |

| 3a. NAME OF SUBMITTING ORGANIZATION | 4. TYPE OF ORGANIZATION (Mark one) |
|---|---|
| | ☐ VENDOR |
| | ☐ USER |
| b. ADDRESS (Street, City, State, ZIP Code) | ☐ MANUFACTURER |
| | ☐ OTHER (Specify) |

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

| 7a. NAME OF SUBMITTER (Last, First, MI) – Optional | b. WORK TELEPHONE NUMBER (Include Area Code) – Optional |
|---|---|
| c. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional | 8. DATE OF SUBMISSION (YYMMDD) |

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.