A-A-53587
29 July 1988

COMMERCIAL ITEM DESCRIPTION

WRAPPER, STERILIZATION

The General Services Administration has authorized the use of this Commercial Item Description in preference to Federal Specification DDD-W-631.

This Commercial Item Description covers a cotton, double thickness sterilization wrapper suitable for autoclaving and storing surgical supplies.

Salient characteristics:

The sterilization wrapper shall be a cotton, double thickness sterilization wrapper suitable for autoclaving and storing surgical supplies. The wrapper shall be furnished in the following sizes:

Size 1 - 12 inches square + 1 inch, - 1/2 inch.
Size 2 - 18 inches square ± 1 inch.
Size 3 - 24 inches square ± 1 inch.
Size 4 - 36 inches square ± 1 inch.
Size 5 - 54 inches square ± 2 inches.

The wrapper shall be fabricated from a uniformly dyed, colorfast, unshrunk, cotton cloth, commercial designation No. 140 (muslin). The color of the cloth shall approximate the grayish green 180 standard shade (misty green). The wrapper shall show "good" color-fastness to laundering.

Minimum weight of the wrapper shall be 4.2 ounces per square yard. Total yarns per inch shall be 140 minimum. Number of ends per inch shall be 74 minimum. Number of picks per inch shall be 66 minimum. Breaking strength of the cloth shall be 60 pounds minimum. The weave of the cloth shall be plain.

The thread shall be a soft finish cotton thread, ticket No. 30, 3 ply. It shall have a length per pound between 6,441 yards and 8,400 yards, and a minimum breaking strength of 3.0 pounds.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

The wrapper shall consist of two plies of the wrapper cloth sewn together along the sides and ends. The wrapper shall not have diagonal or center stitching.

All wrapper sewing shall be accomplished with stitch type 301 with ten to twelve stitches per inch. All top stitching shall be 1/16 to 1/8 inch from the edges of the wrapper. The seam allowance for all wrappers shall be a minimum of 1/4 inch. The seam allowance shall be maintained so that no runoff or open seam shall occur. All the seams shall start and finish uniformly. No seam shall be twisted, puckered, or pleated. All the seam edges shall be properly forced out and shall not contain any folds. All the thread ends shall be trimmed to a maximum length of 1/4 inch. No cut or raw edge shall be visible on the exterior of the wrapper.

Each row of stitching shall be straight and parallel to the seam edge. The straightness of the stitching, in any row, shall be maintained within a tolerance of plus or minus 1/16 inch. The same tolerance for spacing and parallelism shall be applicable for the double rows of stitches. The ends of the stitching that are not caught in other seams or stitching shall be backstitched, by overlapping on themselves, a minimum of 1/2 inch. Each thread break, skip, and run-off shall be overstitched not less than 1/2 inch. The thread tension shall be maintained so that there shall not be any loose or tight stitching and the lock shall be embedded in the materials sewn together.

The wrapper shall be fabricated from one piece of cloth folded in half or from two pieces of cloth. The wrapper shall be sewn together along the sides and ends with a single row of stitching.

The wrapper shall not contain any spot, stain, hole, abraded area, tear, cut, mend, patch, scorch, needle chew, dust particles, lint, thread and cloth scraps, or any other surface contaminant.

Workmanship. The sterilization wrapper shall be free from defects which detract from its appearance or impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing six wrappers, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any or the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective and an inspection level of II. The unit of product for examination purposes shall be one sterilization wrapper as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Defects.

Any component not as specified or any defect of component or defect of assembly.

Any component, component part, or required operation omitted, or any operation improperly performed.

Ends of stitching, that are not caught in other seams or stitching, not securely backstitched.

Thread breaks are not securely overstitched.

Open seams more than 1/2 inch.

Raw edge more than 1/2 inch.

Seam or stitch not as specified.

Any loose stitching.

Any tight stitching resulting in breakage of the stitches when normal pull is applied.

Two more or two less stitches than specified.

Any hole, scissors or knife cut, tear, mend, burn, or weakening defect such as smash, multiple float, loose slub, needle chew, or abraded area.

Any portion of the wrapper stiffened or scorched by any process of manufacturing.

Any missing or broken yarn, thin place, misweave, woven-in waste, loose yarn, snagged thread, area of no dye penetration, or dye streak on the exterior of the wrapper.

Color not as specified.

Any unsightly stain, spot, or stain.

Any stitching irregular, uneven or not gaged as specified, or not uniform.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one sterilization wrapper.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

6530-01-246-1935 (size 1)
6530-01-246-4812 (size 2)
6530-01-244-2776 (size 3)
6530-01-244-9946 (size 4)
6530-01-246-0156 (size 5)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-2006

PGC 70026

Location:CID/WRAPPER/STERILIZATION /H13-32

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (DO NOT STAPLE), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53587	WRAPPER, STERILIZATION

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
	☐ USER
b ADDRESS (Street, City, State, ZIP Code)	☐ MANUFACTURER
	☐ OTHER (Specify)

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (Last, First, MI) – Optional	b WORK TELEPHONE NUMBER (Include Area Code) – Optional
c MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8 DATE OF SUBMISSION (YYMMDD)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.