A-A-53600
29 July 1988

COMMERCIAL ITEM DESCRIPTION

TRAYS, IMPRESSION, DENTAL (Non-water Cooled)

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification GG-T-00626A (DSA-DM).

Salient characteristics:

This commercial item description covers dental impression trays suitable for taking full impressions with dental irreversible hydrocolloid materials and rubber base materials. The trays shall be furnished for upper or lower, dentulous or edentulous impressions, and in the size as listed in the "NOTE" section for the applicable National Stock Number and Item Identification.

Material.

Tray body and handle. The body and handle of the tray shall be fabricated of brass conforming to Copper Alloy No. 260 as specified in ASTM B36 for composition only. The body shall be annealed temper and the handle shall be half-hard temper.

Rim wire. The peripheral rim and palatal area wire shall be soft, solid, copper wire.

Solder. The solder used for attaching the handle and the wire to the tray body shall be 95 percent tin and 5 percent antimony as specified in ASTM B32 for composition only.

Style, design and dimensions. The impression trays shall be suitable for use in taking full impressions with dental irreversible hydrocolloid materials and rubber base materials. The style, design and dimensions shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 24302. The trays shall be non-water cooled trays of the rim lock design with a solid surface. The rim lock shall serve to lock the impression material in the tray. The upper trays shall have a wire rim in the palatal area to reinforce the retention of the impression material.

Construction. The handle of the tray shall be secured to the body of the tray by soldering. The wire rim around the periphery of the body of the tray and in the palatal area shall also be secured by soldering.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Finish. Trays shall be nickel or chromium over nickel plated. Trays shall have a smooth satin or bright polished finish. Finish on edges and surfaces shall be uniform and free of burrs, sharp edges, crevices, rough areas and cracks. Wire rims shall be smoothly and uniformly contoured. Soldered joints shall be continuous and free from projections, fractures and voids. Plating shall be uniform in appearance. Plating shall be free from blisters, visible porosity, peeling, flaking, or other defects. Soldered connections shall be plated over.

Identification marking. Trays made in U.S.A. shall be marked with name or registered trademark of the contractor. Trays of foreign (other than U.S.) manufacture shall be marked with the name of origin, together with the name or registered trademark of the contractor. All marking shall be on the top face of the handle and shall be capable of being read when the tray is hanging on the trayboard. All marking shall be permanent and legible.

Workmanship. The trays shall be free of defects which detract from their appearance or impair their serviceability and intended use. The trays shall be uniform in quality and shall be smooth and free from burrs, seams, cracks, and free from blisters. Solder joints shall be uniform, continuous and free of flux residue or foreign material.

Unit. Each (EA). One impression tray, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification, where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II, except for dimensions the inspection level shall be S-1. The unit of product for examination purposes shall be one tray. Defects shall be those listed below.

Defects.

Style and design not as specified.
Tray not constructed as specified.
Edges and surfaces not free of burrs, sharp edges, or cracks.
Plating not free from blisters or peeling.
Soldered connections not plated over.
Wire rim not securely soldered to tray body.
Handle not securely soldered to tray body.
Dimensions not as specified.
Identification marking not present, not properly located, illegible, or incomplete.
Letter or letters designating size missing or incorrect for the applicable size.
Trays not free from dirt, polishing compound, or foreign matter.
Finish not smooth or brightly polished.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Number	Item Identification
6520-00-563-1650	IMPRESSION TRAY, DENTAL, Dentulous, Lower, Small
6520-00-563-0650	IMPRESSION TRAY, DENTAL, Dentulous, Lower, Medium
6520-00-563-1150	IMPRESSION TRAY, DENTAL, Dentulous, Lower, Large
6520-00-563-0150	IMPRESSION TRAY, DENTAL, Dentulous, Lower, Extra Large

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

National Stock Number	Item Identification
6520-00-042-8402	IMPRESSION TRAY, DENTAL, Dentulous, Lower, Extra-Extra Large
6520-00-563-3650	IMPRESSION TRAY, DENTAL, Dentulous, Upper, Small
6520-00-563-2650	IMPRESSION TRAY, DENTAL, Dentulous, Upper, Medium
6520-00-563-3150	IMPRESSION TRAY, DENTAL, Dentulous, Upper, Large
6520-00-563-2150	IMPRESSION TRAY, DENTAL, Dentulous, Upper, Extra Large
6520-00-042-8396	IMPRESSION TRAY, DENTAL, Dentulous, Upper, Extra-Extra Large
6520-00-165-6569	IMPRESSION TRAY, DENTAL, Edentulous, Lower, Small
6520-00-165-6568	IMPRESSION TRAY, DENTAL, Edentulous, Lower, Medium
6520-00-165-6567	IMPRESSION TRAY, DENTAL, Edentulous, Lower, Medium Large
6520-00-165-6566	IMPRESSION TRAY, DENTAL, Edentulous, Lower, Large
6520-00-165-6565	IMPRESSION TRAY, DENTAL, Edentulous, Upper, Small
6520-00-165-6564	IMPRESSION TRAY, DENTAL, Edentulous, Upper, Medium
6520-00-165-6563	IMPRESSION TRAY, DENTAL, Edentulous, Upper, Medium Large
6520-00-165-6562	IMPRESSION TRAY, DENTAL, Edentulous, Upper, Large

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

A-A-53600

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

05500

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

PROJECT NO. 6520-2078

Location: NCR Disk#6

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in ...itting comments and suggestions for improvements. All users of military standardization documents are invited to provide ...gestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATE

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53600	TRAYS, IMPRESSION, DENTAL (Non-water Cooled)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street City State ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:22Z
Check the source to verify that this is the current version before use.