MILITARY SPECIFICATION

CABINET, CASSETTE TRANSFER, WALL MOUNTED

1. SCOPE

1.1 This specification covers one type of lead lined, two compartment, lightproof, manual cassette transfer cabinet, designed for mounting into the wall. The unit is used for passing loaded exposed and unexposed medical x-ray film cassettes into and out of the darkroom in military medical x-ray facilities.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-L-201	Lead Sheet.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.

Military

MIL-L-10547	Liners, Case, and Sheet Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6525

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:29Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Steel sheet. The internal and external shells of the transfer cabinet shall be constructed of cold rolled steel not less than 20 gage (.035 inch nominal) in wall thickness. Commercial tolerances on the gage of material used shall apply.

3.1.2 Lead. Lead sheet used for x-ray protection in the construction of the transfer cabinet shall be grade C or D in accordance with QQ-L-201. The nominal thickness of the lead sheet shall be 1/16 inch, equivalent to a unit weight of 4 pounds per square foot.

3.1.3 Hardware. Hardware shall be of the highest quality throughout, suitable in all respects for the purpose intended.

3.2 Style, design and construction. The cabinet shall have two light-proof compartments and shall be designed for mounting into a wall. The unit shall have two separate lead lined doors at each end, each fitted with a metal or plastic knob to facilitate opening and closing. Corresponding doors at each end shall be clearly stencilled "EXPOSED" and "UNEXPOSED" for segregation and temporary storage of cassettes in the appropriate compartment. The door markings shall be stencilled with letters not less than 5/8 inch in height. The cabinet shall include a special interlocking device to prevent the opening of both doors in either compartment at the same time. Cabinet shall be constructed from materials specified in 3.1. The top, bottom, sides and ends (including doors) shall be lined with lead. Each door shall be mounted with not less than two hinges, spot welded or fastened by machine screws tightly secured with nuts. The method of fastening the doors shall not reduce the equivalent lead protection offered by the lining. Hardware for the door locking devices shall be similarly mounted to retain equivalent lead protection. The assembled doors and door locking devices shall be securely mounted and operate smoothly. The cabinet shall be furnished with two mounting flanges (one of which may be fixed to the cabinet body) not less than 1-1/2 inches wide with nominal 1/16 inch lead lining. The bottom of each compartment

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:29Z
Check the source to verify that this is the current version before use.

shall be fitted with matting or other suitable material to retain stored cassettes in a stable position, prevent slippage and minimize abrading cassette surfaces when effecting transfers in each of the compartments. Each compartment shall hold at least ten 14 by 17 inch cassettes. The overall dimensions of the cabinet, excluding mounting flanges, shall be 22-1/2 ± 1/2 inches wide, 22-1/2 ± 1/2 inches high and 19 ± 1/2 inches deep (front to back).

3.3 Finish. All hardware parts not constructed of corrosion-resisting base metal shall be protected by chrome-plating over a nickel underlay. As an alternate, aluminum alloy parts may be treated for corrosion by anodizing. All internal and external cabinet surfaces shall be degreased, cleaned and treated to improve the adhesion of the finish coats. The interior compartment surfaces shall be finished in flat black. Exterior surfaces shall be finished by the application of two coats of enamel in the color normally furnished by the manufacturer. Stencilled door markings shall be readily visible and legible. The resultant internal and external finishes shall be hard, smooth and highly resistant to abrasion, chipping or cracking.

3.4 Identification marking. Each cabinet shall have a nameplate permanently attached bearing the following information:

Federal stock number
Name or registered trademark of the manufacturer.
Contract or purchase order number.
U.S.

3.5 Service data. The contractor shall enclose, with each unit, two copies each of one or more booklets which contain complete instructions for installation, operation, maintenance, and a list of component parts. Each part of the unit other than common fasteners, etc., shall be identified in the service data and parts list by the manufacturer's part number. In addition, the contractor shall mail 15 copies of the service data to the contracting officer for distribution.

3.6 Workmanship. Workmanship shall be first class throughout. Cabinets shall be free from defects which detract from their appearance or may impair their serviceability. All sharp edges shall be removed from the cabinet. Lockwashers or locknuts shall be used wherever machine screws or bolts are utilized for attachment.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:29Z
Check the source to verify that this is the current version before use.

3.6.1 Welding. All welding shall be done with appropriate equipment properly regulated and operated by thoroughly trained and experienced welders, highly skilled in the operations of welding. Seam and fillet welds shall be continuous, even, and smooth. Spot welds shall be accomplished to attain sound welds of strength and penetration sufficient to tear the spot out of the fused joint when test coupons are pulled apart. All exposed seam and fillet welds shall be ground smooth to eliminate traces of weld marks, and polished to blend with adjacent surfaces. All exposed spot welds shall be ground smooth, or filled and sanded smooth so that they will not be visible after the enamel finish is applied.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105.

4.3 Examination. The cabinets shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:29Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) cabinet, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit preparation. All surfaces of the unit shall be suitably protected by wrapping with kraft paper in such a manner as to prevent marring or scratching. Wrapping shall be adequately secured.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Each unit protected as specified in 5.1.2.1 shall be overpacked in an exterior container designed for type 3 load, and constructed in accordance with PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table IV. The unit shall be adequately secured within the exterior container in such a manner to insure against shifting or breakage. The exterior container design shall include skids, so positioned as to facilitate handling by forklift truck.

5.2.1.2 Case liner. Each level A box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 Strapping. Strapping, when required, shall be in accordance with applicable paragraphs of appendix thereto.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Each unit protected as specified in 5.1.2.1 shall be overpacked in an exterior container conforming to PPP-B-601, table I, using type III, class 1 plywood, or shall be designed for type 3 load, and constructed in accordance with PPP-B-621, table II. The unit shall be adequately secured within the exterior container in such a manner to insure against shifting or breakage. The exterior container design shall include skids, so positioned as to facilitate handling by forklift truck.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:29Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military Government requirements.

Military custodians:
Army. - MD
Navy. - MS
Air Force. - 03

Preparing activity:
Defense Supply Agency. - DM

Review activities:
Army. - MD
Navy. - MS
Air Force. - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:29Z
Check the source to verify that this is the current version before use.

Stock No. 11A 0101 807 2000

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to *preparing activity (as indicated on reverse hereof).*

SPECIFICATION MILITARY SPECIFICATION dated,
CABINET, CASSETTE TRANSFER, WALL MOUNTED MIL-C-36373(DM), 5 February 1965.

ORGANIZATION *(Of submitter)*

CITY AND STATE

CONTRACT NO.

QUANTITY OF ITEMS PROCURED

DOLLAR AMOUNT
$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES.

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity)*

DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:29Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

FOLD UNDER

DEFENSE SUPPLY AGENCY
DEFENSE MEDICAL SUPPLY CENTER
3RD AVENUE AND 29TH STREET
BROOKLYN, N.Y., 11232

POSTAGE FEES PAID
DEFENSE SUPPLY AGENC

OFFICIAL BUSINESS

DEFENSE SUPPLY AGENCY

DEFENSE MEDICAL SUPPLY CENTER

ATTN: DIRECTORATE OF TECHNICAL OPERATIONS (CODE T:5)

3RD AVENUE AND 29TH STREET

BROOKLYN, N.Y., 11232

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:29Z
Check the source to verify that this is the current version before use.