MIL-C-36398A
15 June 1989
SUPERSEDING
MIL-C-36398(DM)
25 FEBRUARY 1965

MILITARY SPECIFICATION

COVER, SURGICAL INSTRUMENT STAND

(Bag Type)

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a suitable bag type protective cover for use with an adjustable, surgical instrument stand (see 6.1).

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications, standards, and handbooks. The following specifications, and standards and handbooks form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents are those listed in the issue of the Department of Defense Index of Specifications and Standards (DODISS) and supplement thereto, cited in the solicitation (see 6.2).

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center ATTN: Directorate of Medical Materiel DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC/NA FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:31Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

V-T-276	Thread, Cotton
CCC-C-430	Cloth, Sheeting, Cotton, and Polyester and Cotton
CCC-C-444	Cloth, Jean, Cotton
DDD-L-20	Label: for Clothing, Equipage, and Tentage, (General Use)
DDD-T-86	Tape, Textile; Cotton, General Purpose, (Unbleached, Bleached or Dyed)
PPP-B-601	Boxes, Wood, Cleated Plywood
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner
PPP-B-636	Boxes, Shipping, Fiberboard

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible

STANDARDS

Federal

FED-STD-191	Textile Test Methods
FED-STD-751	Stitches, Seams, and Stitching

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage
MIL-STD-147	Palletized Unit Loads
MIL-STD-794	Part and Equipment, Procedures for Packaging

(Unless otherwise indicated, copies of federal and military specifications, standards, and handbooks are available from the Naval Publications and Forms Center, (ATTN: NPODS), 5801 Tabor Avenue, Philadelphia, PA 19120-5099.)

2.1.2 Other Government documents, drawings, and publications. The following other Government documents, drawings, and publications form a part of this document to the extent specified herein. Unless otherwise specified, the issues are those cited in the solicitation.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:31Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center (DPSC)

24281 Cover, Surgical Instrument Stand,

(Copies of specifications, standards, and drawings, required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

2.2 Order of precedence. In the event of a conflict between the text of this document and the references cited herein (except for related associated detail specifications, specification sheets or MS standards), the text of this document takes precedence. Nothing in this document, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 Materials.

3.1.1. Cloth, cotton. The basic material used in the fabrication of the cover shall conform to type II, class 2, style A of CCC-C-430, or shall be jean cloth in accordance with CCC-C-444, class 2.

3.1.1.1 Color and colorfastness. The color of the cloth shall be a uniform grayish green, vat dyed, shade 180. When cloth conforming to CCC-C-430 is used it shall conform to the colorfastness requirements therein. When cloth conforming to CCC-C-444 is used it shall show colorfastness to laundering, bleaching, crocking and light either equal to or better than the standard sample. When no standard sample is available the dyed cloth shall show good fastness to laundering, bleaching, crocking and light. Testing shall be as specified in 4.2.1.2.

3.1.2 Thread, cotton. The thread used for stitching and seaming shall be cotton conforming to type IA3, 3 ply, ticket no. 50 of V-T-276.

3.1.2.1 Color of thread. The color of the thread shall be the same as specified for the cloth employed for the cover (see 3.1.1.1).

3.1.2.1.1 Colorfastness. The dyed thread shall show colorfastness to the bleaching, laundering and light requirements specified in V-T-276.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:31Z
Check the source to verify that this is the current version before use.

3.1.3 Tape. The tape shall be cotton, 3/4 inch wide, conforming to type I, class 1 of DDD-T-86.

3.2 Construction. The cover shall be constructed by either of the two methods shown on DPSC drawing 24281. The finished cover shall have two plies of cloth on the front and two plies of cloth on the back. Seams and stitching shall be in accordance with FED-STD-751. Stitch type 301 shall be used throughout. The closed end of the cover shall be enclosed with binding tape and stitched 1/16 inch to 1/8 inch from the edge of the tape. The cover shall finish so that the face of the cloth shall be on the outside and the hem and binding tape shall be on the inside.

Not more than four pieces of cloth shall be used in fabricating the cover. Each piece shall be at least the full length or full width of the cover, as applicable in one direction, and not less than 9 inches in the other (finisned dimensions).

After forming a double length tube per stitching type EFu-1, the top of the tube may be pulled down over the outside of itself to the midpoint.

3.2.1 Thread breaks and ends of seams. Ends of all seams and stitching, when not caught in other seams or stitching, shall be backtacked not less than 1/4 inch. Thread breaks shall be secured by stitching back of the break not less than 1/2 inch.

3.2.2 Stitches per inch. Stitches per inch shall be a minimum of 10 and a maximum of 14.

3.3 Identification marking. The cover shall be marked on the inside of the cover, 1/2 to 1 inch below the hem, using type III or IV markings as specified in DDD-L-20. Markings shall be fast to laundering. Markings shall include the following:

National stock number (NSN).
Item identification.
Contractor's name or registered trademark.
Contract number.

Letters shall be at least 1/8 inch high.

3.4 Workmanship. The finished cover shall conform to the quality of product established by this specification. The occurrence of defects shall not exceed the limits set by the applicable acceptable quality levels.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:31Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements (examinations and tests) as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to ensure supplies and services conform to prescribed requirements.

4.1.1 Responsibility for compliance. All items shall meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of ensuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling inspection, as part of manufacturing operations, is an acceptable practice to ascertain conformance to requirements, however, this does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to accept defective material.

4.1.2 Certificate of compliance. When certificates of compliance are submitted the Government reserves the right to check test to determine the validity of the certification.

4.2 Quality conformance inspection. Unless otherwise specified, sampling for inspection shall be performed in accordance with MIL-STD-105.

4.2.1 Component and material inspection. In accordance with 4.1 above, components and materials shall be inspected in accordance with all the requirements of referenced documents unless otherwise excluded, amended, modified, or qualified in this document or applicable purchase document.

4.2.1.1 Sampling for testing. The sample for testing colorfastness requirements of the cotton jean cloth shall be two yards. The lot size shall be expressed in units of one yard. The inspection level shall be S-1, and the acceptable quality level shall be 4.0 defects per 100 units.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:31Z
Check the source to verify that this is the current version before use.

4.2.1.2 The method of testing specified in FED-STD-191, as listed in table I, shall be followed. All requirements are applicable to the sample unit. The number of determinations per sample unit, and the results reported, shall be as specified by the referenced test method.

Table I. Test methods

Component	Characteristic	Requirement paragraph	Test method
Cotton, jean cloth	Color	3.1.1.1	Visual1/
	Colorfastness to:		
	Laundering	3.1.1.1	5610
	Bleaching	3.1.1.1	5600
	Crocking	3.1.1.1	5651
	Light		5660

1/ One determination per sample unit is required and the result shall be reported as "Pass" or "Fail".

4.2.2 Inspection of the end item. An examination shall be made on the finished cover for visual and dimensional defects.

4.2.2.1 Examination for visual defects. The defects found during this examination shall be classified in accordance with the defects listed in 4.2.2.2 and 4.2.2.3. The sample unit for this examination shall be one cover. The lot size shall be expressed in units of one cover. The sample size shall be selected in accordance with inspection level II of MIL-STD-105. The acceptable quality levels shall be as specified below:

Major defects	2.5 defects per 100 units
Total (major and minor)	6.5 defects per 100 units

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:31Z
Check the source to verify that this is the current version before use.

4.2.2.2 General Defects

Table II. Classification of defects. - Continued.

Defects	Classification Major	Minor
Material defects and damages:		
a. Holes or seriously weakening defects such as smashes, floats, loose and weakening slubs, etc., expected to develop into a hole.	X	
b. Heavy shade bars (ones that are clearly visible at 3 feet, 3 inches in width) or slubs and knots (larger than Sears Scale E for slubs and D for knots).	X	
c. Needle chews which may develop into a hole, scissors or knife cuts, tears, mends, or burns.	X	
Shaded parts:		
Any part noticeably shaded.	X	
Cleanness:		
a. Spots or stains of a permanent nature.	X	
b. Thread ends not trimmed, loose threads, and removable spots or stains not removed.		X

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:31Z
Check the source to verify that this is the current version before use.

Table II. Classification of defects. - Continued.

Defects	Classification Major	Minor
Component and assembly:		
a. Any defect of component or assembly.	X	
b. Any component part or required operation omitted.	X	
c. Length of any piece less than full width or full length of the cover, as applicable.	X	
d. Width of any piece less than 9 inches.	X	
e. Not of double thickness construction.	X	
Cutting:		
a. Component part not cut in conformance with those shown on the drawing, as indicated in paragraph 3.2.	X	
b. More than four pieces used to fabricate the cover.	X	
Seams and stitching:		
a. Seams puckered, twisted, or pleated, affecting appearance.	X	
b. Thread breaks.		X
c. Gauge of stitching: Irregular, i.e., unevenly gauged or not uniformly gauged, affecting appearance.		X

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:31Z
Check the source to verify that this is the current version before use.

Table II. Classification of defects. - Continued.

Defects	Classification Major	Minor
Seams and stitching: (continued)		
d. Open seam more than 1/4 inch.	X	
e. Hem less than 3/4 inch or more than 1 inch.		X
f. Raw edges more than 1/4 inch.		X
g. Raw edges of cloth exposed.	X	
h. Less than 10 or more than 14 stitches per inch.		X
Identification marking:		
a. Not accomplished as specified.	X	
b. Missing, incorrect, or illegible.		X

4.2.2.3 Examination for dimensional defects. During this examination, any measurement deviating from the nominal dimension and tolerance specified on the drawing shall be scored as a size measurement defect. The sample unit for this examination shall be one cover. The lot size shall be expressed in units of one cover. The sample size shall be in accordance with inspection level S-3 of MIL-STD-105. The acceptable quality level shall be 4.0 (one class) defects per 100 units.

4.3 Inspection of packaging. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:31Z
Check the source to verify that this is the current version before use.

5. PACKAGING.

5.1 Preservation. Preservation shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Each cover shall be clean, neatly folded and packaged in a polyethylene bag of an appropriate size. Closure of the bag shall be effected by heat sealing.

5.1.1.2 Intermediate package. Twelve unit packages shall be packaged in a snug-fitting box conforming to PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.1.2 Level C.

5.1.2.1 Unit package. Preservation for covers shall conform to the requirements of MIL-STD-794 for level C.

5.1.2.2 Intermediate package. Twelve unit packages shall be packaged in accordance with the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Two intermediate packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621, class 2, or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Two intermediate packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1, or PPP-B-636, class domestic. Closure shall be as specified in the box specification.

5.2.3 Level C. Two intermediate packages shall be packed in accordance with the requirements of MIL-STD-794 for level C.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:31Z
Check the source to verify that this is the current version before use.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination exceed 250 pounds or 20 cubic feet. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 20 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in length, 52 inches in width, and 54 inches in height. Pallet shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. Covers packed as specified in 5.2.1 and 5.2.2, shall be unitized on pallets as specified in 5.2.5.

5.2.5.2 Level C. Covers packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B, and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.3.

6. NOTES

(This section contains information of a general or explanatory nature that may be helpful, but is not mandatory.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:31Z
Check the source to verify that this is the current version before use.

6.1 Intended use. The bag-type, surgical instrument stand cover, is a suitable protective cover intended for use with NSNs 6530-708-1590, Stand, Surgical Instrument, Adjustable, Corrosion-Resisting Steel, and 6530-551-8681, Stand, Surgical Instrument, Adjustable, Steel, Olive Drab Enameled, Field.

6.2 Acquisition Requirements. Acquisition documents must specify the following.

a. Title, number, and date of this specification.
b. Issue of DODISS to be cited in the solicitation and if required, the specific issue of individual documents referenced (see 2.1).
c. National Stock Number (NSN).
d. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet military requirements.

6.4 All changes in this specification or modification at time of procurements shall be coordinated with the Directorate of Medical Materiel, Defense Personnel Support Center.

6.5 Standard sample. For access to standard sample, address the procuring office issuing the invitation for bids.

6.6 This specification covers the following item shown in the Federal Supply Catalog.

NSN	Item Identification
6530-00-890-2026	COVER, SURGICAL INSTRUMENT STAND, (Bag Type)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:31Z
Check the source to verify that this is the current version before use.

6.7 Subject Term (Keyword) Listing

Protective bag
Wrapper
Envelope
Cotton tool bag

6.8 Changes from previous issue. Marginal notations are not used in this revision to identify modifications, additions, corrections, and deletions from the previous issue due to the extensiveness of the changes.

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:

DOD-MB

Agent:

DLA-DM

Civil Agency Coordinating Activities:

USPHS
VA

Project No. 6530-1764

Location: MILSP/COVER/H13-20

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:31Z
Check the source to verify that this is the current version before use.