| Inch/Pound |

MIL-H-36364B
15 June 1990
SUPERSEDING
MIL-H-36364A
21 May 1970

MILITARY SPECIFICATION

HOLDER, RADIAC DETECTING ELEMENT, FILM BADGE TYPE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE.

1.1 This specification covers a badge type film holder, suitable for holding two packets of personal monitoring film for determining quantity of radiation exposure of an individual.

2. APPLICABLE DOCUMENTS.

2.1 Government Documents.

2.1.1 Specifications, Standards, and handbooks. The following specifications, standards, and handbooks form a part of this document to the extent specified herein. Unless otherwise specified, the issues of these documents shall be listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation (see 6.2).

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC/NA FSC 6665

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Setup.

Military

MIL-P-116	Preservation, Methods of.
MIL-L-10547	Liner, Case, and Sheet, Overwrap, Water Vaporproof or Waterproof, Flexible.

STANDARDS

Federal

FED-STD-66	Steel, Chemical Composition and Hardenability.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.
MIL-STD-147	Palletized Unit Loads.
MIL-STD-794	Parts and Equipment, Procedures for Packaging of.

(Unless otherwise indicated copies of federal and military specifications, standards, and handbooks are available from the Standardization Documents Order Desk, Building 4D, 700 Robbins Avenue, Philadelphia, PA 19111-5094.

2.1.2 Other Government documents, drawings, and publications. The following other Government documents, drawings, and publications form a part of this document to the extent specified herein. Unless otherwise specified, the issues are those cited in the solicitation.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center (DPSC)

20002 Holder, Radiac Detecting Element, Film Badge Type

(Application for copies should be addressed to the Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-RST, 2800 S. 20th Street, Philadelphia, PA 19101.)

2.2 Order of precedence. In the event of a conflict between the text of this document and the references cited herein (except for related associated detail specifications, specification sheets, or MS standards), the text of this document takes precedence. Nothing in this document, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 Material. Holder shall be fabricated of material specified on Drawing No. 20002. Commercial tolerances shall apply to all gages.

3.2 Construction. The holder shall be constructed in accordance with the latest revision of Drawing No. 20002. The holder shall consist of two corrosion-resisting steel sections. A cadmium filter or shield shall be attached to each section. A corrosion-resisting steel clip and a corrosion-resisting steel channel shall be securely attached to the rear section.

3.2.1 Case. The front section shall be capable of sliding into the rear section to form a secure case and shall be removable from the rear section with finger pressure. The case shall be capable of holding two individual personal monitoring film packets, 1-1/4 inches wide by 1-5/8 inches long.

3.2.2 Clip. The clip shall have teeth as shown by the design on Drawing No. 20002. The clip shall be capable of being fastened to the clothing of the wearer and shall maintain its position when tested as specified in 4.4.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

3.3 Workmanship. Holder shall be free from defects which detract from its appearance or may impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements (examinations and tests) as specified herein. Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to ensure supplies and services conform to prescribed requirements.

4.1.1 Responsibility for compliance. All items must meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of ensuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling for inspection, as part of manufacturing operations, is an acceptable practice to ascertain conformance to requirements, however this does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to accept defective material.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory) and testing (determination by technical means of physical and chemical properties).

4.1.3 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the steel, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and Table I.

Table I

	Inspection level	AQL defects per 100 units
For visual examination Major defects	S-4	2.5
For dimensional examination	S-2	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with level S-4 of MIL-STD-105.

4.3 Examination. The holder shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Table I. Categories and defects *

Major	
101	Holder not free of burs or sharp edges.
102	Holder not free of dents.
103	Cadmium shield and clip not secured to outer case.
104	Cadmium shield not secured to inner case.
105	Inner case not secured when slid into outer case.
106	Inner case cannot be removed from outer case (without use of tools).
107	Clip does not function.
108	Materials not as specified.
109	Spring clip pin not sufficiently peened or upset to prevent its removal.
Dimensional	
201	Any dimension not within tolerance specified.

* Examination shall not be restricted to the classified possible defects listed above.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

4.4 Tests. The following tests shall be conducted, to determine compliance with specification requirements.

4.4.1 Pull test. The holder shall be attached by means of the spring clip to leather or denim material and subjected to a straight pull or static load of 5 pounds for 10 seconds.

4.4.2 Composition. The composition of the steel used in the fabrication of the holder shall be determined by the applicable Methods of FED-STD-66.

4.5 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING.

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Cleaning and drying. Holder shall be cleaned and dried in accordance with MIL-P-116.

5.1.1.2 Preservative applications. Preservatives shall not be used.

5.1.1.3 Unit package. Each holder shall be packaged in accordance with MIL-P-116, method IC-1. Closure shall be as specified in the applicable box specification.

5.1.2 Level C.

5.1.2.1 Unit package. Preservation for holder shall conform to the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. One hundred individually packaged holders shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621, class 2, or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. One hundred individually packaged holders shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1 or PPP-B-636, class domestic. Closure shall be as specified in the box specification.

5.2.3 Level C. One hundred individually packaged holders shall be packed in accordance with the requirements of MIL-STD-794 for level C.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination exceed 250 pounds or 20 cubic feet. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 20 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in length, 52 inches in width, and 54 inches in height. Pallet shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. Holders packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

5.2.5.2. Level C. Holders packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

(This section contains information of a general or explanatory nature that may be helpful, but is not mandatory.)

6.1 Intended use. This specification covers a badge type film holder, suitable for holding two packets of personal monitoring film for determining quantity of radiation exposure of and individual.

6.2 Acquisition requirements.

a. Title, number, and date of the specification.

b. Issue of DoDISS to be cited in the solicitation and if required, the specific issue of individual documents referenced (see 2.1).

c. Applicable National Stock Number.

d. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 National Stock Number (NSN) Coverage. This specification covers the following National Stock Number:

National Stock No.	Item Identification
6665-00-299-9925	HOLDER, RADIAC DETECTING ELEMENT

6.4 Subject term (key word) listing.

Holder
Badge type
Film

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6665-0560

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-H-36364B	HOLDER, RADIAC DETECTING ELEMENT, FILM BADGE TYPE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.