MIL-I-36409A
15 June 1978
SUPERSEDING
MIL-I-36409 (DM)
15 March 1965

MILITARY SPECIFICATION

INSERT CABINET, MEDICAL INSTRUMENT AND SUPPLY SET CHEST

This specification is approved for use by all Departments and Agencies of the Department of Defense

1 SCOPE

1.1 Scope This specification covers a metal insert of drawers, suitable for insertion in the standard medical service aluminum field chest (see 6.1)

2 APPLICABLE DOCUMENTS

2.1 Issues of documents The following documents, of the issue in effect on date of invitation for bids or request for proposals, forms a part of this specification to the extent specified herein

SPECIFICATIONS

FEDERAL

QQ-A-200/9	- Aluminum Alloy Bar, Rod, Shapes and Tube, Extruded, 6063
QQ-A-225/6	- Aluminum Alloy Bar, Rod, and Wire, Rolled, Drawn, or Cold Finished, 2024
QQ-A-250/8	- Aluminum Alloy, 5052, Plate and Sheet

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter

FSC 6545

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:33Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

FEDERAL

TT-E-527	- Enamel, Alkyd, Lustreless
TT-P-1757	- Primer Coating, Zinc Chromate, Low Moisture Sensitivity
PPP-B-585	- Boxes, Wood, Wirebound
PPP-B-601	- Boxes, Wood, Cleated-Plywood
PPP-B-621	- Boxes, Wood, Nailed and Lock Corner
PPP-B-636	- Boxes, Fiberboard

MILITARY

MIL-C-5541	- Chemical Conversion Coating or Aluminum and Aluminum Alloys
MIL-A-8625	- Anodic Coatings, for Aluminum and Aluminum Alloys
MIL-L-10547	- Liners, Case, and Sheet, Overwrap, Water-Vaporproof or Waterproof, Flexible

STANDARDS

FEDERAL

Fed Test Method Std No 151	- Metals Test Methods
Fed Std No 595	- Colors

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	- Marking for Shipment and Storage
MS16686	- Medical Services Symbol

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER

10017	- Insert Cabinet, Medical Instrument and Supply Set Chest (with Bill of Materials)

(Copies of specifications, standards, and drawing required by supplier in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:33Z
Check the source to verify that this is the current version before use.

3, REQUIREMENTS

3 1 Materials Except as specified herein, materials shall conform to the applicable portions of the specifications, standards, and drawing (with Bill of Material) referenced herein and shall meet the requirements shown therein When materials to be used are not specifically designated, those used shall be of the highest commercial quality, and suitable for the purpose intended In addition, the metals shall conform to the chemical requirements of their applicable specifications when tested as specified (see 4 5 2)

* 3 1 1 Primer The primer shall conform to color Y of TT-P-1757

* 3 1 2 Enamel Enamel shall conform to TT-E-527, and to color 34087, lusterless olive drab (see 6 3), of Fed Std 595, except that for the marking specified in 3 6, the color shall be 37038, lusterless black (see 6 3), of Fed Std 595 The color for the numerals on the drawer fronts shall be color 33538, lusterless yellow (see 6 3) of Fed Std 595

3 2 Component parts The insert cabinet shall consist of a frame assembly, seven (7) drawers, a tray, a drawer-locking device and two (2) drawer inserts, all of metal, two (2) plastic trays, four (4) truss head screws, fourteen (14) suspension drawer slides, and four (4) keys

3 2 1 Frame assembly The drawer retaining frame assembly consists of two (2) similar halves joined by four (4) metal brackets, fourteen (14) slide frames and the sliding member of the drawer slides, and upper frame plate with a tray guide, a lower frame plate, and a locking bar retainer

* 3 2 2 Drawers The components of each drawer, one (1) through seven (7) are a one-piece drawer body (including sides) an attached drawer front and drawer back, two (2) dust covers, and two (2) drawer slides Each drawer front is provided with a card holder

3 2 2 1 Drawers No 1 and 2 The Drawers No 1 and 2 shall be as specified in 3 2 2

3 2 2 2 Drawer No 3 Drawer No 3 shall be as specified in 3 2 2 and is provided with an attached metal box insert with a hinged cover and lock for narcotics

3 2 2 3 Drawers No 4 5 and 7 Drawers No 4 5 and 7, shall be as specified in 3 2 2, and shall be divided with compartments by partitions

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:33Z
Check the source to verify that this is the current version before use.

3 2 2 4 Drawer No 6 Drawer No 6 shall be as specified in 3 2.2, partitioned and provided with a removable metal box insert with a partition and a hinged cover

3 2 2 5 Drawer locking device The drawer locking device includes a lock, a locking bar, a locking bar rest, a catch (and a locking bar retainer on frame)

3 2 3 Tray The tray is one-piece, bottom, sides, front and back

* 3 3 Construction Construction shall conform in all details to the requirements specified herein and DPSC Drawing No 10017 Construction shall be sound and durable in all respects Assembly shall be welded on jigs capable of maintaining required tolerances The insert cabinet shall fit into a No 3 medical field chest conforming to MIL-C-16776 (see 6.4), when tested as specified (see 4 5 3)

* 3 3 1 Welding Welding shall be done by qualified welders using appropriate aluminum welding equipment in a qulified process Spot welds shall be capable of developing the tensile strength of the parent metal All exposed welds shall be smoothed sufficiently to blend with the adjacent surfaces after the enamel finish specified in 3 5 shall have been applied Welds shall be clean, nonporous, sound, and finished smooth, free from pitmarks and imperfections Other welding shall be continuous on all surfaces, using welding rod of composition similar to parent metal and suitable for the purpose intended

3 4 Keys Lock for the metal box insert for narcotics in Drawer No 3, and the lock for the drawer locking device shall each be provided with two (2) keys Each lock shall be keyed individually so that keys for different inserts are not interchangeable The keys shall be secured within the insert cabinet so that they will not move around or be lost

3 5 Finish After welding has been completed but before final assembly and attachment of cardholders to drawer fronts, all accessible aluminum alloy surfaces of the insert cabinet, except the working surfaces of the drawer slides, shall be finished as specified herein All aluminum alloy surfaces shall be thoroughly cleaned (free from dirt, oil, grease, welding flux, and loose scale) using solvents, detergents, or other processes which have no deleterious effect on the surfaces

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:33Z
Check the source to verify that this is the currentversion before use.

3 5.1 Topcoating. After thorough cleaning and degreasing, surfaces to be enameled shall be prepared for topcoating by one of the following three methods·

(1) Chemical treatment in accordance with MIL-C-5541
(2) Anodic treatment in accordance with MIL-A-8625.
(3) Priming treatment using primer as specified in 3 1 1, thinned and applied by spray in accordance with the directions specified in TT-P-1757, and air-dried for not less than 30 minutes before topcoating

Topcoating shall be accomplished with one coat of enamel, as specified in 3 1 2, applied by spray and cured by a heat treatment equivalent to baking at 250°F (Fahrenheit) for not less than 45 minutes The finish shall be hard, smooth, and highly resistant to abrasion, chipping, cracking, acid, alkali, and solvents Cardholder shall be attached to the drawer fronts before topcoating is applied

3 5 2 Film thickness of paint Contractor shall endeavor to control film thickness so that the thickness of a single dried coat of paint is 0.3 to 0 7 mil (008 to 018 mm) for primer and 1 0 to 1 5 mil (025 to 038 mm) for enamel

* 3 6 Identification markings The insert cabinets shall be permanently and legibly marked with enamel as specified in 3.1.2, as follows

(a) Suitably located on the front of Drawer No 3, there shall be a 1 inch diameter Medical Services Symbol as shown for type I in MS16686

(b) Centrally located on the underside of the lower frame plate (Part No 26), there shall appear, the following information

(1) In vertical block letters 3/8 inch (9 53 mm) high, the legend

"Insert Cabinet, Medical Instrument and Supply Set Chest"

(2) In vertical block numerals 3/8 inch (9 53 mm) high, the National Stock Number "6545-00-914-3485"

(3) In vertical letters 1/4 inch (6 25 mm) high, the name or registered trademark of the manufacturer

(4) In vertical letters 1/4 inch (6 25 mm) high, the contract or purchase order number

Each of the drawers shall be marked with its designated number (see drawing), positioned on each drawer front as shown on the drawing

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:33Z
Check the source to verify that this is the current version before use.

3 7 Workmanship Workmanship shall be first class throughout The frame and drawers shall be square and free from twist Drawers shall operate smoothly and easily All burs shall be removed and all sharp corners broken Completed insert cabinets shall be free from defects which detract from their appearance or impair their serviceability

4 QUALITY ASSURANCE PROVISIONS

* 4 1 Responsibility for inspection Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification to assure supplies and services conform to prescribed requirements

4 1 1 Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

4 1 2 Inspection Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

4 1 3 Certificates of quality Certificates of quality, supplied by the manufacturer of the aluminum sheet and channels and corrosion-resisting steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government The certificates shall include the name of the contractor the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material the lot number, the lot size, the sample size, the date of testing, the test method, individual test results and the specification requirements

* 4 2 Quality conformance inspection The examination and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables I, I' and III (see 4 2 1, 4 2 2 and 4 3)

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:33Z
Check the source to verify that this is the current version before use.

* 4 2 1 Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105 The inspection level and acceptable quality level (AQL) shall be as indicated in table I The unit of product for examination shall be one insert cabinet.

TABLE I Sampling for examination

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	4 0

* 4 2.2 Sampling for tests Sampling for tests shall be conducted in accordance with table II, and MIL-STD-105, except that the acceptance number shall be zero for all sample sizes Minimum sample size shall be three The unit of product for tests purposes shall be one insert cabinet

TABLE II Sampling for tests

Characteristic	Requirements	Test procedure
Metals	3 1	4 5.2
Fit	3 3	4 5 3

4 3 Examination The insert cabinet shall be examined to determine compliance with all requirements contained in this specification Examination shall be in accordance with, but shall not be limited to the defects listed in Table III

TABLE III Classification of defects

Categories	Defects
Critical	None defined
Major	
101	Unit not complete with all specified component parts
102	Component parts not free of cracks, tears, ruptures, or deformed surfaces
103	Exposed and accessible edges not free of burs and sharpness

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:33Z
Check the source to verify that this is the current version before use.

TABLE III Classification of defects (cont'd)

Categories	Defects
Major	
104	Welds not sound or not smoothly finished
105	Unit does not fit in a 6545-00-914-3480 field chest
106	Drawers do not completely close
107	Finish not smooth, even, and adherent
108	Unit identification markings not complete and correct
109	Exterior container markings not complete and correct
110	Keys for one lock, or keys for both locks, missing
111	Key does not operate designated lock
Minor A	
201	Color of plastic basins off white
202	Identification marking not completely legible
203	Enameled surfaces not free of excessively chipped or scratched areas
204	Color of enamel not as required
205	Movable parts do not operate smoothly
206	Not clean, free of grease, oil, dirt, or other foreign matter
207	Duplicate keys not supplied
208	Keys not secured within insert cabinet
Minor B	
251	Dimensions not within tolerances shown or the drawing

4.4 Packaging inspection The inspection of the individual and unit packaging, packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:33Z
Check the source to verify that this is the current version before use.

4.5 Tests

* 4 5 1 Sampling for chemical testing of metals. A sufficient amount of each lot of each metal used in the construction of the cases shall be selected for the chemical analysis as specified in 4 5 2

* 4 5 2 Chemical analysis of metals Samples selected in accordance with 4 5 1 shall be analyzed in conformance with Methods 111 or 112 of Fed-Std-151 In case of dispute, Method 111 shall be used for referee analysis

* 4 5 3 Fit test The frame of the insert cabinet shall be positioned in a No 3 medical field chest (see 6 4) and secured in place with four (4) locking screws The chest shall be set on end, and each of the 7 drawers shall be inserted in the frame to their full depth

5 PACKAGING

5 1 Packaging Packaging shall be level A or C, as specified (see 6 2)

5 1 1 Level A

5 1 1 1 Unit preparation The drawers shall be secured in a closed position in such a manner as to prevent movement or damage during transit or handling The unit shall be further protected by use of corrugated fiberboard or other suitable cushioning material to insure against marring, scratching, or damage to the unit The cushioning material shall be secured by the use of tape, non-metallic strapping or other suitable material

5 1.2 Level C Each insert shall be packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from the supply source to the first receiving activity

5 2 Packing Packing shall be level A, B or C as specified (see 6 2)

5 2 1 Level A Each insert shall be packaged in a shipping container designed for a-type 3 load and conforming to PPP-B-585, class 3, style 3, PPP-B-601 overseas type, PPP-B-621, class 2 or PPP-B-636, class weather-resistant The insert shall be blocked, braced and cushioned within the shipping container in such a manner to prevent movement and damge during shipment and handling Closure and strapping shall be as specified in the appendix of the applicable box specification

5 2 1 1 Case liner Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547 Closure and sealing shall conform to applicable paragraphs of appendix thereto Case liner shall not be required for fiberboard boxes Each fiberboard box shall be waterproofed as specified in the appendix of PPP-B-636

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:33Z
Check the source to verify that this is the current version before use.

5 2 2 Level B Each insert shall be packaged in a shipping container designed for a type 3 load and conforming to PPP-B-585, class 1, style 3, PPP-B-601, domestic type, PPP-B-621, class 2, or PPP-B-636, class domestic The insert shall be blocked, braced and cushioned within the shipping container in such a manner to prevent movement or damage during shipment and handling Closure shall be as specified in the appendix of the applicable box specification

5 2 3 Level C Each packaged insert shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity These packs shall conform to the applicable carrier's rules and regulations

5 3 Marking Each shipping container shall be marked as specified in MIL-STD-129

6 NOTES

6 1 Intended use The insert cabinet covered by this specification, is intended for use with the following Chest appearing in the Federal Supply Catalog

National Stock Number	Item Identification
6545-00-914-3480	CHEST, MEDICAL INSTRUMENT AND SUPPLY SET, Field, No 3, 30 inches long by 18 inches wide by 10 inches deep, empty

6 2 Ordering data Procurement documents should specify the following

(a) Title, number and date of this specification
(b) National Stock Number (NSN)
(c) Selection of applicable levels of packaging and packing (see 5 1 and 5 2)

6 3 Availability of color chips Individual chips of the colors specified in 3 1 2, are available from Commanding Officer, U.S Naval Publications and Forms Center, 5801 Tabor Avenue, Philadelphia, PA 19120

6 4 Sample chest A NSN 6545-00-914-3480, No 3 Medical Instrument and Supply Set, Field Chest may be obtained on a loan basis upon application to the Contracting Officer and payment of the required deposit

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:33Z
Check the source to verify that this is the current version before use.

6 5 This specification does not cover all types, classes, grades or sizes of the commodity indicated by the title of this specification or those which are commercially available, but is intended to cover the types normally procured to meet military requirements

6 6 This specification covers the following item as listed in the Federal Supply Catalog

National Stock Number	Item Identification
6545-00-914-3485	INSERT CABINET, MEDICAL INSTRUMENT AND SUPPLY SET CHEST, Aluminum

* Changes from previous issue The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue

Custodians
Army - MD
Navy - MS
Air Force - 02

Review activity
DoD-MB

Preparing activity
DLA-DH

Project number 6545-0036

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS This form is provided to solicit beneficial comments which may improve this document and enhance its use DoD contractors government activities manufacturers vendors or other prospective users of the document are invited to submit comments to the government Fold on lines on reverse side staple in corner and send to preparing activity Attach any pertinent data which may be of use in improving this document If there are additional papers attach to form and place both in an envelope addressed to preparing activity A response will be provided to the submitter when name and address is provided within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed

NOTE This form shall not be used to submit requests for waivers deviations or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

DOCUMENT IDENTIFIER (Number AND TITLE

MIL-I-36409A - INSERT CABINET, MEDICAL INSTRUMENT AND SUPPLY SET CHEST

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1 ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID RESTRICTIVE LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDED WORDING CHANGE

C REASON FOR RECOMMENDED CHANGE(S)

2 REMARKS

SUBMITTED BY (Printed or typed name and address — Optional)	TELEPHONE NO
	DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:33Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

POSTAGE AND FEES PAID

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN. DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:33Z
Check the source to verify that this is the current version before use.