MILITARY SPECIFICATION

PRISM SET, OPHTHALMOLOGICAL, BERENS

1. SCOPE

1.1 This specification covers prisms for use in ophthalmology.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-up, Paperboard.

Military

MIL-L-10547	Liners, Case, and Sheet Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:29Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

Fed. Test Method Std. No. 406 Plastics: Methods of Testing.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.
MS16686	Medical Services Symbol.

DRAWING

Defense Medical Supply Center

20257 PRISM SET, OPHTHALMOLOGICAL, BERENS

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. The prisms shall be fabricated from a crystal clear acrylic plastic. Plastic shall have an index of refraction of 1.49 to 1.52 with a light transmittance of not less than 90 percent, when tested as specified in section 4.

3.2 Dimensions, style and design. The ophthalmological set shall consist of individual, square prisms of the Berens design plus a red multiple Maddox rod handle. The set shall be in accordance with the Drawing 20257.

3.2.1 Diopter value. Diopter values shall be as indicated in drawing and shall be tested as specified in section 4.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:29Z
Check the source to verify that this is the current version before use.

3.3 Case. A leatherette covered wooden carrying case shall be supplied with each unit. The carrying case shall be compartmented and velvet lined to adequately accommodate the components listed above and prevent against their shifting and damage during carrying of the complete set and when the set is not in use. The carrying case (inside or outside) shall be permanently and legibly marked with the name or registered trademark of the contractor. The carrying case (inside or outside) shall also be permanently and legibly marked with the company model number of the set, if such number is used. The outside surface of the top of the case shall be permanently and legibly marked with a caduceus centered between the letters "U" and "S" in accordance with type I or III of MS16686 plus the Federal stock number for the set. The caduceus shall be approximately 1 inch in height and the letters "U" and "S" and Federal stock number of appropriate size. A contrasting color to that used for the leatherette shall be suitable for the marking of the caduceus and the letters "U" and "S". In addition, the inside surface of the top of the case shall bear a label stating that the prisms shall be cleaned with mild soap and water and that alcohol shall never be used.

3.4 Workmanship. Workmanship shall be first class throughout. Set shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:29Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 Specimens. Specimens for determining compliance with index of refraction and light transmittance requirements may be prepared from coupons fabricated from the same batch of material.

4.2.2 Lot. A lot, for inspection purposes (other than for the index of refraction, transmittance, handle, case, or packaging, packing and marking) shall be defined in terms of the number of prisms of the same diopter. Lot for index of refraction and light transmittance tests shall be defined as material from the same batch.

4.2.3 Sampling for visual examination. Sampling procedure for visual examination shall be conducted in accordance with MIL-STD-105.

4.2.4 Sampling for tests.

4.2.4.1 Diopter test. Shall be conducted in accordance with MIL-STD-105 for level S-1 and acceptance number shall be zero (0) for all sample sizes.

4.2.4.2 Index of refraction and light transmittance. Sufficient samples shall be selected from each lot to perform the tests. Failure to comply with test requirements shall be cause for rejection of the lot represented.

4.2.5 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The prism sets shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with this specification.

4.4.1 Index of refraction. Test shall be conducted in accordance with Standard No. 406, method 3011.

4.4.2 Light transmittance. Test shall be conducted in accordance with Standard No. 406, method 3022.

4.4.3 Diopter. The diopter power shall be measured on any standard lens measuring instrument capable of yielding the required accuracy.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:29Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) prism set, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	24 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Unit package shall be a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable, or PPP-B-636. Closure shall be adequate to prevent spillage of contents under normal handling.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:29Z
Check the source to verify that this is the current version before use.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix of PPP-B-636. In addition, waterproofing shall be as specified in 30.4 of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood, or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic, fabricated from not less than 275-pound-test fiberboard. Closure of fiberboard boxes shall be in accordance with Method I of appendix to box specification.

5.3 Marking.

5.3.1 Interior packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:29Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.

(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:29Z
Check the source to verify that this is the current version before use.