MIL-S-36366A
Amendment-1
22 March 1974

MILITARY SPECIFICATION

STERILIZER, SURGICAL INSTRUMENT

This amendment forms a part of Military Specification MIL-S-36366A, dated 27 November 1972 and is approved for use by all Departments and agencies of the Department of Defense.

Page 2

3.1 Fire and casualty hazards. Delete heading and substitute: "Fire and casualty hazards."

Page 4

3.3.1 Housing. Second sentence after "housing" insert ", excluding knobs and handles,".

Page 5

3.4.4 Thermostat. Second sentence, delete "calibrated";after "knob" insert "and graduated scale". Third sentence delete "±3" and substitute "± 5".

Page 10

4.4.1 Performance. Second sentence delete "petrolatum" and substitute "clean sand".

4.4.2 Temperature rise. Second sentence after "heat position" insert "of 360 degrees F.".

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review Activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DSA-DM

Project No. 6530-1312

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.