MIL-T-36368A

30 September 1977

SUPERSEDING
MIL-T-36368(DM)
22 January 1965

# MILITARY SPECIFICATION

## TUBES, NASAL FEEDING, SURGICAL, INFANT, DISPOSABLE

### 1. SCOPE

1.1 Scope. This specification covers two types of sterile, disposable plastic nasal feeding tubes suitable for feeding infants.

1.2 Classification. The tubes shall be furnished in either of the following types:

Type I. With integral Luer connector formed from plastic tubing and cap.

Type II. With attached but separate Luer connection and plug.

### 2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

| | | |
|---|---|---|
| PPP-B-566 | - | Boxes, Folding, Paperboard. |
| PPP-B-585 | - | Boxes, Wood, Wirebound. |
| PPP-B-601 | - | Boxes, Wood, Cleated-Plywood. |
| PPP-B-621 | - | Boxes, Wood, Nailed and Lock-Corner. |
| PPP-B-636 | - | Boxes, Shipping, Fiberboard. |
| PPP-B-676 | - | Boxes, Set-up. |

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Hq, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 20260-21T08:28Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-P-116 - Preservation-Packaging, Methods of.
MIL-L-10547 - Liners, Case, and Sheet Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)
No. 20524 TUBE, NASAL FEEDING

(Copies of specification, standards, drawings, and publications, required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

* 2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise specified, the issue in effect on date of invitation for bids or request for proposals shall apply.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)
ANSI Z70.1 Standard for Luer Tapers for Medical Applications

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, NY 10018).

UNITED STATES PHARMACOPEIA CONVENTION, INC.
Pharmacopeia of the United States (USP).

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042).



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE,
FOOD AND DRUG ADMINISTRATION:
Federal Food, Drug, and Cosmetic Act and
Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204).

(Technical society and technical association specifications and standards are generally available for reference from libraries. They are also distributed among technical groups and using Federal agencies.)

3. REQUIREMENTS

3.1 Material. All components coming in contact with solutions being administered shall contain not more than 5 parts per million (ppm) of lead when tested as specified in 4.4.1.

3.1.1 Plastic tubing. Plastic tubing shall be transparent, medical-grade, polyvinyl chloride, and shall be sterile, pyrogen-free and meet the requirements for Class IV plastics of the USP as specified in 4.4.2.

3.1.2 Other plastic parts. All other plastic parts shall be of the best quality, and suitable in all respects for the purpose intended.

3.2 Style, design, and construction. The nasal feeding tubes shall be either type I or type II in accordance with Drawing 20524, and as hereinafter specified.

3.2.1 Type I. Type I nasal feeding tube shall be fabricated with a Luer fitting that is an integral part of the plastic tubing (formed from this tubing) with a cap attached to the tubing. The tube shall comply with the requirements of 3.1, 3.1.1, 3.3, 3.4, 3.7 and 3.8.

3.2.2 Type II. Type II nasal feeding tube shall be fabricated with an attached, but separate Luer fitting and plug. The tube shall comply with the requirements of 3.1, 3.1.1, 3.3, 3.4, 3.7 and 3.8.

* 3.3 Sterility. The nasal feeding tubes shall be sterile when tested with a biological indicator as specified in 4.4.3.

* 3.3.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

a. Item identification.
b. National stock number (NSN).
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.)
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pa., 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

* 3.3.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and sterilized simultaneously in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.4 Freedom from pyrogens. Tubes shall be pyrogen-free when tested for freedom from pyrogens as specified in 4.4.4.

3.5 Certificates of freedom from pyrogens and compliance with USP Class IV requirements. A certificate(s) listing all lot or control numbers, shall state that representative sample(s) from each lot or control have been tested and found to be free from pyrogens, and have passed the USP Class IV requirements. This certificate(s) shall include the specific tests performed and the test results. The record of the animal implantation test and its results shall be certified by the pathologist who performed or supervised performance of the animal implantation test. For destination inspection only, one copy of certificate(s) shall be submitted by the contractor directly to



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101 and one copy shall be forwarded with each shipment to the consignee.

3.6 Manufacturer's record. The manufacturer shall place a code or lot number on the paper label, or on the printed carton, which can be verified against the manufacturer's records of all tests made on nontoxic animal implantation test, sterile, pyrogen-free tubes. These records shall be made available for inspection by any duly authorized Government representative. No shipment shall be made until all tests have been completed, verified, and recorded in the permanent records of the manufacturer.

3.7 Security of assembly.

3.7.1 Type I. The plastic cap shall not separate or loosen from the plastic tubing when tested as specified in 4.4.5.

3.7.2 Type II. Plastic Luer connector and caps shall not separate or loosen from the plastic tubing when tested as specified in 4.4.5.

3.8 Leakage of tubes. The female Luer cavity shall make a secure watertight union with the male Luer slip connector when tested as specified in 4.4.6. In addition, the type II set shall be watertight when tested as specified in 4.4.6. There shall be no falling drop in ten seconds.

3.9 Workmanship. Workmanship shall be first class throughout. Tubes shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specifications to assure supplies and services conform to prescribed requirements.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

* 4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of the plastic may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified in table I. The unit of product for examination purposes shall be one nasal feeding tube.

TABLE I. Sampling criteria for examination.

| Type of Defect | Inspection level | AQL (percent defective) |
|---|---|---|
| Major | S-3 | 2.5 |
| Minor | S-2 | 4.0 |

* 4.2.2 For Tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table II. The unit of product for test purposes shall be one nasal feeding tube. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

TABLE II. Sampling criteria for tests.

| Characteristic | Require-ment | Test Procedure | Inspection level | AQL (percent defective) |
|---|---|---|---|---|
| Lead | 3.1 | 4.4.1 | S-1 | 1.5 |
| Class IV USP | 3.1.1 | 4.4.2 | S-2 | 1.0 |



Source: https://assist.dla.mil -- Downloaded: 202610-21T08:28Z
Check the source to verify that this is the currentversion before use.

TABLE II. Sampling criteria for tests, Continued.

| Characteristic | Requirement | Test Procedure | Inspection level | AQL (percent defective) |
|---|---|---|---|---|
| Dimensions | 3.2 | --- | S-3 | 1.5 |
| Sterility | 3.3 | 4.4.3 | S-2 | 1.0 |
| Freedom from pyrogens | 3.4 | 4.4.4 | S-2 | 1.0 |
| Security of assembly | 3.7 | 4.4.5 | S-2 | 1.5 |
| Leakage | 3.8 | 4.4.6 | S-2 | 1.5 |

* 4.3 Examination. The nasal feeding tubes shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in tables I and III. Examination, however, shall not be restricted to defects listed in table III. Any tubes in the sample containing one or more defects shall be rejected, and if the number of defective tubes in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

TABLE III. Classification of defects.

| Category | Defects |
|---|---|
| Major | |
| 101 | Not free from rough edges on exterior of tube. |
| 102 | Cap or plug difficult to remove from proximal opening. |
| 103 | Cap or plug not providing effective seal. |
| 104 | Tubing permanently kinked. |
| 105 | Presence of foreign material. |
| 106 | Crack in plug or cap. |
| Minor | |
| 201 | Adherent burs on cap or plug not interfering with function. |
| 202 | Tubing not clear. |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

4.3.1 Packaging inspection. The inspection of the individual package shall be in accordance with group A and B quality conformance inspection requirements of MIL-P-116. Leakage test specified for group B is not applicable. The inspection of the unit package, packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specifications and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Test for lead. Ten (10) grams of sample shall be taken, ignited to char, and then further ignited in a muffler of 700°Centigrade (°C) to a white ash. The ash shall be boiled in 15 milliliters (ml) of 10 percent nitric acid filtered into a volumetric flask and made up to 100 ml. An aliquot portion, equivalent to one gram of sample, shall be taken and placed in a separator. The dithizone test for lead, as given in the USP, shall be applied using a control of 5 ppm or 12 ppm of lead solution as given in the USP test. If the sample solution is darker than the limit, smaller aliquots shall be used until the color of the control is matched by the sample solution and the concentration of lead in ppm is determined.

4.4.2 USP Class IV tests. The catheters shall be tested in accordance with the requirements of the current edition of the USP for class IV plastics.

4.4.2.1 Animal implantation test. Samples of the catheters shall be tested in accordance with the method given in the USP for Implantation Test for class IV plastic, except for the changes which follows: Last paragraph, second line, after "72 hours" add "and not to exceed 144 hours". Last paragraph, third and fourth sentences, delete entirely and substitute: "The tissues surrounding the implants shall be examined, both grossly and microscopically, by a pathologist certified by the American Board of Pathology or the American College of Veterinary Pathologists."

* 4.4.3 Sterility test. The nasal feeding tubes shall be tested for sterility in accordance with the method specified in the USP for non-parenteral devices using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

temperature shall be as indicated in paragraph 4.4.3.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for Type I Sterilization and 30 for Type II Sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in Table II, Sampling Criteria for Tests.

* 4.4.3.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

| Type of Sterilization | Organism | Incubation Temperature |
|---|---|---|
| Steam | B. stearothermophilus | 55° - 65°C |
| Dry Heat | B. stearothermophilus, or | 55° - 65°C |
| | B. subtilis var.niger (or globigii) | 30° - 35°C |
| Gas | B. subtilis var. niger (or globigii) | 30° - 35°C |
| Radiation | B. pumilus | 30° - 35°C |

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

4.4.3.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.3 Sterility test and 4.4.3.1 Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.3.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

* 4.4.3.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

4.4.3.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP.

4.4.4 Freedom from Pyrogens. Tubes shall be tested for freedom from pyrogens in accordance with the USP, for transfusion and infusion assemblies.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

4.4.5 Security of assembly. The tubes shall be conditioned at 240° Fahrenheit (°F) and 100 percent relative humidity for 20 minutes. The security of assembly of cap (type I) or Luer connection and cap (type II) shall then be tested by gradually applying a 5-pound weight to cap or Luer connection (as applicable) and maintaining the load for 15 seconds, while that component is held in a clamping device so that the tubing hangs vertically downward.

* 4.4.6 Leakage of set. Tubes, after sterilization, shall be subjected to an internal pressure of 4 pounds per square inch for ten seconds. This test shall be performed with a gage shown in ANSI Z70.1 (and having a suitable fluid path throughout the length of the gage) fitted to the female Luer connection. This test shall be performed with the tube in a horizontal position and with the tube sealed at the distal end.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C as specified (see 6.1).

5.1.1 Unit of issue. One package, containing 50 tubes of type I or type II as specified, constitutes one unit of issue.

5.1.1.1 Level A.

5.1.1.1.1 Individual package. Each tube shall be packaged in a sealed peel type envelope capable of maintaining sterility until manually opened. The envelope shall be designed with tabs to facilitate opening.

5.1.1.1.2 Unit package. Fifty (50) tubes packaged as specified in 5.1.1.1.1 shall be packaged in a box of appropriate size and constructed in accordance with PPP-B-566, PPP-B-676, or PPP-B-636, type CF, class domestic. Closure of box shall be as specified in the appendix of the applicable box specification.

5.1.1.2 Level C. Tubes shall be individually packaged in a manner that will afford adequate protection against deterioration and physical damage during shipment from supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A. Ten (10) unit packages shall be packed in an exterior container designed for a type 2 load constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.



Source: https://assist.dla.mil -- Downloaded: 20260-21T08:28Z
Check the source to verify that this is the current version before use.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof as specified in the appendix of PPP-B-636.

5.2.2 Level B. Ten (10) unit packages shall be packed in an exterior container designed for a type 2 load constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, type CF, class domestic. Closure of box shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The packaged tubes shall be packed in shipping containers in a manner that will afford adquate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier's rules and regulations.

5.3 Marking.

5.3.1 Individual package. In addition to commercial marking, each package shall be marked as required by the Federal Food, Drug and Cosmetic Act. Marking shall include the legends: "STERILE and STERILITY GUARANTEED UNLESS PACKAGED IS DAMAGED OR OPENED" or similar statement.

5.3.2 Unit package and exterior container. Each unit package and exterior container shall be marked as specified in MIL-STD-129. In addition, marking shall include the method of sterilization (G-Gas, R-Radiation, C-Cold, S-Steam, D-Dry, Heat Sterilization). Marking shall follow the level of packaging and date of packaging. Example: "A-1/77-G." Final sterilization shall be on or after indicated packaging date.

6. NOTES

6.1 Ordering data. Purchasers should exercise any desired options offered herein, and procurement documents should specify the following:

a. Number, title, and date of this specification.
b. Type (see 1.2).
c. Selection of applicable levels of packaging and packing.
d. Applicable NSN.

6.2 NSN coverage. This specification covers the following items listed in the Federal Supply Catalog:



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

| National Stock Number | Item Identification |
|---|---|
| 6515-00-841-3664 | TUBE, NASAL FEEDING, SURGICAL, Premature Infant, Disposable, 5 Fr, 50's |
| 6515-00-841-3665 | TUBE, NASAL FEEDING, SURGICAL, Infant, Disposable, 8 Fr, 50's |

* 6.3 Changes from previous issues. The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Offerors and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue.

Custodians:
Army - Md
Navy - Ms
Air Force - 03

Review activity:
DOD-MB

Preparing activity:

Defense Logistics Agency - DM

Project No. 6515-1979



Source: https://assist.dla.mil -- Downloaded: 202610-21T08:28Z
Check the source to verify that this is the current version before use.

**INSTRUCTIONS:** In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

*NOTE:* This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

*(Fold along this line)*

*(Fold along this line)*

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

**OFFICIAL BUSINESS**
**PENALTY FOR PRIVATE USE $300**

**BUSINESS REPLY MAIL**
FIRST CLASS PERMIT NO. 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-LS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:28Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

*(See Instructions – Reverse Side)*

| 1. DOCUMENT NUMBER | 2. DOCUMENT TITLE |
|---|---|
| | |

| 3a. NAME OF SUBMITTING ORGANIZATION | 4. TYPE OF ORGANIZATION *(Mark one)* |
|---|---|
| | ☐ VENDOR<br>☐ USER<br>☐ MANUFACTURER<br>☐ OTHER *(Specify)* |
| b. ADDRESS *(Street, City, State, ZIP Code)* | |
| | |

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason for Recommendation:

6. REMARKS

| 7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional | b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional |
|---|---|
| c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional | 8. DATE OF SUBMISSION *(YYMMDD)* |
| | |

DD FORM 1426

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026l0-21T08:28Z
Check the source to verify that this is the current version before use.