MIL-T-36381 (DM)
9 February 1965

MILITARY SPECIFICATION

TUBE, CAPILLARY, BLOOD SAMPLE, GLASS,
0.8 MM INSIDE DIAMETER, 90 MM LONG

1. SCOPE

1.1 This specification covers one type of plain capillary tube, suitable for use in the C-reactive protein determination (see 6.1).

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

FSC 66307

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:30Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material. Capillary tube shall be made of clear, transparent, borosilicate or soda-lime glass, suitable for the intended purpose.

3.2 Construction. Tube shall be a plain capillary glass tube (see 3.1), and shall be suitable for use in the C-reactive protein determination test. Tube shall have an inside diameter of 0.8 millimeter (mm.) ± 0.15 mm., an outside diameter of not greater than 1.22 mm., and a length of 90 mm. ± 1 mm.

3.3 Count. The average count shall be not less than 75 tubes per package, and no package shall contain less than 70 tubes, when tested as specified in 4.4.1.

3.4 Workmanship. Workmanship shall be first class throughout. Capillary tubes shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. Unit of product shall be the tube.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:30Z
Check the source to verify that this is the current version before use.

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	2.5
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105.

Component (unit of product)	Characteristic	Requirement	Test procedure	Inspection level	AQL
Package	Count	3.3	4.4.1	S-4	*

* Acceptance number shall be zero.

4.3 Examination. The capillary tubes shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Categories and defects *

Major
101 Tube not free of sharp ends.
102 Tube not free of chips and cracks.
103 Ends not cut perpendicular to tube axis.
104 Ends of tube uneven or jagged.
105 Tube not clear and transparent.
106 Tube not straight or free of bends.
107 Tube not free of dirt and foreign matter.

* Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensional examination. The capillary tubes shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Count. The number of tubes in the package shall be determined.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:30Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. Seventy-five (75) tubes, as specified, constitute one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	576 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Unit package shall be a transparent glass or plastic container of appropriate size and design. Container closure shall be adequate to prevent accidental opening.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable, or PPP-B-636. Closure shall be adequate to prevent spillage of contents under normal handling.

NOTE: When unit package is a glass container, intermediate package design shall include suitable interior fittings or shall contain sufficient cushioning to insure against shifting of, or damage to, the contents.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:30Z
Check the source to verify that this is the current version before use.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform to 30.4, of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic, and method I closure shall be utilized.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown. For unit packages, the required information may be shown on a label inserted within the container in such a manner as to permit ready identification.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

Note: The quantity and unit shall be designated on the interior packages and exterior container as follows:

Unit package:	1 PKG (of 75 tubes)
Intermediate package:	12 PKG
Exterior container:	576 PKG

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:30Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. Capillary tube covered by this specification is intended for use with Federal stock number 6630-542-1269, Rack, BLOOD SAMPLE CAPILLARY TUBE, Plastic and Federal stock number 6505-660-1604, ANTISERUM, C-REACTIVE PROTEIN, 1 cc.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable level of packing (see 5.2).

6.3 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:30Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

Commander
Defense Personnel Support Center
ATTN; DPSC-ATTS
2800 South 20th Street
Philadelphia, Pa. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:30Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

S/N 0102-014-1802

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:30Z
Check the source to verify that this is the current version before use.