MIL-W-36401A
11 November 1974
SUPERSEDING
MIL-W-36401(DM)
3 March 1965

MILITARY SPECIFICATION

WARMER, PLASTIC SPECTACLE FRAME

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a bath type warming device for heating plastic spectacle frames.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Boxes, Shipping, Fiberboard

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.
MIL-M-7298	Manual, Technical, Commercial Equipment.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

FSC 6540

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:32Z
Check the source to verify that this is the current version before use.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Fire and casualty hazards.

3.1.1 The subject item shall comply with the fire and casualty hazards provisions of the applicable standards and be an approved type for the purpose intended, as determined by a nationally recognized testing agency, adequately equipped and competent to perform such services, as shall be evidenced by the attachment of its seal, stamp or label to the article, or issuance of a certificate. The stamp, label, seal or certificate of the Underwriters' Laboratories (New York, N.Y.), the United States Testing Laboratories, Inc. (Hoboken, New Jersey), the Electrical Testing Laboratories, Inc. (New York, N.Y.), or other testing laboratory approved by the contracting agency shall be considered acceptable evidence of such approval; however, in addition to the aforementioned, the contracting agency reserves the right to require the contractor to submit a laboratory test report which shall include test results and methods used to determine conformance.

3.1.2 Prior acceptance of a particular design after satisfactory use by the government may be used as evidence of approval.

3.1.3 Compliance with the above requirements regarding fire and casualty hazards, does not absolve the contractor from complete compliance with the requirements of this specification in order to secure the acceptance of his commodity.

3.2 Materials.

3.2.1 Outside shell (wrap around). The outside shell shall be aluminum or cold rolled steel sheet.

3.2.2 Pan. The pan used to contain the bath (glass beads) shall be aluminum plate or sheet.

3.2.3 Border trim. The border trim shall be case aluminum.

3.2.4 Bottom plate. The bottom plate shall be aluminum or cold rolled steel.

3.2.5 Heat control knob. The heat control knob shall be phenolic, styrene or other suitable plastic.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:32Z
Check the source to verify that this is the current version before use.

3.2.6 Heater assembly. Commercial type heating elements (mica or resistance wire) shall be used.

3.2.7 Feet or Grommets. Feet or grommets shall be black synthetic rubber.

3.3 Style, design, and construction. The plastic spectacle frame warmer shall be an indirect bath type warming device to heat and soften frames for glazing aligning, and adjusting, and shall operate on 115 volt, 60 Hz, alternating current. The pan size shall be 7 ± 1/4 inches long by 5 ± 1/4 inches wide and 2-3/4 ± 1/4 inch deep. The warmer shall be of all-metal construction and shall have a built-in heating element which shall be thermostatically controlled for constant bath temperature. The unit shall have a rheostat control for temperature selection. The warmer shall be provided with an airspace between the inner container and the outer shell to allow for circulation of air to prevent the outside of the shell from reaching a high temperature. The unit shall be equipped with 6 feet of electrical cord and plug.

3.4 Identification marking. Each warmer shall be permanently and legibly marked with the serial and model number, electrical characteristics, and the name or registered trademark of the manufacturer. In addition, the National stock number and the contract number shall be permanently and legibly marked on the bottom.

3.5 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance and lists of component parts. The manual shall comply with the requirements of MIL-M-7298 MANUAL, TECHNICAL, Commercial Equipment. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual, for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of 15 additional copies shall be at the option of the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:32Z
Check the source to verify that this is the current version before use.

3.6 Operational test (100 percent). An operational check shall be performed on each completed assembled warmer to demonstrate capability of warmer to perform specified function.

3.7 Workmanship. Workmanship shall be first class throughout. The spectacle frame warmer shall be free from defects which detract from its appearance or may impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, inspection level II.

4.3 Examination. The warmers shall be examined visually to determine compliance with requirements of section 3, including dimensions, identification marking and workmanship. The items shall also be examined for conformance to preparation for delivery requirements. Defects shall be classified in accordance with MIL-STD-105 with an AQL of 1.0 for major defects and 4.0 for minor defects.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:32Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Operational test (100 percent). Each warmer shall be given an operational test by the supplier. The test shall consist of connecting to the proper power supply and observation of the warmer for significant functional defects. Additional testing for acceptance shall be conducted as described in 4.4.2, unless otherwise specified.

4.4.2 Performance test. The warmer pan shall be filled with glass beads, approximately 0.021 inch in diameter (NSN 6640-00-299-9818, or equivalent) to within 1 inch of the top. This will be 1-1/2 to 1-3/4 deep. Connect the warmer to a power source and set the dial at 4. Allow 1 hour for warming up and temperature stabilization. A standard spectacle frame front (NSN 6540-00-482 Series) shall be immersed to a depth of 3/4 $\pm$ 1/4 inch. Temperature at this depth shall be 260° $\pm$ 10°F. After 30 $\pm$ 2 seconds the front shall be examined. It should be sufficiently pliable for insertion of lenses and adjustment.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. Each warmer constitutes one unit of issue.

5.1.2 Level A. Each warmer shall be wrapped in kraft paper or inserted into a plastic bag. The wrap or bag shall be secured. The unit shall be further packaged in a box of an appropriate size constructed in accordance with PPP-B-566, (E-flute) or PPP-B-636, type CF, class domestic. The box shall be designed with suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of or damage to the contents. Closure of the box shall be secured to prevent spilling of contents.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Twelve unit packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification. Special requirements as specified in PPP-B-636 shall apply for fiberboard boxes.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:32Z
Check the source to verify that this is the current version before use.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraph of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed in accordance with 30.4 of PPP-B-636.

5.2.2 Level B. Twelve unit packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method I as specified in PPP-B-636. Special requirements as specified in PPP-B-636 shall apply for fiberboard boxes.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2.4 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.3 Marking.

5.3.1 Unit packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The warmer is used to heat and soften plastic spectacle frames for glazing, aligning and adjusting.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:32Z
Check the source to verify that this is the current version before use.

6.3 The warmer covered by this specification is listed in the Federal Supply Catalog under the following stock number:

National Stock Number	Item Identification
6540-00-299-9590	WARMER, PLASTIC SPECTACLE FRAME, 115 Volt, 60 Hz, AC

6.4 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6540-1036

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:32Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS. The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers' vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s), or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-W-36401A

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:32Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

Commander
Defense Personal Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:32Z
Check the source to verify that this is the current version before use.