FED. SUP CLASS
6640



PART NUMBER	NOMINAL DIMENSIONAL DATA (mm)			CAPACITY ML
	TOP ID BOWL	STEM LG	STEM OD	
MS36577-1	—	CANCELED		—
MS36577-2	50	40	14	20
MS36577-3	76	50	11	75
MS36577-4	—	CANCELED		—
MS36577-5	101	76	19	190
MS36577-6	152	89	19	600
MS36577-7	184	70	25	720
MS36577-8	254	140	40	3190

MATERIAL: Conventional polyethylene in accordance with Federal Specification L-P-390, Natural Color.

FINISH: Funnels shall be translucent or opaque with smooth surfaces that are free of bubbles and blisters.

DESIGN: The inner wall of the bowl shall diverge from the stem at an angle of $30^{o} \pm 2^{o}$. External ribbing may be included to eliminate vapor or air lock. The stem shall have an oblique termination. A tab with one hole may be provided on the top rim for hanging storage.

NOTES:
1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.
2. Referenced documents shall be of the issue in effect on date of invitations for bid.
3. FSCM: 05178, 11134, and 25578.

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
AIR FORCE - 82

APPROVED 4 June 1969 . REVISED (A) 3 FEB 1984

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

(A) ENTIRE STANDARD REVISED

P.A.	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
EA		FUNNELS, COMMON, LABORATORY	MS36577
Other Cust MS 99			
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1216

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:43Z
Check the source to verify that this is the current version before use.