FED. SUP CLASS
6630

PART NUMBER	DESCRIPTION	ELECTRODE LENGTH (CM)	CONCENTRATION RANGE	pH RANGE	TEMP RANGE °C	TEST FOR
MS36584-1	Glass	13	1 to 6 x 10^{-7}M	3 - 12	0-80	NH_4^+
MS36584-2	Solid State	14	1 to 5 x 10^{-6}M	0 - 14	0-80	Br^-
MS36584-3	PVC	14	1 to 10^{-5}M	4 - 9	0-40	Ca^{++}
MS36584-4	Solid State	14	1 to 5 x 10^{-5}M	0 - 14	0-80	Cl^-
MS36584-5	Solid State	14	10^{-2} to 10^{-6}M	0 - 14	0-80	CN^-
MS36584-6	Solid State	14	Saturated to 10^{-6}M	0 - 11	0-80	F^-
MS36584-7	Solid State	14	1 to 5 x 10^{-8}M	0 - 14	0-80	I^-
MS36584-8	PVC	14	1 to 10^{-6}M	2.5 - 11	0-40	No_3^-
MS36584-9	PVC	14	1 to 10^{-6}M	2 - 12	0-40	K^3
MS36584-10	Glass	13	1 to 10^{-6}M	0 - 14	0-80	K^3
MS36584-11	Solid State	14	1 to 10^{-7}M	0 - 14	0-80	AG^+
MS36584-12	Glass	13	Saturated to 10^{-6}M	3 - 12	0-80	AG^+
MS36584-13	Glass	13	Saturated to 10^{-6}M	3 - 12	0-80	Na^+
MS36584-14	Solid State	14	1 to 10^{-7}M	0 - 14	0-80	$S^=$

MATERIALS: Electrodes shall be made with internal elements and external materials to meet the requirements for temperature and resistance to chemical attack

DESIGN Electrodes shall be either glass, solid state, or PVC/liquid ion exchange types as specified in the table. The leads may be detachable or permanently attached and shall be a minimum of 76 cm in length. The electrode plugs shall be designed so that they are readily adaptable for use with most pH meters.

PURPOSE. For selective ion determination measurements

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein

USER ACTIVITIES ARMY - CE NAVY - MC

REVIEWER ACTIVITIES ARMY - MD

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document

P.A. EA | INTERNATIONAL INTEREST | TITLE ELECTRODE, HYDROGEN METER (ION SELECTIVE TYPE)

Other Cust MS 99

Procurement Specifications | SUPERSEDES MS36465

MILITARY STANDARD

MS36584

SHEET 1 OF 1

PROJECT 6630-0

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE STANDARD DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:43Z
Check the source to verify that this is the current version before use.