MILITARY SPECIFICATION

CASE, MEDICAL INSTRUMENT AND SUPPLY SET,
Nylon, Non-Rigid, No. 3

This amendment forms a part of Military Specification MIL-C-36856, dated 26 September 1972, and approved for use by all Departments and Agencies of the Department of Defense.

Page 3

2.2 Other publications. Delete, "200 Madison Avenue" and substitute, "343 Lexington Avenue."

Page 6

3.2.15 Snaphooks - delete and substitute, "Snaphooks shall conform to MIL-S-43770/1, Material 1."

Add: "3.2.16 Stitching sealents. Exposed stitching in the cover flap and all exposed stitching in the vertical surfaces of the case not protected by the secured cover flap, shall be sealed by the application of one of the sealents listed in, 6.7 Stitching sealants or equal in accordance with the manufacturers directions."

Page 11

"4.4 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129."

5. PREPARATION FOR DELIVERY. Delete and substitute:

"5. PACKAGING

"5.1 Preservation. Preservation shall be level A or C, as specified (see 6.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:27Z
Check the source to verify that this is the current version before use.

"5.1.1 Level A.

"5.1.1.1 Unit package. Each case shall be clean, dry and packaged in accordance with MIL-P-116, method III.

"5.1.2 Level C. The level C preservation for instrument cases shall conform to the MIL-STD-794 requirements for this level.

"5.2 Packing. Packing shall be level A, B or C, as specified (see 6.1).

"5.2.1 Level A. Forty-eight unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type PPP-B-621, class 2 or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

"5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

"5.2.2 Level B. Forty-eight unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type PPP-B-621, class 1 or PPP-B-636, class domestic weather-resistant. Closure shall be as specified in the applicable box specification.

"5.2.3 Level C. The level C packing for instrument cases shall conform to the MIL-STD-794 requirements for this level.

"5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

"5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 40 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in height, 48 inches in width, and 40 inches in length. Pallet shall conform to Type IV or Type V.

"5.2.5.1 Levels A and B. Instrument cases packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:27Z
Check the source to verify that this is the current version before use.

"5.2.5.2 Level C. Instrument cases packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

"5.3 Marking.

"5.3.1 Levels A, B, and C. Each unit pack, exterior container and unitized loads shall be marked as specified in MIL-STD-129.

"5.4 General.

"5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

"5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.4".

Page 13

6.5 Delete "6545-912-9870"; add "6545-00-912-9870".

Add, "6.7 Stitching sealants - The two sealents referenced in 3.2.15 are:

I. Bostik No. 7058
Bostik Chemical Group
USM Corporation
Middleton, MA 01949

II. K-Kote Seam Sealer
Keyon Consumer Products
200 Main Street
Kenyon, RI 02836

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6545-0081
Location: MIL-C-36856,
NCR, disk 5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:27Z
Check the source to verify that this is the current version before use.