MILITARY SPECIFICATION

CATHETER, URETHRAL, SILICONIZED RUBBER,
DISPOSABLE, FOLEY, ROUND TIP, VALVE TYPE

This Amendment forms a part of Military Specification MIL-C-36891, dated 16 October 1972, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 3

3.1.2 Valve. In second sentence, delete "Spring" and substitute "Flexible member."

3.2 Style, design and dimensions. Delete third sentence. In seventh sentence after "DPSC Drawing 24268", insert "(for Style 1 or Style 2 design),". In Table I, delete tolerance of "+1 Fr, -0 Fr" for Sizes A and B and substitute "± 1 Fr".

Page 8

4.3 Examination. In Table IV, delete Major Category 111, and renumber 112 through 116 to 111 through 115, respectively.

Page 12

5.1.3.2 Unit package. Delete the last sentence and substitute: "In addition, each unit package shall be wrapped in waterproof barrier material and the barrier sealed.

5.2.1.1 Exterior container. Delete the last sentence and substitute: "Fiberboard boxes shall conform to the special requirements specified in Table II of PPP-B-636."

5.2.2.1 Exterior container. Delete the last sentence and substitute: "Fiberboard boxes shall conform to the special requirements specified in Table I of PPP-B-636."

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency

Project No. 6515-1788

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.