MIL-C-36891
16 October 1972

MILITARY SPECIFICATION

CATHETER, URETHRAL, SILICONIZED RUBBER,

DISPOSABLE, FOLEY, ROUND TIP, VALVE TYPE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers sterile, disposable, rubber, urethral catheters, of the Foley design, with balloon and self-sealing inflation system.

1.2 Classification. Catheters shall be furnished in the following types and sizes, as indicated below (see 6.1):

Types

I - Catheter with 5 cc balloon
II - Catheter with 30 cc balloon

Sizes

A - 16 French
B - 18 French
C - 22 French
D - 24 French

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein. These documents may be obtained as directed by the contracting officer.

SPECIFICATIONS

Federal

PPP-A-250	Adhesive, Water-Resistant, (For sealing Fiberboard Boxes).
PPP-B-585	Box, Wood, Wirebound.
PPP-B-601	Box, Wood, Cleated-Plywood.
PPP-B-621	Box, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing).

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

Military

MIL-P-116	Preservation, Methods of.
MIL-S-36586	Sterilization Strip Test, Bacterial Spore.

STANDARDS

Federal

Fed-Std-601	Rubber; Sampling and Testing.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)
24268 CATHETER, URETHRAL, RUBBER, FOLEY

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

THE UNITED STATES PHARMACOPEIAL CONVENTION, INC.

The Pharmacopeia of the United States of America (U.S.P.)

(Application for copies shall be addressed to Mack Publishing Company, Easton, Pennsylvania 18042).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material

3.1.1 Catheter proper. The catheters shall be fabricated from durable, soft, natural latex rubber. The entire lumen and external surface of the catheter shall be bonded to the rubber, with layer of suitable silicone rubber. No plasticizers or residual vulcanizing agents shall be permitted in the silicone rubber formulation. The finished catheters shall be sterile, pyrogen-free, and shall pass the animal implantation test as specified herein.

3.1.2 Valve. Valve shall be fabricated from suitable, rigid plastic compound for housing and plunger. Spring and sealing member shall be fabricated from safe, corrosion-resisting materials, suitable in all respects for the purpose intended.

3.2 Style, design and dimensions. The catheters shall be fabricated from materials specified in 3.1. Catheters shall be sterile, disposable, of the Foley design, possessing a reinforced hollow round tip at the distal end, and two eyes. Eyes shall be oriented as shown on drawing. The catheters shall be constructed with an integral self-retaining balloon of the capacity indicated in Table I for appropriate type and size. The balloon shall be inflated by means of a self-sealing inflation valve. The Foley catheters shall be suitable for use in the treatment of urethral strictures and urinary diversion. In addition to the dimensions and physical properties specified in Table I, the catheters shall be in accordance with DPSC Drawing 24268, for appropriate type, size, stock number and item identification.

Table I. Dimensions and Physical Properties

Size	*French size (Fr)	Balloon capacity cubic centimeters (cc) Type	Inside diameter-minimum (inches) (measured at any point)	Breaking strength minimum-pounds
A	16 + 1 Fr - 0 Fr	I	0.084	58
B	18 + 1 Fr - 0 Fr	I	0.098	70
C	22 ± 1 Fr	II	0.140	94
D	24 ± 1 Fr	II	0.171	106

* The diameter of the balloon section of the catheter shall not exceed the diameter of the catheter by more than 1 millimeter (3 French equals 1 millimeter).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

3.2.1 Inflation system. The inflation system shall consist of a self-sealing valve, an inflation channel, and a balloon. The balloon shall be constructed integral to the outer wall of the catheter with the inflation channel imbedded or permanently attached within the wall. A self-sealing valve shall be securely attached to the proximal end of the inflation segment. The balloon shall be capable of inflation and deflation by penetration of a Luer lock or Luer slip syringe (without use of a needle) into the valve inlet. The inflation and deflation of the balloon shall be accomplished in a simple and leakproof manner. The valve shall seal immediately upon removal of the syringe after balloon inflation. The valve shall be rigidly constructed, self-contained, and shall not be capable of opening, after balloon inflation, when a patient is lying on the valve.

3.3 Fluid and pneumatic tightness of inflation system. The catheters shall be capable of withstanding the fluid tightness test and pneumatic tightness test as specified in 4.4.5 and 4.4.6, respectively. There shall be no leakage between the syringe tip and the valve inlet when the balloon is tested as specified in 4.4.5 and 4.4.6. The valve shall not leak or become detached during the tests. There shall be no leakage at any point in the inflation channel, including the balloon. The balloon shall not burst or collapse, and shall not cover the catheter eyes, nor cause excessive buckling or collapsing of the tip or drainage lumen during the tests. Catheters that fail these tests in all respects, shall be rejected.

3.4 Physical properties. When tested as specified in 4.4.4, the ultimate elongation (before aging) shall be 750 percent minimum and the breaking strength shall be as specified in Table I (see 3.2). In addition, when the catheters are subjected to the accelerated aging test as specified in 4.4.4, the average ultimate elongation and average breaking strength shall not decrease by more than 25 percent, nor shall the catheters crack, become tacky, or brittle.

3.5 Identification marking. Each catheter shall clearly and legibly bear the manufacturer's name or registered trademark, balloon size in cc, and French size at the proximal end.

3.6 Sterility. The catheters shall be sterile when tested with a biological indicator as specified in 4.4.2.

3.6.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have be tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. Federal stock number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.

m. Signature of head of laboratory or a designated qualified alternate.

When inspection is made at destination, the certificate of sterility shall be sumbitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pa., 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.6.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.7 Animal implantation test. The outer surface of the finished catheter only shall be tested, and shall not show evidence of toxicity or adverse tissue reaction upon gross and microscopic examination, when tested as specified in 4.4.1. Tissue reaction shall not exceed that defined as a "mild tissue reaction".

3.8 Certificates of freedom from pyrogens and animal implantation test. A certificate listing all lot or control numbers shall state that representative samples from each lot or control have been tested in accordance with methods described herein, and found to be free from pyrogens, and have passed the animal implantation test. This certificate shall include the specific tests performed and the test results. The record of the animal implantation test and its results shall be certified by the pathologist who performed or supervised performance of the animal implantation test.

These certificates shall be submitted by the contractor in the manner as specified in 3.6.1.

3.9 Manufacturer's record. The manufacturer shall place a code or lot number on the immediate container or unit package which can be verified against the manufacturer's records of all tests made on sterile, pyrogen-free units. These records shall be made available for inspection by any duly qualified medical officer of the Armed Forces, National Institutes of Health, or any official of the Federal Food and Drug Administration. No shipment shall be made until all the tests have been completed, verified, and recorded in the permanent records of the manufacturer.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

3.10 Workmanship. Workmanship shall be first class throughout. The catheters shall be free from defects which detract from their appearance or impair their serviceability. The catheters shall show a smooth finished outer surface free from pits, flash, scale, cracks, holes, and inner surface free from scale, deep pits, deep furrows, or die marks; the lumen of the catheters shall be free from burrs, chips, dirt, flash, or other foreign matter.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of all rubber and plastics used in the construction of the catheter, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in Table II. The unit of product for examination purposes shall be one catheter.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:31Z
Check the source to verify that this is the current version before use.

Table II. Sampling Criteria for Examination

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	None defined	None defined

4.2.2 For tests.

4.2.2.1 For sterility test. Sampling for sterility test shall be in accordance with the U.S.P. for parenteral devices and as specified in 4.4.2.

4.2.2.2 For pyrogen tests. Sampling for pyrogen tests shall be accomplished by selecting samples which are to be representative of each lot (not less than two catheters) of material used in the fabrication of the catheters.

4.2.2.3 For other tests. Sampling for other tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table III. The unit of product for testing purposes shall be one catheter. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics. Defective catheters shall be rejected.

Table III. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Catheter	Dimensions	3.2	---	S-1	1.0
	Fluid tightness	3.3	4.4.5	S-1	0.65
	Pneumatic tightness test	3.3	4.4.6	S-1	0.65
	Inside diameter	3.2	4.4.4	*	**
	Ultimate elongation	3.4	4.4.4	*	**
	Breaking strength	3.4	4.4.4	*	**
	Accelerated aging	3.4	4.4.4	*	**
	Sterility test	3.6	4.4.2	-----	**
	Animal implantation test	3.7	4.4.1	-----	**
	Pyrogen test	3.1.1	4.4.3	-----	**

* A minimum of 3 samples shall be tested.

** No defects permitted.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:31Z
Check the source to verify that this is the current version before use.

4.3 Examination. The catheters shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Table II and IV. Examination, however, shall not be restricted to defects listed in Table IV. Any units in the sample containing one or more defects shall be rejected and if the number of defective units in any sample exceeds the acceptance number for the sample, the lot represented by the sample shall be rejected.

Table IV. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Catheter not of the proper size for appropriate stock number and item identification.
102	Catheter not free from kinks, dents, blisters, cuts, holes, and scratches.
103	Catheter lumen not free from foreign matter.
104	Catheter outer surface not smoothly finished.
105	Balloon does not inflate when a Luer lock or Luer slip syringe is activated into valve inlet.
106	Catheter does not reseal when valve inlet is inflated with syringe.
107	Valve not permanently secured to inflation segment.
108	Balloon does not deflate by means of Luer lock or Luer slip syringe insertion after balloon has been inflated.
109	Catheters not symmetrical in shape.
110	Catheter does not possess two eyes.
111	Eyes not in proper plane with inflation channel.
112	Eyes not complete, not well defined, and not free from flash.
113	Catheters not identified as specified in 3.7.
114	Immediate container markings and legend not as specified.
115	Color of catheters not uniform.
116	Catheter coiled in immediate container.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:31Z
Check the source to verify that this is the current version before use.

4.4.1 Animal implantation test. Samples of the finished catheters (end item) shall be tested in accordance with the U.S.P. for Implantation Test, Plastic Containers, for Class VI plastics, except after "72 hours" add "plus or minus 4 hours." Examine macroscopically the area of the tissue surrounding the center protion of each implant strip and rate and record it according to the following procedure:

No tissue reaction -	reaction not distinguishable.
Trace tissue reaction-	film around the implant consists of a precipitated serum protein. Not measurable in thickness.
Slight tissue reaction-	film around the implant in form of capsule 0.5 millimeters (mm) in thickness maximum.
Mild tissue reaction-	film around the implant in form of capsule 0.5 mm to 1.0 mm in thickness.
Moderate tissue reaction-	film around the implant in form of capsule greater than 1.0 mm in thickness. Slight inflammation around implant site.

The testing shall be performed by a pathologist certified by the American Board of Pathology or the American College of Veterinary Pathologists.

4.4.2 Sterility test. The catheters shall be tested for sterility in accordance with the method specified in the USP for parenteral devices using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.4.2.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for Type I sterilization and 30 for Type II Sterilization including inoculated carrier and product test units.

4.4.2.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C
Dry Heat	B. stearothermophilus, or	55° - 65°C
	B. subtilis var. niger (or globigii)	30° - 35°C
Gas	B. subtilis var. niger (or globigii)	30° - 35°C
Radiation	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

4.4.2.2 Inoculated carrier. Inoculated carriers (paper spore strips), if used, shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.2 Sterility test and 4.4.2.1 Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.2.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

4.4.2.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer , and approval obtained before employment of the alternate method on a government contract.

4.4.2.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the U.S.P.

4.4.3 Pyrogen test. Pyrogen test shall be performed by selecting representative samples and cutting the catheters into small portions so that no two dimensions exceed 3 mm and place the pieces (total weight 10 grams) into a clean flask. Add 200 milliliters of physiological saline solution. Heat the solution to 85 degrees ± 1 degree Centigrade for 1 hour (± 5 minutes) with occasional agitation. Decant, cool, and test the saline eluate.

4.4.4 Physical properties test. The tests shall be conducted in accordance with the following test methods of Federal Test Method Standard No. 601:

Test	Test Method No.
Inside diameter	2211
Ultimate elongation	4121*
Breaking strength	4111*
Accelerated Aging	7221*

* A straight specimen shall be used. To determine breaking strength, the value defined as "breaking force" shall be used as the value for breaking strength.

4.4.5 Fluid tightness of inflation system. This test shall be performed on samples that have been subjected to the accelerated aging test specified in 4.4.4.

The catheter balloon shall be slowly inflated with water, (volume as specified in Table V) through the inflation valve using a syringe that has a standard Luer lock or Luer slip tip. The syringe shall be removed, and the balloon shall remain inflated for at least 4 hours. The entire inflation system shall be observed for leakage. The balloon shall also be observed for all characteristics specified in 3.3.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

Table V. Water Volume

Type	Volume of water (cc)
I	5
II	30

In addition, Type I catheters shall be further tested, as described above, by inflating them with 10 cc of water through the inflation valve and the balloon observed for rupture and balloon retention.

4.4.6 Pneumatic tightness of inflation system. After completion of the fluid tightness test specified in 4.4.5, the balloons shall be tested for pneumatic tightness of balloon inflation system by injecting 10 cc of air through the inflation valve of Type I catheters and 30 cc of air with Type II catheters. The catheters shall be observed for balloon inflation and leakage when the catheters are submerged under water for a minimum period of 30 minutes with the above appropriate volume of air applied to the inflation system.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One box containing twelve catheters, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table VI. shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table VI. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not Required	12 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Individual package. Each catheter shall be packaged uncoiled in a sealed, peel-type and tamperproof container which shall maintain sterility of contents unless damaged or opened.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

5.1.3.2 Unit package. Twelve individual packaged catheters shall be packaged in a box of appropriate size constructed in accordance with PPP-B-566, PPP-B-676 or PPP-B-636, type CF, class domestic. Commercial colors are acceptable on unit packages and the closure shall be adequate to prevent spilling of contents under normal handling. In addition each unit shall be shrink film wrapped.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification. Special requirements of applicable table shall apply for PPP-B-636 box.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed in accordance with [illegible] of PPP-B-636.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United Stated for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard box shall be closed by method II of PPP-B-636. Special requirements of applicable table shall apply for PPP-B-636 box.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Individual package. In addition to any commercial markings, each individual package shall be marked with item name, type, French size, balloon size, instructions for use, lot number, manufacturer's name and the legend or similar: "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED". Marking shall be printed directly on the container or on a label inserted within the container in such a manner as to permit ready identification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

5.3.2 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. Lot (control) number, contract or purchase order and name of contractor shall be shown. The method of sterilization (G-Gas, R-Radiation, S-Steam, D-Dry Heat Sterilization) shall be shown following the level of packaging and date of packaging. Example: "A - 11/71 - G". Final sterilization shall be on or after indicated packaging date.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Lot (control) number and method of sterilization as specified in 5.3.2 shall be shown.

6. NOTES

6.1 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Type and size required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification covers the following items appearing in the Federal Supply Catalog:

Federal Stock No.	Item Identification
6515-104-8694	CATHETER, URETHRAL, Siliconized Rubber, Disposable, Round Tip, Foley, with 5 cc Balloon, Valve Type, 16 Fr, 12s.
6515-104-8695	CATHETER, URETHRAL, Siliconized Rubber, Disposable, Round Tip, Foley, with 5 cc Balloon, Valve Type, 18 Fr, 12s.
6515-104-8696	CATHETER, URETHRAL, Siliconized Rubber, Disposable, Round Tip, Foley, with 30 cc Balloon, Valve Type, 22 Fr, 12s.
6515-104-9006	CATHETER, URETHRAL, Siliconized Rubber, Disposable, Round Tip, Foley, with 30 cc Balloon, Valve Type, 24 Fr, 12s.

It does not include all types or sizes as indicated by the title nor all types which are commercially available.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1666

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-C-36891

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

Defense Supply Agency

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.