MILITARY SPECIFICATION

HANDPIECE, DENTAL, CONTRA-ANGLE, WITH FRICTION GRIP CHUCK, ROUND SLEEVE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a conventional, contra-angle, round sleeve, dental handpiece with friction grip chuck for attachment and use with a belt or air driven straight handpiece with a round nose.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Setup.

Military

MIL-P-116	Methods of Preservation.
MIL-M-7298	Manual, Technical Commercial Equipment.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-vaporproof or Waterproof, Flexible.

STANDARDS

Federal

FED-STD-151	Metal, Test Methods.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

24298	Handpiece, Dental, Contra-Angle, With Friction Grip Chuck, Round Sleeve.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM Designation E 18	Tests for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Material.
ASTM Designation E 92	Tests for Vickers Hardness of Matallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, Pa. 19103).

3. REQUIREMENTS

3.1 Drawings. The drawings forming a part of this specification are specification control drawings. The supplier is responsible for preparing his own shop drawings. Where tolerances prescribed could cumulatively result in incorrect fits, the supplier shall provide tolerances within those prescribed on the drawings to insure correct fit, assembly, and operation of the item. No deviation from the prescribed dimensions or tolerances is permissible without prior approval from the contracting officer.

3.2 Materials. Materials shall be new and shall conform to the requirements specified herein. Materials not specifically designated in table I shall be of high commercial quality throughout.

Source: https://assist.dla.mil -- Downloaded: 20260-21T09:31Z
Check the source to verify that this is the current version before use.

Table I Materials, Hardness and Platings

Part	Component	Material	Rockwell Hardness	Plating
Spindle assembly	Spindle	AISI 416	C26-36	-
	Upper bearing	Phosphor bronze #544	-	-
	Lower bearing	Phosphor bronze #544	-	-
	Gear	AISI 4140	C40-50	Electrolize
	Washer	Shim stock	-	-
	Slinger	AISI 416	C26-36	-
Head and neck assembly	Collet	AISI 17-4 PH	C26-36	-
	Drive screw	AISI 17-4-PH	C26-36	-
	Head and neck shell	ASTM B-16-brass	-	Ni.-Cr.
	Head cap	AISI 416	C26-36	-
	Head to elbow bearing	AISI 12L14	-	-
	Head lock nut	AISI 12L14	-	-
	Pinion gear	AISI 4140	C40-50	Electrolize
	Elbow gear	AISI 4140	C40-50	Electrolize
Elbow and rear end assembly	Elbow	ASTM B-16 brass	-	Ni.-Cr.
	Elbow gear	AISI 4140	C40-50	Electrolize
	Rear end housing	ASTM B-16 brass	-	Ni.-Cr.
	Rear end shaft	AISI 4140	C30-40	-
Chuck tool assembly	Wrench handle	AISI 416	-	-
	Wrench spanner	AISI 416	C25 min.	-
	Wrench square	AISI 303	-	-
Wrench	Wrench	AISI 1010, 1/2 HD	-	Ni.

3.3 Design and construction. The handpiece shall be a conventional contra-angle, round sleeve, dental handpiece with friction grip chuck suitable to receive friction grip burs, diamond points and other rotary instruments for the performance of dental procedures. Shall be suitable for attachment and use with belt or air driven, round nose straight handpieces. The handpiece shall consist essentially of a spindle assembly, a head and neck assembly, and an elbow and rear end assembly. A chuck tool assembly and two wrenches shall be furnished with each handpiece. Materials, hardnesses and platings shall be as indicated in table I. The handpiece shall be a durable, precision made dental handpiece entirely suitable for operation at handpiece speeds up to 15,000 revolutions per minute (r.p.m). The handpiece shall be suitable for attachment and use with Ritter Company's "Torq-Master" Round Nose Straight Handpiece (FSN 6520-181-8332), Midwest

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

American "Tru-Torc" Round Nose Straight Handpiece (FSN 6520-874-1995), and S. S. White Company's "No. 14" Round Nose Straight handpiece (FSN 6520-584-3463).

The handpiece shall conform to Defense Personnel Support Center (DPSC) Drawing No. 24298 for design, dimensions and construction.

3.3.1 Spindle assembly. The spindle assembly shall consist of a spindle, upper and lower bearings, gear, washer and slinger. The upper bearing shall be soldered on the spindle and the lower bearing shall slip free on the spindle shaft. The gear shall be soldered on the spindle. The washer acts as a spacer between the gear and the lower bearing. The slinger shall be pressed into place on the operative end of the spindle. The lower bearing shall be threaded, thus permitting the spindle assembly to screw into and fit securely in the head of the handpiece.

3.3.2 Head and neck assembly. The exterior portion of the head and neck assembly shall consist of the head and neck shell, and the head cap. The head shall contain the spindle assembly (3.3.1), and the collet and drive screw. The collet shall securely hold friction grip angle handpiece burs with shank diameter of 0.0624 to 0.0631 inch. The neck shall house the head to elbow bearing. The pinion gear on one end of the head to elbow bearing shall mesh with the gear on the spindle of the spindle assembly. The elbow gear on the other end of the head to elbow bearing shall mesh with the elbow gear of the elbow and rear end assembly. The head to elbow bearing shall be threaded near the elbow gear end, and shall screw into the elbow end of the elbow and rear end assembly. The head lock nut shall secure the head and neck assembly to the elbow and rear end assembly.

3.3.3 Elbow and rear end assembly. The elbow and rear end housing assembly shall consist of the elbow and the rear end housing. The elbow gear on one end of the rear end housing shall mesh with the elbow gear of the head and neck assembly. The other end of the rear end housing shall contain the drive shaft. The rear end housing shall be threaded near the elbow gear end, and shall screw into the elbow to form the elbow and rear end assembly. The rear end housing shall be knurled on the outside to provide a gripping surface, and the interior surface portion of the round sleeve shall fit the round nose portion of the nosepieces on the straight handpieces (see 3.3). The drive shaft shall be of such diameter that it will fit into the chuck of the straight handpieces.

3.3.4 Chuck tool assembly. The chuck tool assembly shall consist of a wrench handle, wrench spanner and wrench square. The spanner on one end of the assembly is designed to remove and insert the drive screw in the head of the handpiece. The handle and square on the other end of the assembly is designed to remove the collet from the head.

Source: https://assist.dla.mil -- Downloaded: 20261O-21T09:31Z
Check the source to verify that this is the current version before use.

3.3.5 Wrench. Two wrenches of the same size and construction shall be furnished. The center hole of the wrench is designed to fit over the rear end housing, and into the T-slot of the rear end housing, to obtain a firm grip on the handpiece while using the other wrench to disconnect or connect parts. The large U-shape opening at one end of the wrench is used to disconnect or connect the elbow to the rear end housing, or to disconnect or connect the head lock nut to the head and neck assembly. The small U-shape opening at the other end of the wrench is used to disconnect or connect the head to elbow bearing to the elbow.

3.4 Hardness. Hardness of components shall be in accordance with Table I. Wherever a value is not specified, the hardness shall be suitable for the material used and purpose intended.

3.5 Plating. Components shall be plated as specified in Table I for the applicable components. Plating shall be uniform in thickness and appearance. Plating shall be free from blisters, visible porosity, or other defects.

3.6 Runout. The total runout of the angle handpiece shall not exceed 0.0010 inch, when tested as specified in 4.4.3.

3.7 Vibration and heat rise. The handpieces shall pass the test for vibration and heat rise specified in 4.4.4.

3.8 Drive train smoothness and backlash. The handpiece shall pass the test for drive train smoothness and backlash when tested as specified in 4.4.5.

3.9 Instructions. Instructions shall be furnished with each unit of issue for cleaning, lubricating and use of handpiece.

3.10 Interchangeability. The following parts shall be interchangeable and replaceable as either a component part or as an assembly (where stated):

a. Contra angle with metal chuck, less rear end.

b. Head and gear barrel assembly.

c. Spindle assembly.

d. Elbow and rear end assembly.

e. Rear end assembly.

f. Elbow assembly.

g. Heat to elbow bearing assembly.

h. Pinion gear.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

i. Heat nut.

j. Drive screw.

k. Collet.

l. Wrench.

m. Chuck tool.

3.11 Finish. Exposed surfaces of brass components shall be chromium over nickel plated. Exposed surfaces shall have a mirror or satin finish except for knurled portions and lower bronze bearing of spindle assembly.

3.12 Identification Marking. Each handpiece shall be permanently and legibly marked with the name or registered trademark of the manufacturer. Handpiece of foreign (other than U.S.A.) manufacture shall be marked with the name of the country of origin.

3.13 Service Data. The contractor shall furnish with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance, and lists of component parts. The manual shall comply with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of 15 additional copies shall be at the option of the Government.

3.14 Workmanship. Workmanship shall be first class throughout. Handpieces shall be free from defects which detract from their appearance or which may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspection are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials used in the handpieces, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered are classified as specified in Tables II and IV (see 4.2.1 and 4.3) and Table III (see 4.2.2).

4.2.1 Sampling for examination. Sampling for examination for the handpiece shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in Table II. The unit of product for examination purposes shall be one handpiece.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:31Z
Check the source to verify that this is the current version before use.

Table II. Inspection and Acceptable Quality Levels

Type defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-1	2.5
Minor A	II	2.5
Minor B	S-1	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table III for the handpiece. The acceptance number shall be zero for all sample sizes. The unit of product for test purposes shall be one handpiece. Where feasible, the same sample unit may be used for the determination of two or more test characteristics.

Table III. Sampling criteria for Testing

Component	Characteristic	Requirement	Test procedure	Inspection level
All components in table I	Chemical Analysis	3.2	4.4.1	*
Components indicated in table I	Hardness	3.4	4.4.2	S-1
Assembled handpiece	Runout	3.6	4.4.3	S-1
Assembled handpiece	Vibration and heat rise	3.7	4.4.4	S-1
Assembled handpiece	Drive train smoothness and backlash	3.8	4.4.5	100%

* A sufficient amount from each lot of raw materials used in the construction of the item shall be selected for chemical analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

4.3 Examination. The handpiece shall be examined to determine compliance with requirements contained in this specification. Examination shall be conducted in accordance with the following classification of defects. Examination shall include, but shall not be limited to, the defects listed in table IV.

Table IV. Classification of Defects

Categories	Defects
Critical	None defined
Major A	
101	Drive screw will not lock bur in place.
102	Finish not as specified.
103	Slivers and burs present.
Major B	
151	Dimensions affecting fit of handpiece to round nose straight handpieces not as specified.
152	Dimensions affecting fit of bur shank to bur socket (collet) not as specified.
Minor A	
201	Service data, technical manuals missing.
202	Identification marking missing.
203	Identification marking not permanently applied and not as specified.
204	Pits, knicks and discoloration present on outer surfaces.
Minor B	
251	Dimensions (other than 151 and 152) not as specified.

4.4 Tests.

4.4.1 Chemical analysis. Metal analysis shall be conducted in accordance with method 111 or 112 of FED-STD-151.

4.4.2 Hardness test. Hardness shall be determined in accordance with ASTM Standard E 18 (Rockwell) or E 92 (Diamond Pyramid Hardness).

4.4.3 Runout. The runout shall be measured by inserting a true-running arbor or shank 0.0624 to 0.0631 inch in diameter into the collet. A dial indicator sensitive to 0.0002 inch and equipped with a knife-edge foot, or a shadow-graph or other suitable equipment of equal accuracy shall be employed. If a dial indicator is used, the indicator shall be mounted with the shaft perpendicular to the axis of the arbor or shank, and so positioned that the foot contacts the arbor or shank 1/8 inch from the collet. The maximum and minimum readings obtained during one complete 360-degree rotation of the arbor or shank shall be measured and recorded to the nearest 0.0005 inch. The difference between these readings shall be taken as the total runout.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

4.4.4 Vibration and heat rise. The assembled and lubricated handpiece shall be tested by running the handpiece, containing a dental bur, for three minutes at 10,000 r.p.m. under free load. The handpiece shall be tested for excessive vibration when hand held, and for excessive heat rise (above 115° Fahrenheit) when tested with a pyrometer held against the head housing and against the back end housing.

4.4.5 Drive train smoothness and backlash test. The assembled and lubricated handpiece shall be tested by spinning the back end drive shaft through at least two full revolutions, both clockwise and counterclockwise, using finger pressure only as applied by the thumb and index finger. Any detectable binding, slipping, backlash, or irregular mechanical interference shall cause the handpiece to be rejected.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.1).

5.1.1 Unit of issue. One handpiece, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table V shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table V. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	288 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IC-2.

5.1.3.5 Intermediate package Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-566 or PPP-B-676, except commercial colors will be acceptable, or PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type I load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-F-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Fiberboard boxes shall conform to the special requirements as specified in PPP-B-636.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of wood boxes, shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II of PPP-B-636. In addition, fiberboard boxes shall conform to the special requirements as specified in PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit and intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 Limitations. This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Military requirements.

6.3 Federal Stock Number (FSN) coverage. This specification covers the following item appearing in the Federal Supply Catalog:

Stock Number	Item Identification
6520-469-3503	HANDPIECE, DENTAL, Contra-angle, With Friction Grip Chuck, Round Sleeve

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6520-1331

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-H-36887

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

Defense Supply Agency

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S. MAIL

HEADQUARTERS, DEFENSE PERSONNEL SUPPCRT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:31Z
Check the source to verify that this is the current version before use.