MILITARY SPECIFICATION

PAD, SANITARY, HEAVY

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers an absorbent cellulose filled type pad suitable for perineal use after rectal surgery, for vaginal bleeding, and for obstetrical and gynecological purposes.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

L-C-166	Cellulose, Absorbent, Surgical.
MMM-A-250	Adhesive, Water-resistant, (For Sealing Fiberboard Boxes).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-sensitive Adhesive Paper, Water Resistant, (For Carton Sealing).

STANDARDS

Federal

FED-TEST-METHOD STD NO. 191	Textile Test Methods.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:28Z
Check the source to verify that this is the current version before use.

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publication. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

UNITED STATES PHARMACOPOEIAL CONVENTION, INC.

Pharmacopoeia of the United States

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pennsylvania).

3. REQUIREMENTS

3.1 Material.

3.1.1 Cellulose filler. The cellulose filler shall be made from absorbent surgical cellulose conforming to the requirements and tests of L-C-166 except that, as an alternate, the material may be supplied in the form of a particulate fluff and the absorbency of the filler shall be tested by the procedure for absorbency given on page 906 of the USP XVII. In addition, the cellulose shall be relatively free of any water-soluble coloring matter. The cellulose filler shall be bleached white and shall not contain water-soluble coloring matter, or other foreign matter.

3.1.2 Covering. The covering of the absorbent filler shall be a suitable soft, nonwoven, non-irritant fabric with sufficient porosity to permit the assembled pad to meet the absorbency requirements specified in 3.3. The covering shall be wrapped around the filler and sealed or secured in a manner as to contain the absorbent filler and prevent filler fibers from coming in contact with the wound or sutures. The covering shall extend beyond each end of the absorbent filler to form tabs for securing in use. The breaking strength of each tab shall be not less than 14 pounds when tested as described in 4.3.1. The covering shall not adhere to the skin, sutures, or surgical site. The finished pad shall not disintegrate during application, use, or after sterilization by steam.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:28Z
Check the source to verify that this is the current version before use.

3.2 Construction. Shall be a bleached, absorbent cellulose filler (3.1.1) in a soft non-woven covering (3.1.2). Finished pads shall be 11 inches ± 1/4 inch in length and 3 inches ± 1/8 inch in width. The covering on each pad shall extend a minimum of 4-1/2 inches beyond each end of the cellulose filler to form the required tabs. Minimum weight of one dozen finished pads shall be 7.00 ounces. Pads shall not contain moisture retarding side strips or other moisture retarding elements.

3.3 Absorbency, finished pad. The finished pad shall possess such absorbency that when tested as specified in 4.3.2, it will submerge completely within twelve seconds.

3.4 Capability of autoclaving. Unit packages (12 finished pads) shall be capable of being sterilized in an autoclave without disintegrating. Finished pads shall not disintegrate when tested as specified in 4.3.3.

3.5 Workmanship. The pads and containers shall be free from defects which affect their appearance or which may affect their utility. The pads shall be manufactured under clean and sanitary conditions.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:28Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the cellulose, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables I and II (see 4.2.1 and 4.2.2) and 4.2.3.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in table I. Unit of product for examination purposes shall be as indicated in table I.

Table I. Sampling for examination

Class	Series	Inspection level	Unit of product	AQL
Major A	101-150	II	Pad	1.0
Major B	151-200	S-2	Pad	1.0
Minor A	201-250	II	Pad	2.5
Minor B	251-290	S-2	Ext. cont.	2.5
Dimensional	301-400	S-3	Pad	2.5

4.2.2 Examination. The pads shall be examined for defects including, but not limited to, those listed in table II.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:28Z
Check the source to verify that this is the current version before use.

Table II. Classification of defects

Categories	Defects
Critical	None defined.
Major A	
101	Pad does not contain all component parts.
102	Pad not free of lumps.
103	Pad not free of tears.
104	Pad not free of rough areas.
105	Filler not completely enclosed by covering.
Major B	
151	Filler not free of sharp particles.
152	Filler not relatively free from impurities such as oil, pitch, slime, specks and foreign matter.
153	Filler not free of water-soluble coloring matter.
Minor A	
201	Pad not free of frays.
Minor B	
202	Marking on interior packages and exterior containers not present, complete, correct, legible, or permanent.
Dimensional	
301	Weight not within the tolerance specified.
302	Any dimension not within the tolerance specified.

4.2.3 Sampling for tests. Except for 4.3,3, sampling for tests shall be conducted in accordance with MIL-STD-105, inspection level S-1, except that the acceptance number shall be zero for all sample sizes. Unit of product for test purposes shall be one pad.

4.3 Tests. Tests shall be conducted as necessary to determine compliance with specification requirements.

4.3.1 Tab breaking strength. The breaking strength of the tab shall be determined in accordance with Fed. Test Method Std. No. 191, Test Method 5100 with the following modifications: The test specimen shall be the full width of the tab cut from the pad at the end of the filler material. The design of each clamp shall be such that the width of the face of the jaws is greater than the width of the specimen. Elongation measurements are not required. Test shall be made in the longitudinal direction of the specimen only. The average breaking strength for the sample shall be used as the breaking strength of the tab.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:28Z
Check the source to verify that this is the currentversion before use.

4.3.2 Absorbency-finished pad. The absorbency of the finished pad shall be determined as follows: Hold the pad horizontally by the tab ends, over, and within one half inch of, a container of water at 25° ± 1° C; allow finished pad to drop lightly upon the water. Complete submersion takes place within the time specified in 3.3.

4.3.3 Autoclaving. One unit package from each lot shall be subjected to autoclaving at 250° Fahrenheit and 15 pounds pressure for not less than 20 minutes. Unit package shall be opened and finished pads examined.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be Level A or C as specified (see 6.1).

5.1.1 Unit of issue. One package (12 pads), as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table III shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table III. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	24 units	48 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in a bag fabricated of kraft paper of not less than 30 pounds basis weight. Bag with contents shall be capable of withstanding sterilization. Closure shall be adequate to prevent accidental opening.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:28Z
Check the source to verify that this is the current version before use.

5.1.3.2 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.1).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification. Special requirements of applicable table shall apply for PPP-B-636 box.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Fiberboard box shall be closed by method II of PPP-B-636 box. Special requirements of applicable table shall apply for PPP-B-636 box.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown. In addition, each unit package shall be marked with the legend: "NOT STERILE".

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:28Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data.

(a) Title, number, and date of this specification.
(b) Selection of the applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not include all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or which are normally commercially available, but is intended to cover the types which are normally procured to meet Military requirements.

6.3 This specification covers the following item appearing in the Federal Supply Catalog:

FSN	Item Identification
6510-559-6130	PAD, SANITARY, Heavy, 12s

6.4 Supersession data. This specification includes the requirements of Purchase Description No. 7, dated 24 March 1970 and Modification No. 3, dated 11 September 1970.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6510-0579

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:28Z
Check the source to verify that this is the currentversion before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-P-36863

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:28Z
Check the source to verify that this is the current version before use.

FOLD

DEFE SE PERSONNEL SUPPOR
:800 South 20th St
. !ladelphia, Pa.

Defense Supply Agenc

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

'' OQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
'.: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
00 SOUTH 20TH STREET
HILADELPHIA, PA. 19[illegible]

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:28Z
Check the source to verify that this is the current version before use.