MILITARY SPECIFICATION

TABLE, AUTOPSY

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a lightweight, portable autopsy table for use in the field.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

V-T-285	Thread, Polyester
V-T-295	Thread, Nylon.
QQ-A-200/8	Aluminum Alloy Bar, Rod, Shapes, Tube and Wire, Extruded and Structural Shapes, 6061.
QQ-A-250/2	Aluminum Alloy 3003, Plate and Sheet.
QQ-A-250/11	Aluminum Alloy 6061, Plate and Sheet.
QQ-A-371	Aluminum Alloy Ingot (For Remelting).
PPP-B-601	Boxes, Wood, Cleated-Plywood.

Military

MIL-W-4088	Webbing, Textile, Woven Nylon.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Federal

FED-STD-751	Stitch, Seam and Stitching.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:30Z
Check the source to verify that this is the current version before use.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

25064 Autopsy Table, Lightweight, Field, Portable.

(Copies of specifications, standards and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Material shall be as specified in DPSC Drawing 25064, the Bill of Material thereto, and as specified herein.

3.1.1 Aluminum.

3.1.1.1 Sheet. Aluminum sheet of alloy 6061-T6 shall conform to QQ-A-250/11. Alloy 3003-H14 sheeting shall conform to QQ-A-250/2.

3.1.1.2 Tubing, rod and angle. Aluminum alloy 6061-T6 shall be used for tubing, rod and angles, and shall conform to QQ-A-200/8.

3.1.1.3 Castings. Aluminum castings shall be alloy A356-T51 in conformance with QQ-A-371.

3.1.2 Corrosion-resisting steel. Corrosion-resisting steel shall be AISI 302, 303 and 304.

3.1.3 Fabric.

3.1.3.1 Thread. Thread shall be nylon in conformance with V-T-295, or polyester in conformance with V-T-285. In either case, thread shall be Type I, class 1, Size FF, O.D. Shade S-1, Cable # 66022.

3.1.3.2 Nylon. Hypalon-coated nylon, 12.5 oz., O.D. # 7 shall be used.

3.1.3.3 Webbing. Webbing shall be nylon, Type XXI of MIL-W-4088.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:30Z
Check the source to verify that this is the current version before use.

3.2 Construction. The autopsy table shall be in accordance with DPSC Drawing 25064 and as specified herein, except that the Lamp and Gooseneck Assembly consisting of Part Number 19, 23, 24, 40 to 43, 53 and 54 shall not be provided. The table shall consist of, and be supplied with, the following:

Each	Component
1	Container/Cover
1	Container/Case
2	Center Supports
4	Removable legs with braces
	Body boards; 2 ends and 1 center board
2	Sink Waste Fittings (drains)
1	Cutting board assembly
5	Mats, Cutting Board (spares)
1	Jar Holder assembly
1	Electrical outlet box assembly
1	Water nozzle
1	Foam retainer for scale and electric saw
1	Foam separator for scale and basin
1	Nylon carrying case
2	Foam packing cushions

The following additional items, not covered by DPSC Drawing 25064, shall be supplied and packed within the table container:

2 each	Clamp, Rubber Tubing, Regulating (See 6.4)
2 feet	Tubing, Pure Gum, Amber, 1/2" I.D. X 3/32" Wall (See 6.4)
12 ft. 3 in.	Tubing, PVC, 1/2 "I.D. X 3/32" Wall (See 6.4).
3 feet	Tubing, PVC, 1/2 "I.D. X 3/32" Wall (See 6.4).
	2 Connectors, "T-Type", Polypropylene, 1/2" O.D. (See 6.4)
	1 Head Block (See 6.4).

Lengths listed above are minimum. The above listed clamp and pure gum tubing shall be furnished by the Government to the contractor.

When tested as specified in 4.4.1, the table shall not collapse.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:30Z
Check the source to verify that this is the current version before use.

3.2.1 Container/Cover, and Container/Case. The container for the components referenced in 3.2 shall consist of two similar halves, the container/cover and the container/case, which shall fold in the manner of a suitcase and shall be provided with hinges, carrying handles and catches. With the container open, each half shall form part of the table. In the fully assembled condition, the table shall measure 79 inches long by 27 1/4 inches wide by 36 3/8 inches high. In the closed condition, the container shall fit in the carrying case specified in 3.2.4 The container/cover and the container/case shall be made in conformance with DPSC Drawing 25064, except that the following alternate method of fabrication and welding shall be acceptable. In this alternate method, each container shall be formed, and then welded along the complete length of both inside and outside of each corner seam. All outside welds shall be ground along their entire length, so as to be undetectable, when painted. Internal welds shall be neat and even in appearance, but shall not be ground smooth with container surface. Welded seams shall be located on each long side of the container and at the junction of the flat surface and the radius. All corners shall be spherical and not appear as mitered joints. Dimensions shall comply with DPSC Drawing 25064.

3.2.2 Cutting board mats. Each cutting board mat shall be waterproof and shall be as shown in the drawing.

3.2.3 Retainer and separator. A foam retainer for the scale and electric saw and a foam separator for the scale and basin shall be provided. The retainer and separator shall be free of voids and bubbles on the surface of the furnished products, and shall be in conformance with DPSC Drawing 25064.

3.2.4 Carrying case. A carrying case shall be provided for the container described in 3.2.1. Construction shall be in accordance with DPSC Drawing 25064. Material shall be in accordance with the drawing and with 3.1.3. Stitching shall be Type 301 or 304 of FED-STD-751.

3.2.5 Packing cushions. Two urethane foam cushions, one each for the table bottom and table top, shall be provided. The cushions shall be free of voids and bubbles on the surface of the finished products, and shall be in conformance with DPSC Drawing 25064.

3.3 Finish. All grease shall be removed from the corrosion-resisting steel components. The aluminum parts shall be degreased and etched (see 6.5). After cleaning, all metal surfaces shall be coated with one coat of primer (see 6.5) in accordance with the paint manufacturer's instructions, masking all surfaces which mate or fit other surfaces. After primer has dried, one coat of epoxy paint (see 6.5) formulated to produce a dull surface, shall be uniformly applied to all metal surfaces. Finish shall be oven dried for 20 minutes at 200°F. Painted finish shall be free of runs, chips, cracks, blisters, and orangepeel.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:30Z
Check the source to verify that this is the current version before use.

3.4 Identification marking. The outer surface of the container and cover shall be permanently and legibly marked with the following:

Manufacturer's name or registered trademark
Contract or purchase order number
Federal Stock Number
Item Identification
Letters U.S.

Marking secured by adhesives shall not be acceptable.

The carrying case shall be marked on an exterior surface with black waterproof ink or stencil paint with the same information specified above.

3.5 Service Data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for packing, assembly, operation, maintenance, and lists of component parts. Each part of the unit other than common fasteners, shall be identified in the manual and parts list by the manufacturer's part number. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract, or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of the contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon requests to the Contracting Officer. Waiver of submission of drafts and submission of 15 additional copies shall be at the option of the Government.

3.6 Workmanship. Workmanship shall be first class throughout. Welding shall be continuous, free of blow holes, splatter, pitting, cracks, or slag. Autopsy tables shall be free of defects which detract from their appearance or which may impair their serviceability. Table shall be free of all burrs and sharp edges.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:30Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturers of the materials used in the autopsy table may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the Item Identification, the name of the component or material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and Table I.

Table I.

	Inspection level	AOL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Minor defects	II	4.0
For dimensional examination	S-1	2.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:30Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table II.

Table II

Characteristic	Requirement	Test	Inspection Level	AQL
Load capability	3.2	4.4.1	S-1	4.0

4.3 Examination. The autopsy table shall be examined to determine compliance with all requirements of this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with Table III.

Table III.

Major	Categories and defects*
101	Any component not present.
102	Exposed edges not free of burrs, sharpness.
103	Table does not rest evenly on flat surface.
104	Painted finish not free of runs, chips, cracks, blisters, orangepeel.
105	Welded areas not ground smooth where specified.
106	Welds not free from cracks, porosity, gas pockets.
107	Carrying case not free of cuts, tears.
108	Electrical connections not tight, not properly accomplished.
109	Warning marking on electrical connection box not supplied, not complete, not legible.
Minor	
201	Finish of metal components not free of scratches, marred surfaces.
202	Identification marking not supplied, not complete, not legible.
203	Foam components not free of voids and bubbles on the surface of the finished products.

* Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensional examination. The table and components shall be examined for defects in dimensions. Any dimension not within the specified tolerances shall be classified as a defect.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:30Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Table load bearing capability. A load of 500 pounds shall be evenly distributed on the load bearing surface of the erected table.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. Each autopsy table, as specified, constitutes one unit of issue.

5.1.2 Level A. The table shall be disassembled and the disassembled components shall be positioned within the body of the table along with all other components or kits as specified in 3.2. The table shall be folded, secured and inserted within the carrying case and the case secured.

5.2 Packing.

5.2.1 Level A. Each unit prepared as specified in 5.1.2 shall be overboxed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-601, overseas type. The box shall be lined with a case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the case liner specification. Closure and strapping of the box shall be as specified in the box specification

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking. The exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The autopsy table, with specified components, is intended for use in post mortem examinations, as part of autopsy set FSN 6545-165-6603.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:30Z
Check the source to verify that this is the currenversion before use.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packing (see 5.1 and 5.2).
(c) Government-furnished equipment (see 3.2).

6.3 This specification covers the following item appearing in the Federal Supply Catalog:

Federal Stock Number	Item Identification
6530-043-0915	TABLE, AUTOPSY, Field, Lightweight, Portable

6.4 Connectors referenced in 3.2 are available from Nagle Sybron Corporation, Labware Division, Rochester, N.Y. The head block referenced in 3.2 is available from Lipshaw Manufacturing Company, Detroit, Michigan, Model No. 407. PVC tubing referenced in 3.2 is available from U.S. Stoneware, Akron, Ohio as "Tygon" tubing. Clamp,Rubber Tubing referenced in 3.2 is listed under Federal Stock Number 6640-024-2279. Tubing, Pure Gum, Amber referenced in 3.2 is listed under Federal Stock Number 4720-142-2238.

6.5 A suitable etching solution for that referenced in 3.3 is Oakite 33 available from Oakite Products Inc., Berkeley Heights, N.J. The primer and paint required in 3.3 are available from Pettit Paint Company, Inc., Belleville, N.J., Pettit's No. 6455 metal primer, and their No. 4718 platinum color epoxy paint mixed with their No. 4032 satin hardener.

6.6 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6530-1258

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:30Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:30Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS. The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers, vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO (If "Yes", in what way?)

4. REMARKS

SUBMITTED BY (Printed or typed name and address - Optional)

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:30Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

Defense Supply Agency

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:30Z
Check the source to verify that this is the current version before use.