MILITARY SPECIFICATION

TEST STRIPS AND COLOR CHART, URINARY GLUCOSE AND PROTEIN

This specification is mandatory for use by all Departments and Agencies of the Department of Defense

1. SCOPE

1.1 This specification covers test strips and color chart, urinary glucose and protein, in bottles of 100s. (See 6.2).

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATION

PPP-C-00186a	Containers, Packaging and Packing for Drugs, Chemicals, and Pharmaceuticals.

STANDARDS

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specification and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:32Z
Check the source to verify that this is the current version before use.

2.2 Other publication. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE,
FOOD AND DRUG ADMINISTRATION

Federal Food, Drug and Cosmetic Act and Regulations
Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration Washington, D. C. 20402.)

3. REQUIREMENTS

3.1 Material. Shall be clear plastic strips, each strip approximately 1/5 inch wide by 3-1/4 inches long with two separate and distinctly divided bibulous paper areas which are impregnated with chemicals, enzymes, or mixtures thereof, in each such area, which produce sensitive and reliable qualitative colorimetric determination of glucosuria and proteinuria or the specific absence of glucose and protein when moistened with urine.

3.1.1 A water impervious barrier shall separate the impregnated areas so that the dye of the test areas shall not bleed across the area spacing.

3.2 Shall be suitable for use as a rapid dip and read test for urinary glucose protein.

3.3 Impregnation of test areas. The separate test areas shall be impregnated with the specified compositions for each test, as follows, to produce the anticipated color response charted in the incremented table provided on the label of the immediate container:

(a) Protein Test Portion (distal test area of strip):
Pale yellow reagent misture containing:
Tetrabromphenol blue
Citrate buffers

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:32Z
Check the source to verify that this is the currentversion before use.

(b) Glucose Test Portion (proximal test area of strip):
Red reagent mixture containing:
Glucose oxidase
Catalyst-chromogen combination

3.3.1 The paper (cellulose) used for the reactive portion of the plastic strip shall be Eaton and Dikeman No. 204 white filter paper, or equal, and shall be 0.38 mm plus or minus 10 percent in thickness.

3.3.2 The strips when used shall react to indicate:

(a) the presence of 30 mg/100 ml, 100 mg/100 ml, 300 mg/100 ml, and over 1000 mg/100 ml amounts of protein in urine (time of reading not critical).

(b) the presence of "light" "medium" and "dark" amounts of glucose in urine, when read at 10 seconds after dipping.

The quality of color which is developed shall be evenly distributed and not spotty.

3.3.2.1 The strips shall be tested as specified in 4.3.

3.4 Delivery. Not more than six months shall have elapsed from the date of manufacture of the product, to the date of delivery to the Government.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:32Z
Check the source to verify that this is the currentversion before use.

4.1.1 Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time or from time to time during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 No company supplying any ingredient(s) to the contractor will be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.2 Lot. For purpose of this specification, a lot, batch or control is that single, uniform quantity of test strips produced from one formulation, subjected to the same manufacturing operation, and placed in the immediate containers.

4.3 Tests. The strips shall be tested by appropriate, accurate, and reproducible method(s) as used for commercial release. A copy of the test results shall be submitted to Defense Personnel Support Center, Directorate of Medical Materiel, 2800 South 20th Street, Philadelphia, Pa. 19101. Attention: Quality Assurance Branch DPSC-ATQD.

5. PREPARATION FOR DELIVERY

5.1 Shall be in accordance with all applicable requirements of Interim Federal Specification PPP-C-00186a, dated 15 May 1969, and Amendment-1, dated 27 October 1969, and as specified herein:

5.1.1 Immediate containers. Shall comply with the following classification:

GROUP A CLASS 1 TYPE e STYLE 2 GRADE 1

CLOSURE B or F

A suitable dessicant, appropriately packaged and marked, shall be packaged in each immediate container.

5.2 Labeling. Labeling shall be in accordance with the requirements of the Federal Food, Drug, and Cosmetic Act, and shall include the information required below:

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:32Z
Check the source to verify that this is the current version before use.

5.2.1 Immediate containers. Each immediate container label shall bear the following information. However, the information is not required to appear in the sequence indicated:

(a) label information in accordance with commercial practice

(b) the quantity of strips

(c) the lot or control number

(d) a scaled color chart imprinted and consisting of comparative color reactions at:

1. 30 mg/100 ml, 100 mg/100 ml, 300 mg/100 ml and over 1000 mg/100 ml amounts of protein in urine (time of reading not critical)

2. Light, medium, and dark, for the amount of glucose in urine, when read at 10 seconds after dipping.

(e) instructions advising not to remove desiccant and proper storage.

(f) directions for using the reagent strips, to include a color chart for qualitative colorimetric estimation of readings.

Note: Paragraph 4.3.2 Label adhesion test. In Supplement No. 2, FSN 6505 and FSN 6810 Items, on page 7, delete the "Note" at the bottom of the page and substitute the following:

"Note: Wrap-around label. The wrap-around label is one that is longer than the circumference of the immediate container to which it is affixed, and may or may not have a free flap portion."

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:32Z
Check the source to verify that this is the current version before use.

5.2.2 Unit packages. Each unit package label shall bear the following information. However, the information is not required to appear in the sequence indicated:

(a) the item identification designated as "TEST STRIPS AND COLOR CHART, URINARY GLUCOSE, AND PROTEIN"

or

"Test for Glucose and Protein in Urine"

(b) the quantity of strips designated as "100s" or "100 Reagent Strips"

(c) the Federal Stock No.

(d) the lot or control number

(e) the date of manufacture

(f) the name and address of the manufacturer. When the manufacturer is not the contractor, the name and address of the contractor shall also appear.

When both names are placed on the label, the following designations shall precede the names:

"MFR" for the manufacturer and
"CONTR" for the contractor.

(g) the following statement or similar statement: "Carefully read instructions for use on inside of carton before performing test."

5.2.2.1 Instructions. Complete detailed instructions and precautions for use and handling of the reagent strips, and directions for "Interpretation of the Color Reactions" shall be printed on the inside area of the enclosing carton surrounding the immediate container.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:32Z
Check the source to verify that this is the current version before use.

5.3 Packaging and packing.

5.3.1 Unit of issue. One bottle containing 100 strips, as specified, constitutes one unit of issue.

5.3.2 Unit package. Each unit shall be packaged as specified in 5.2.5 of PPP-C-00186a.

5.3.3 Procedure code. Procedure code No. 3 of PPP-C-00186a applies, except in Table I, column 2, delete "12" and substitute "24" and in column 4, delete "12/144" and substitute "24/144."

5.4 Marking.

5.4.1 Intermediate package. In paragraph 5.5.3 of PPP-C-00186a, add: "Date of manufacture shall be shown in lieu of date packed. Marking shall include the legend:

'STORE AT CONTROLLED ROOM TEMPERATURE (59° - 86° F.).'

5.4.2 Exterior container. In paragraph 5.5.4 of PPP-C-00186a, add: Date of manufacture shall be shown in lieu of date packed. Marking shall include the legend:

'STORE AT CONTROLLED ROOM TEMPERATURE (59° - 86° F.). "

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification
(b) the Federal Stock No.
(c) applicable levels of packaging and packing (see PPP-C-00186a).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:32Z
Check the source to verify that this is the current version before use.

6.2 This specification covers the following item:

FSN 6505-226-1203 TEST STRIPS AND COLOR CHART, URINARY GLUCOSE AND PROTEIN, 100s.

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6505-1298

Review activities:

Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

6505-226-1203

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:32Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

Defense Supply Agency

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:32Z
Check the source to verify that this is the current version before use.