MIL-W-36849
25 May 1972

MILITARY SPECIFICATION

WRAPPER, STERILIZATION, PAPER, BROWN, DISPOSABLE, 25 INCHES SQUARE, 500s

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a surgical paper wrapper, suitable for wrapping supplies for sterilization by autoclaving.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

UU-P-31	Paper, General Specifications and Methods of Testing.
PPP-B-636	Box, Fiberboard.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:27Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material. The basic material of the paper wrapper shall be a good quality brown kraft paper, free from natural impurities such as; dirt, foreign bodies, or pieces of undigested bark. The paper shall be creped, and then embossed at right angles to the direction of the crepe, so that a two way stretch is provided. The paper shall have the following properties:

Thickness.	0.0110 minimum.
Elongation:	
Machine direction*.	35 percent minimum.
Cross direction.	5 percent minimum.
Tensile strength dry:	
Machine direction.	7 pounds per inch minimum.
Cross direction.	7 pounds per inch minimum.
Tensile strength wet**	
Machine direction.	1 pound per inch minimum.
Cross direction.	1 pound per inch minimum.

* Machine direction is direction of the crepe.
** Fifteen minute immersion period.

3.2 Construction. The paper, before cutting into sheets, shall have the following legends, or equivalents, printed alternately throughout on one side of the sheet:

"REUSABLE WRAPPER. RETURN TO CENTRAL SERVICE".

"REUSABLE WRAPPER. TO UNWRAP PEEL BACK ADHESIVE TAPES WITH CARE TO AVOID TEARING".

Ink shall be a black or blue indelible ink. Letters shall be at least 3/16 inch high. The vertical and horizontal distance between each legend shall be approximately 6-1/2 inches. The paper shall be cut into sheets 25 inches square, and shall be packaged 500 sheets per package. Dimension of sheets shall be 25 inches ± 3/8 inch. Weight of 500 sheets shall be not less than 23 pounds.

3.3 Workmanship. Workmanship shall be first class throughout. The wrappers shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:27Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the wrappers, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in table I and table III.

Table I.

Categories and defects**	
Major A	
101	Creping inadequate or not present.
102	Embossing inadequate or not present.
103	Instructions not present.
Minor A	
201	Labeling not legible and free from defacing marks.

** Inspection is not restricted to the classified possible defects listed above.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in table II. Unit of product for examination purpose shall be one wrapper.

4.2.2 Examination. The wrappers shall be examined for defects including, but not limited to, those listed in table I.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:27Z
Check the source to verify that this is the currentversion before use.

Table II - Sampling for examination

	Inspection Level	AQL Percent Defective
Major A	S-2	1.5
Minor A	S-2	4.5
Total		6.0
For dimensional examination	At least 10 samples from each lot	6.5
(Unit of inspection is one wrapper.)		

4.2.3 Sampling for tests. Minimum number of samples tested from each lot shall be as indicated in the test methods (Table III). The acceptance number shall be zero for all sample sizes. This does not apply to dimensional examinations. Acceptance shall be as indicated in Table II.

4.3 Tests and Dimensional Examination. Tests and dimensional examination shall be performed in accordance with Table III.

Table III Tests and Examinations

Characteristic	Requirement reference paragraph	Test method.
Thickness.	3.1	UU-P-31b (method 173)
Elongation, percent.	3.1	UU-P-31b (method 171)
Tensile strength, dry.	3.1	UU-P-31b (method 171)
Tensile strength, wet.	3.1	UU-P-31b (method 172)
Weight, 500 sheets.	3.2	UU-P-31b (method 110)
Dimensions.	3.2	Ruler or yardstick.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:27Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packing.

5.1.1 Unit of issue. One box containing 500 wrappers, as specified, constitutes one unit of issue.

5.1.2 Level A. Five hundred wrappers shall be packed in a box of appropriate size constructed in accordance with PPP-B-636, grade V3c. Closure, strapping, and waterproofing shall be as specified in the appendix of the box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.1.3 Level B. Five hundred wrappers shall be packed in a box of appropriate size constructed in accordance with PPP-B-636, type CF, class domestic. Closure shall conform to method II as specified in the appendix of PPP-B-636.

5.1.4 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2 Marking. Marking shall be as specified in MIL-STD-129.

Add the following paragraphs:

"6. NOTES

6.1 Intended use. The wrapper covered by this specification is intended for wrapping supplies for sterilization by autoclaving.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) Federal stock number (FSN).
(c) Selection of applicable level of packing (see 5.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:27Z
Check the source to verify that this is the current version before use.

6.3 The wrapper covered by this specification is listed in the Federal Supply Catalog under the following stock number:

Federal Stock Number	Item Identification
6530-926-2234	WRAPPER, STERILIZATION, Paper Brown, Disposable, 25 Inches Square, 500s

6.4 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification or those which are commercially available, but is intended to cover the types which are normally procured to meet the military requirements.

6.5 Supersession Data. This specification includes the requirements of Defense Medical Purchase Description No. 3, dated 17 March 1971, and Modification No. 1, dated 28 December 1971.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity
Defense Supply Agency - DM

Project No. 6530-1252

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:27Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

6530-926-2234 MIL-W-36849

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:27Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DOD-304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:27Z
Check the source to verify that this is the current version before use.