A-A-53507
28 January 1991

COMMERCIAL ITEM DESCRIPTION

PLUGGER, PLASTIC FILLING, DENTAL
CONTOURING INSTRUMENT, MATRIX, DENTAL

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-P-001278a which is cancelled.

This commercial item description covers double-ended dental instruments suitable for manipulating and contouring plastic-filling materials.

Salient Characteristics:

The pluggers shall as specified below:

Ladmore No. 3
No. 1-2
No. 5-7
No. FP-2A
Woodson No. 1
Woodson No. 2
Woodson No. 3

Material. Material shall be fabricated from corrosion-resistant steel in accordance with ASTM Designation F899 for composition only.

One-piece construction	Class 4, type 440A.
Three-piece construction	
Operative ends	Class 4, type 440A.
Handles	Class 3, type 303 or Class 4, type 416.

As an alternate, handle shall be chromium or nickel plated brass. Brass shall conform to copper alloy No. 260 or 330 as specified in ASTM Designation B135 for composition only.

Design and dimensions. The design and dimensions of the instrument shall conform to the drawing listed below for the applicable instrument, and for the respective design and instrument number.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Design/Instrument Number	Drawing Number
Ladmore (LAD) No. 3	20697
No. 1-2	20134
No. 5-7	20134
No. FP-2A	24147
Woodson (WDS) 1, 2 and 3	20650

The instruments shall be suitable for manipulating plastic filling materials. The No. FP-2A instrument shall be suitable for contouring proximal surfaces of dental amalgam matrix bands.

Construction. Instruments shall be double-ended, one piece or three-piece construction, consisting of two operative ends and a handle. The operative ends shall be permanently secured in the handle of the three-piece instruments, and the interface of the shank and handle shall be smoothly blended with no crevices evident.

Handle serrations or knurls. The finger grip portion of the handles shall be finely serrated or fine diamond knurled.

Hardness. The operative ends and shanks of the instruments shall have a Rockwell hardness of C52 to C56. The handles shall have a Rockwell hardness of C20 minimum, when tested in accordance with ASTM Designation E18 or E92 and E140.

Corrosion resistance. The entire instrument shall be tested for the boil and copper sulfate tests in accordance with ASTM Designation F1089.

Finish.

Operative ends and shanks. Operative ends and shanks of the instruments shall have a bright polished finish.

Handles. The handles of the instruments shall have a bright polished finish, except for the serrated or knurled portion.

Identification marking. The handles of the instruments shall be permanently and legibly marked with the applicable instrument design (or abbreviation thereof) and the instruments number as shown herein, the manufacturer's name or registered trademark, and the word "stainless". Instruments of foreign manufacture shall bear the name of the country of manufacture. All markings shall be on the smooth portions of

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

the handles. As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless is acceptable, provided prior approval is received from the cognizant contracting officer.

Workmanship. The instruments shall be free from defects such as nicks, feathers, pits, burrs, corrosion or coarse grind marks which detract from their appearance or impair their serviceability.

Unit. Each (EA). One plugger, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality for raw materials, supplied by the manufacturer of the material, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection, examination, and test characteristics. The unit of product for inspection, examination, test purposes shall be one plugger. Unless otherwise specified, sampling for examination shall conducted at inspection level II, except for dimensions which shall be conducted at inspection level S-1, AQL 2.5 (percent defective) and sampling for tests (when applicable) shall be conducted at inspection level S-1. Acceptance number for tests shall be zero (0) for all samples.

Tests. Tests shall be conducted to determine compliance with specification requirements..

Chemical analysis. In the absence of certificates of quality for the raw material, a chemical analysis shall be conducted. The chemical analysis shall be performed by reagent solutions, combustion methods, nonemission methods, or by spectrochemical analysis. Any ASTM method shall be acceptable.

Contract certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirement.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Number	Item Identification
6520-00-536-5205	PLUGGER, PLASTIC FILLING, DENTAL, Ladmore, No. 3
6520-00-536-5250	PLUGGER, PLASTIC FILLING, DENTAL, No. 1-2.
6520-00-536-5305	PLUGGER, PLASTIC FILLING, DENTAL, No. 5-7.
6520-00-536-5350	PLUGGER, PLASTIC FILLING, DENTAL, Woodson No. 1
6520-00-536-5405	PLUGGER, PLASTIC FILLING, DENTAL, Woodson No. 2
6520-00-536-5450	PLUGGER, PLASTIC FILLING, DENTAL, Woodson No. 3
6520-00-935-7176	CONTOURING INSTRUMENT, MATRIX, DENTAL, Double-Ended, No. 2FP-2A.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2037

LOCATION: Word/Pluggers/kd

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide inputs. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-MSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53507	Plugger, Plastic Filling, Dental Contouring Instrument, Matrix

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording.

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.