A-A-53508
27 June 1988

COMMERCIAL ITEM DESCRIPTION

PLUGGER, PLASTIC FILLING, DENTAL, (Glick No. 1)

The General Services Administration has authorized the use of this Commercial Item Description in preference to Military Specification MIL-P-36959.

Salient characteristics:

This commercial item description covers a plugger of Glick No. 1 design for use in endodontic procedures.

General requirements. The plugger shall be in accordance with the applicable requirements as specified herein.

Material. The plugger shall be fabricated of Class 4, Type 440A martensitic corrosion-resisting steel, as specified in ASTM F899, for composition only.

Style, design and dimensions. The plugger shall be a one-piece, double-ended, plastic filling type plugger of the Glick No. 1 design. One operative end (plastic filling end) shall be suitable for placing temporary fillings in root canal therapy. The other operative end (plugging end) shall be suitable for endodontic plugging and shall be graduated with millimeter markings to permit gaging the depth of plugging. The millimeter markings shall be legible, sharply defined lines of uniform width and depth. The handle shall be octagonal, and shall be serrated except for one face which shall be smooth for identification marking purposes. Serrations shall be fillet or notched-type serrations. There shall be 40 $\pm$ 10 serrations per inch. The depth of serrations shall be 0.002 to 0.004 inch. Handle shall be straight. The style, design and dimensions of the plugger shall conform to Defense Personnel Support center (DPSC) Drawing 24322.

Hardness. After proper stress relieving, the Rockwell hardness of the operative ends, shall be C50 to C56, when tested as specified herein.

Corrosion resistance. The complete plugger shall be properly passivated and shall meet the resistance to copper sulfate and boil test requirements of ASTM F1089 when tested as specified herein.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Finish. The plugger shall have a mirror or bright finish. Serrated portions shall be free of barbs and shall provide adequate finger grip during use. The operative ends shall be smoothly finished including the flat face of the plugging operative end tip, and shall be free from grind marks, gouges, cracks, burrs, feathers or other irregularities. The faces of the plastic filling operative end shall be blended to form smoothly contoured surfaces. The operative end (plugging end) shall be uniformly and smoothly tapered.

Identification marking. Plugger shall be permanently and legibly marked as specified herein. All marking shall be on the smooth face and shall read in the direction from the plastic filling end to the plugging end. The plugger shall be marked with the identification designation "GLICK 1" on the plugging end portion of the handle. The word "stainless" shall be marked on instruments of foreign manufacture. As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is received from the cognizant contracting officer.

Workmanship. The plugger shall be free from defects which either detract from its appearance or may impair its serviceability, proper functioning, or intended use.

Unit. Each (EA). One plugger, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the material, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered are classified as specified herein.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an inspection level II, except the dimensions shall be conducted at inspection level S-1, with an AQL of 2.5 (percent defective). The unit of product for examination purposes shall be one plugger.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The acceptance number shall be zero (0) for all sample sizes. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics. The unit of product for test purposes shall be one plugger. Lots for hardness testing shall be as specified herein.

Tests.

Component	Characteristic	Inspection level
Operative End	Hardness	S-2
Entire Instrument	Copper Sulfate	S-2
Entire Instrument	Boil Test	S-2

Sampling of raw materials. A sufficient amount from each lot of each metal used in the fabrication of the pluggers shall be selected for chemical analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Examination. The plugger shall be examined to determine compliance with specification requirements. Examination shall be conducted in accordance as specified herein. The plugger shall be examined for defects, including but not limited to, those listed below.

Defects.

Style and design not as specified.
Instrument not constructed as specified.
Operative end and shank finish not as specified.
Operative end not free from grind marks, gouges, cracks, burrs, feathers and other imperfections.
Millimeter markings not legible, sharply defined lines.
Plugging operative end not uniformly and smoothly tapered.
Operative end dimensions not as specified.
Handle finish not as specified.
Instrument not permanently and legibly marked with name or trademark of the contractor.
Instrument identification not present or not properly located.
Handle and shank dimensions not as specified.
Instrument not free from dirt, oil, compound or foreign matter.
Handle serrations, either not uniform or not complete.
Handle not straight.

Tests. Tests shall be conducted to determine compliance with the requirements of this specification.

Hardness. Hardness shall be determined in accordance with ASTM Designation E18 (Rockwell) or E92 (Vickers or Diamond Pyramid).

Corrosion test.

Copper sulfate test. The copper sulfate test shall be conducted in accordance with ASTM F1089.

Boil test. The boil test shall be conducted in accordance with ASTM F1089.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document.

National Stock Number	Item Identification
6520-00-116-0843	PLUGGER, PLASTIC FILLING, DENTAL, Glick No. 1

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2028

LOCATION: NCR DISK # 5

47386

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53508	PLUGGER, PLASTIC FILLING, DENTAL, (Glick No. 1)

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.