A-A-53509
18 May 1988

COMMERCIAL ITEM DESCRIPTION

RETRACTOR, SURGICAL (Harrington)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers one type of surgical retractor suitable for use during deep abdominal and thoracic surgery.

Salient characteristics:

Style and design. Shall be a large-size, extra-deep surgical retractor. Style and design of the retractor shall conform to Defense Personnel Support Center (DPSC) Drawing No. 24458. The design specified herein shall be classified as the "Harrington" design. The blade-to-handle joint shall not crack or break under a 15 lb tensile load (See Drawing).

Materials. Material requirements for each of the retractor's components are as follows: The blade shall be fabricated from martensitic corrosion-resisting steel. The handle shall be fabricated from austenitic corrosion-resisting steel. These materials shall conform to ASTM Designation F899. As an alternate, both handle and blade may be fabricated from martensitic corrosion-resisting steel.

Hardness. Retractor components fabricated from Class 4 martensitic stainless steel shall have a Rockwell Hardness C Scale (HRC) of 40 to 47. Retractor components fabricated from Class 3 austenitic stainless steel shall have a Rockwell Hardness B Scale (HRB) of 75 to 100. Each component shall be tested for hardness.

Corrosion-resistance. The retractor shall resist corrosion. The retractor shall be tested in accordance with the Copper-Sulfate Test and Boil Test specified in ASTM Designation F1089. However, the results of the Copper-Sulfate Test shall not apply to a retractor component made of austenitic corrosion-resistant steel.

AMSC/NA FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:20Z
Check the source to verify that this is the currentversion before use.

Finish. The retractor shall have a mirror or mirror and satin finish except where beading and knurling are required. (See Drawing). The retractor shall be passivated. Passivation shall render the retractor corrosion-resistant.

Identification markings. The retractor shall be permanently marked with the manufacturer's name or registered trademark and the word "stainless". In addition, retractors of foreign manufacture shall bear the name of the country of manufacture.

Workmanship. The retractor shall be free from dents, pits, nicks, burrs, corrosion, sharp edges, and any other defects that detract from its appearance or impair its serviceability.

Unit. Each (EA). One retractor, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105 at level II, AQL 1.0%. For inspection purposes, the unit of product shall be one retractor.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-139-8196	RETRACTOR, SURGICAL, Harrington, Curved, Large 21/4 By 7 Inch (5.7 by 17.8 cm) Blade

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and /or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 54263

Project No. 6515-3926

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.