COMMERCIAL ITEM DESCRIPTION

SPLINT, LEG,(Traction)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a leg traction splint, Hare style, for use in immobilizing long bone fractures of the lower extremity of an adult.

Salient characteristics.

Shall be a leg traction splint suitable for use in immobilizing a fractured leg. Splint shall consist of an adjustable leg support frame, an ishial support, a positive traction device, a heel stand, an ankle strap, and leg support straps.

The leg support frame shall consist of two telescoping sections connected at the proximal (thigh) end by the ishial support, and at the distal (ankle) end by the positive traction device. Each telescoping section shall consist of an upper portion fabricated of aluminum tubing 5/8 inch in diameter by 21.5 inches long, and a lower portion fabricated of 3/8 inch diameter by 28.4 inches long aluminum rod which slides freely within the upper portion to control the extended length of the splint. The upper and lower portions of each section shall be held at the desired length by a locking device consisting of a threaded collet permanently attached to the upper portion, and a collet sleeve that shall thread onto the collet, engaging the collet sleeve and collet onto the lower portion. Collet and collet sleeve shall be fabricated of 1018 cold rolled steel.

The ishial support shall be fabricated from 1/2 inch diameter 6061-T6 aluminum rod contoured with various compound angles to form a 7-1/4 inch width between the telescoping sections at the proximal end of the splint. The rod shall be covered with a polyurethane foam ishial pad, 1-5 8 inches in diameter by 7-1/2 inches long, to prevent undue pressure at the ishium which could cause restriction of blood flow. The ishial pad shall have sufficient memory to return to its original shape when not in use.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

The positive traction device shall consist of a ratchet assembly fabricated of A-380 aluminum die castings with spring loaded lock-pin to apply positive locking , 18 degrees per engagement, to the distal (ankle) end of the splint. The positive traction device shall require no more than manual force for operation. A nylon webbing traction strap approximately 2 inches wide by 19-1/2 inches long shall be permanently attached to the ratchet assembly. An "S" hook shall be attached to the free end of the traction strap which provides a means for connection to the tri-ring ankle strap.

The heel stand shall be adjustable and self locking and shall be attached to both sides of the lower frame to elevate the leg at least 4 inches from the ground. Stand shall be "U" shaped and fabricated of corrosion resistant metal.

The tri-ring ankle strap shall be nylon webbed strap partially covered with 1/2 inch thick closed cell foam to protect and immobilize the foot and ankle when attached to the traction strap. The strap shall be "T" shaped to wrap around the ankle with "D" rings attached to each end for connection to the ratchet traction strap. The strap shall be 2 inches wide by approximately 32 inches long to fit an adult.

A nylon webbing thigh strap with hook and pile closure shall be provided to secure the thigh to the splint. The thigh strap shall consist of two parts. One part shall be 10 inches long by 1-1/2 inches wide and the other part 22 inches long by 1-1/2 inches wide.

At least four leg securing straps shall be provided. Each strap shall be composed of two pieces connected in the middle, to form a lower and an upper loop. The lower loops are to be connected around the two aluminum tubes which form the leg support frame, and shall be fabricated from nonelastic polypropylene material with hook and pile closure. Each lower loop shall be 3 inches wide by 20-1/2 inches long.

The upper loop of each leg securing strap shall be fabricated from nylon elastic material with hook and pile closures. The upper loops will enclose the leg at four locations; two above the knee and two below the knee. The upper loops shall be furnished in the following sizes:

Strap 1: Below calf strap shall be 3 inches wide and 17-1/2 inches long.

Strap 2: Below knee strap shall be 3 inches wide and 17-1/2 inches long.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:20Z
Check the source to verify that this is the current version before use.

Strap 3: Above knee strap shall be 3 inches wide and 21 inches long.

Strap 4: Mid thigh strap shall be 3 inches wide and 24 inches long.

All four straps shall have hook and pile closure on both upper and lower loops. All lower loop closures shall be fabricated from 2 inch wide strips. All upper loop strap closures shall be fabricated from 1-1/2 inch wide strips. Each strap closure is composed of hook and pile. Each strap shall have different lengths of hook and pile according to placement numbers 1 thru 4. The following measurements apply to the length of the hook and pile sewn to each closure.

Strap 1. 1-1/2" by 8" hook
1-1/2" by 8" pile
2" by 5" hook
2" by 4" pile

Strap 2. 1-1/2" by 8" hook
1-1/2" by 8" pile
2" by 5" hook
2" by 4" pile

Strap 3. 1-1/2" by 8" hook
1-1/2" by 8" pile
2" by 6" hook
2" by 4" pile

Strap 4. 1-1/2" by 8" hook
1-1/2" by 8" pile
2" by 7" hook
2" by 4" pile

The hooks on all closures shall incorporate a minimum of 500 hooks per square inch.

Each leg securing strap shall be labelled and numbered for proper placement and application.

The splint shall be fabricated of materials which are corrosion resistant or which are treated to render them corrosion resistant. Where painting or plating is used, appropriate cleaning and priming shall be performed prior to finishing.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:20Z
Check the source to verify that this is the currentversion before use.

Each splint shall be contained in a carrying case made of "STAPH CHECK®" brand of vinyl polyester laminate fabric, or equal, to reduce and inhibit bacterial growth, and will have a complete set of instructions for use of the splint.

Workmanship. The splint shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One splint, in a carrying case, complete with [illegible] specified [illegible].

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities [illegible] specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:20Z
Check the source to verify that this is the currentversion before use.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one splint.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-076-4714

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:20Z
Check the source to verify that this is the current version before use.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4293
Location: Disk # 6

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better the DoD provides this form for use in submitting comments and suggestions for improvements All users of military standardization documents are invited to provide suggestions This form may be detached, folded along the lines indicated taped along the loose edge *(DO NOT STAPLE)* and mailed In block 5 be as specific as possible about particular problem areas such as wording which required interpretation was too rigid restrictive loose ambiguous or was incompatible and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents nor to request waivers deviations or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document/s or to amend contractual requirements

(Fold along this line)

Fold along this line

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-[redacted]RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:20Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-52561	SLIP-ON [illegible]

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
	☐ USER
b. ADDRESS (Street, City, State, ZIP Code)	☐ MANUFACTURER
	☐ OTHER (Specify)

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) - Optional	b. WORK TELEPHONE NUMBER (Include Area Code) - Optional
c. MAILING ADDRESS (Street, City, State, ZIP Code) - Optional	8. DATE OF SUBMISSION (YYMMDD)

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:20Z
Check the source to verify that this is the current version before use.