METRIC

A-A-53513A
30 January 1993
SUPERSEDING
A-A-0053513(DM)
8 August 1988

COMMERCIAL ITEM DESCRIPTION

GLOVE, PATIENT EXAMINING AND TREATMENT, Latex Rubber, Pre-Powdered, Disposable, Non-Sterile

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-G-36592 which is cancelled

This Commercial Item Description covers non-sterile, ambidextrous, latex rubber patient examining and treatment gloves suitable for use in oral and rectal examinations.

Salient characteristics:

General. "Gloves", as used herein, shall refer to all patient examining and treatment gloves covered by the scope of this document Gloves shall be ambidextrous. Glove sizes shall be limited to the following: small, medium, and large.

Material Gloves shall be made of latex rubber This material shall comply with the requirements for one of the three types of latex rubber defined in ASTM D1076. Material samples shall be cut from finished gloves and subjected to all inspections specified by ASTM D1076.

Dimensions. Gloves shall comply with the dimensional requirements specified in ASTM D3578. The dimensions of glove samples shall be inspected in accordance with ASTM D3578.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Performance. Gloves shall be leakproof. Sample glove pairs shall be inspected for leaks in accordance with ASTM D5151-92.

Physical Properties. Gloves shall comply with the physical property requirements specified in ASTM D3578. The physical properties of glove samples shall be inspected in accordance with ASTM D3578.

Finish. Gloves shall have a smooth, uniform finish. Gloves shall be pre-powdered with a USP-grade absorbable dusting powder. Gloves shall not be powdered with talcum powder. Glove samples shall be inspected for finish. Samples of glove powder shall be inspected for compliance with USP requirements (See Note 1).

Workmanship. Gloves shall be free from blooming, blisters, fish-eyes, frosting, shorts, wrinkles, foreign matter, visible holes, missing fingers, missing thumbs, fingers that are fused together, thumbs that are fused together, and any other defects that detract from their appearance or impair their serviceability. The definitions of the first seven defect terms shall be as specified in ASTM D883. Glove samples shall be inspected for workmanship.

Unit. Package (PG). One package containing one hundred gloves, as specified, constitutes one unit. The unit package shall be a top opening and dispensing container.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, shall be defined as the examination and testing of an item. Examination shall be defined as unaided visual, auditory, tactile, or olfactory investigation. Testing shall be defined as the use of laboratory equipment and procedures to determine various properties of an item. Inspections shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. For inspection purposes, the unit of product shall be one glove.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Seven samplings shall be performed - one for each salient characteristic specified herein. For each sampling, inspection level S-2 shall be used; the Acceptable Quality Level (AQL) shall be 0.65 percent.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE 1: USP requirements for glove powder are given in the monograph entitled "Absorbable Dusting Powder".

NOTE 2: The following National Stock Number and their Item Identifications are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-364-8553	GLOVE, PATIENT EXAMINING AND TREATMENT, Latex Rubber, Ambidextrous, Medium Size, Pre-Powdered, Disposable, Non-Sterile, 100s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-364-8554	GLOVE, PATIENT EXAMINING AND TREATMENT, Latex Rubber, Ambidextrous, Large Size, Pre-Powdered, Disposable, Non-Sterile, 100s
6515-01-365-6183	GLOVE, PATIENT EXAMINING AND TREATMENT, Latex Rubber, Ambidextrous, Small Size, Pre-Powdered, Disposable, Non-Sterile, 100s

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5450
Location: a:\word\doc\Komada\P1-GLOVE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-53513A	2 DOCUMENT DATE (YYMMDD) 30 January 1993

3 DOCUMENT TITLE
GLOVE, PATIENT EXAMINING AND TREATMENT, Latex Rubber, Pre-Powdered, Disposable, Non-Sterile

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle)	b ORGANIZATION
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME Defense Personnel Support Center ATTN: DPSC-MSE	b TELEPHONE (Include Area Code) (1) Commercial: (215) 737-2870 (2) AUTOVON: 444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT. Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:20Z
Check the source to verify that this is the current version before use.