COMMERCIAL ITEM DESCRIPTION

LID, ANAEROBIC CULTURE APPARATUS

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a lid assembly which is a component of an anaerobic culture apparatus.

Salient characteristics.

The lid shall be the lid assembly of, or lid assembly replacement for, a self contained anaerobic apparatus suitable for use in conjunction with other components entering into the overall apparatus. (See notes)

The lid assembly shall consist of the following components:

A cover, fabricated of a suitable plastic material with an outside diameter equal to the outside diameter of the jar flange. The cover shall have an annular groove on the inner surface fitted with an "O" ring gasket to form a seal with the top surface of the jar flange. A central boss on the internal surface shall contain a female blind thread for mounting the catalyst reaction chamber, and a central boss on the outside surface shall have a depression for use in centering and securing the cover into position on the jar flange in conjunction with the clamp and thumb screw components.

A "C" clamp, fabricated of a suitable plastic material, which clears the outside diameter of the jar flange and has a central tapped hole into which the thumb screw shall fit.

A corrosion resistant metal thumb screw with knurled head, and thread diameter not less than 7/16 inch (1.1 cm). Length of thread and dimensions of the knurled head shall be such that the cover can be easily positioned and sealed on the jar flange or removed therefrom by hand without the need for tools.

AMSC N/A FSC 6640
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

A catalyst reaction chamber constructed from corrosion resistant metals. The top section of the chamber shall be fitted with a short tubular stud with external thread fitting the blind thread in the cover. The internal diameter of the stud shall be as large as practicable to facilitate loading catalyst pellets without reducing the wall thickness below the minimum required for strength and durability. The bottom section of the chamber shall be fabricated from stainless steel wire cloth or mesh and consist of two containers, one symmetrically mounted within the other and permanently secured at the joint between the top and bottom sections.

The lid assembly shall be provided with a charge of catalyst (see notes), capable of producing and maintaining an anaerobic atmosphere within the jar at room temperature indefinitely, unless contaminated by hydrogen sulfide, when used in conjunction with the other components of the overall anaerobic apparatus.

Workmanship. The lid shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One lid as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:20Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for tests purposes shall be one lid.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTES: The following National Stock Number is covered by this document:

6640-00-001-6670

The lid assembly shall be compatible for use in conjunction with the following components entering into the overall anaerobic apparatus:

Jar	6640-00-926-7655
Hydrogen Generator	6640-00-935-6863
Indicator	6640-00-145-1150
Catalyst	6640-00-933-8739

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1488

Location: Disk # 6

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-[redacted]RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53514	LID, ANAEROBIC CULTURE APPARATUS

3a. NAME OF SUBMITTING ORGANIZATION

b ADDRESS *(Street, City, State ZIP Code)*

4 TYPE OF ORGANIZATION *(Mark one)*

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER *(Specify)*

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last First, MI)* – Optional

b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional

c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional

8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.