A-A-53518
3 June 1988

COMMERCIAL ITEM DESCRIPTION

STAPLE UNIT, SURGICAL, (End-To-End Anastomosis)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers three sizes of staple units used to create end-to-end, side-to-side anastomoses.

Salient characteristics:

Style and design. Staple units shall be supplied in three sizes. Style and design of all sizes of staple units shall be in accordance with Figure 1. Staple units shall consist of a cartridge and anvil.

The anvil shall consist of the anvil body, mounting nut, anvil plate, and cutting ring. The anvil body shall be vented. Its proximal portion shall house the non-lubricated mounting nut, and its distal portion shall house anvil plate and a machined cutting ring. The anvil plate shall contain staple-clinching pockets arranged in a double-staggered, annular pattern. The anvil shall be keyed to the drive-rod of the compatible stapler.

The cartridge shall consist of a vented body shell, keyed bayonet-sleeve, keyed face-plate, push-rod assembly, vented push-block, and cartridge knife. The body shell shall house all other cartridge parts and shall be keyed to the face-plate and bayonet-sleeve. The sleeve shall be press-fitted into the distal end of the body shell and shall provide a connection that is keyed to the drive rod and drive tube of the compatible stapler. The face-plate shall be press-fitted into the proximal end of the body shell and shall contain a double-staggered, annular array of staple channels. This array shall match the anvil's array of staple-clinching pockets. Each channel shall contain a surgical staple and corresponding push-rod. The push-rod assembly shall be bonded to the proximal end of the push-block. The cartridge knife shall be hot-staked to the proximal end of the push-block. The knife shall be located within the I.D. of the face plate.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

The cartridge keys and anvil keys shall permit only one rotational orientation between the cartridge and anvil. In this orientation, the staple-channels and clinching pockets shall exhibit a one-to-one correspondence.

Dimensions. For all staple units covered by this document, the diameters of the staple-pattern and cartridge knife shall be as indicated in Table I. Dimensions of the staples shall be as follows: 4.8 ± .5 mm leg length, 4.0 ± .5 mm width, and 0.28 ± .05 mm wire diameter.

Table I.
Staple Unit Dimensions

Size	Cartridge Knife Dia.(1)	Staple-Pattern Dia (2)
1	15 mm	25 mm
2	18 mm	28.6 mm
3	21.2 mm	31.6 mm

(1) Tolerance: ±0.05 mm

Materials. Staples shall be fabricated from corrosion-resisting steel wire that conforms to ASTM Designation F55, Grade 2. The staple unit's anvil, cartridge, cutting ring, and cartridge knife shall be fabricated from the materials specified below:

Cartridge: Anodized aluminum, rigid plastic, corrosion-resisting steel

Anvil: Anodized aluminum, corrosion-resisting steel

Cutting ring: Soft, machinable plastic

Cartridge knife: Martensitic Corrosion-resisting steel

All corrosion-resisting steel parts shall be properly tempered and passivated. All martensitic corrosion-resisting steel parts shall be properly hardened and passivated.

Sterility. Each staple unit shall be sterile in accordance with the latest edition of the USP.

Workmanship. Staple units shall be free from defects that detract from their appearance or impair their serviceability. Staple units shall show no evidence of loose parts, damaged parts, or ejected staples after falling from a height of 18 inches.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing 12 individually packaged sterile staple units, as specified, constitutes one unit. Each staple unit shall be packaged sterile in sealed, peel-open type inner and outer see-through blister paks able to maintain sterility of contents unless package is damage or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspection. Sampling for inspection shall be done in in accordance with MIL-STD-105 at level S-2, AQL 0.65%. For inspection purposes, the unit of product shall be one staple unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:20Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: This document covers the following NSNs:

NSN	ITEM IDENTIFICATION
6515-01-151-1835	Staple Unit, Surgical, Intraluminal, 31.6 mm, Disposable, 12s
6515-01-151-1837	Staple Unit, Surgical, Intraluminal, 25 mm, Disposable, 12s.
6515-01-207-8218	Staple Unit, Surgical, Intraluminal, 28.6 mm, Disposable, 12s.

Note 2. NSN-to- Size relationships are as specified:

Size	NSN
1	6515-01-151-1837

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Size	
2	6515-01-207-8218
3	6515-01-151-1835

Note 3. The above-listed NSNs are compatible with the following:

NSN	Item Identification
6515-01-205-2379	STAPLER, SURGICAL, INTRALUMINAL, Corrosion-Resistant Alloy:

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4110

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.




FIGURE 1

STAPLE UNIT, SURGICAL
(FOR GENERAL STYLE AND DESIGN)

1. ANVIL
2. CARTRIDGE
3. VENTS
4. MOUNTING NUT
5. BAYONET SLEEVE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53518	STAPLE UNIT, SURGICAL (End-To-End Anastomosis)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 1426 32 MAR PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil - Downloaded: 2026-02-26T09:20Z
Check the source to verify that this is the current version before use.