A-A-53537
22 June 1988

COMMERCIAL ITEM DESCRIPTION

SPLINT, ALL PURPOSE, EMERGENCY TRANSPORT

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a universal type splint used to immobilize most arm or leg injuries for emergency transport.

Salient characteristics.

Shall be a universal type transportation splint, adjustable to fit most arm and leg sizes, and suitable for use in stabilizing injuries for patient transport to a treatment center.

The splint shall consist of three curved plates, an "L" bracket, a distal piece with formed heel rest, four cravats and securing hardware. Tools shall not be required for assembly or usage of the splint.

The curved plates and formed heel rest shall be fabricated of 3003H14 aluminum 0.063 inches thick. The "L" bracket and distal piece shall be fabricated of 5052H32 aluminum 0.160 inches thick. All hardware shall be fabricated of aluminum. All material shall be radio-transparent to allow for X-ray of immobilized member, and shall be anodized or similar finish for a pleasing professional appearance.

The curved plates shall each be approximately 12 inches long by 4-1/2 inches wide by 0.063 inches thick with a depth of curvature at the center of approximately 1-1/4 inches. Three plates clearly and permanently numbered 1, 2, and 3 shall be provided. The three plates shall nest together with plate 2 fitting into plate 3 and plate 1 into plate 2. Plates 1 and 3 shall have a 1 inch wide by 5-1/4 inch long longitudinal slot approximately 1-1/4 inch in from each side.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Plate 2 shall have four square holes, one in each corner, sized to accept the carriage bolt hardware, and positioned to mate with the slots in plates 1 and 3. When nested together and secured by the carriage bolt hardware, the plates shall slide within each other to extend or retract to set the desired splint length. Plate 3 shall have two square holes located on the longitudinal centerline at the distal (heel) end for mounting the "L" bracket. Corners of the plates shall have a radius of curvature of approximately 1 inch.

The "L" bracket shall be 1-1/2 inches wide by 0.160 inches thick, bent to form a 90 degree angle with legs approximately 7 inches and 2-1/2 inches long. The 7 inch leg shall have a 6 inch long, 1/4 inch wide, longitudinally centered slot to mate with the two square holes in the distal end of plate 3 to allow longitudinal adjustment of the distal piece. The 2-1/2 inch leg shall have a 1 inch long by 1/4 inch wide longitudinally centered slot to coincide with the slot in the distal piece for vertical adjustment. Corners of the bracket shall have an approximate radius of curvature of 3/8 inches.

The distal piece shall be a 0.160 inch thick plate with a formed heel piece attached to one end. The plate shall be approximately 7 inches long by 1-1/2 inches wide with a 1/4 inch centrally located longitudinal slot 3-1/2 inches long starting 1/2 inch from the end containing the heel piece. The plate corners shall have an approximate radius of curvature of 3/8 inch.

The heel piece shall be approximately 2-1/2 inches wide by 2-1/2 inches long by 0.060 inches thick. The first 1-1/4 inches shall be flat and contain a 1/4 inch wide, centrally located longitudinal open slot 1-1/4 inches long. The remaining 1-1/4 inches shall be made up of the formed heel stop, contoured to the shape of the heel of a shoe. The flat portion shall be riveted to the distal piece in such a way that the 1/4 inch slots in each piece coincide, and the back of the heel piece is even with the end of the distal piece.

The distal piece shall be attached to the inside of the 2-1/2 inch leg of the "L" bracket by a 1-1/4 inch long 1/4-20 carriage bolt.

Cravats shall be 38 inches by 38 inches by 50 inches. Shall be of unbleached natural color muslin 60 x 48 mesh.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Hardware to assemble the splint shall consist of one 1/4-20 carriage bolt 1-1/4 inches long, six 1/4-20 carriage bolts 3/4 inches long, seven 1/4-20 wing nuts, and seven 1/4 inch flat washers 11/16 O.D. Spare hardware consisting of two carriage bolts 1-1/4 inches long, two carriage bolts 3/4 inches long, two wing nuts, and two washers shall be provided with each splint.

An illustrated instruction sheet describing the use of the splint shall be provided with each package of splints.

Workmanship. The splint shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing two splints, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory inspection without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one splint.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-174-9917

Ordering data. (Intermediate exterior package quantities and labelling and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPOAS

Project No. 6515-4296
Location: List = 6

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-55537	SPLINT, ALL PURPOSE, EMERGENCY TRANSPORT

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
	☐ USER
3b. ADDRESS (Street, City, State, ZIP Code)	☐ MANUFACTURER
	☐ OTHER (Specify)

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) – Optional	7b. WORK TELEPHONE NUMBER (Include Area Code) – Optional
7c. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8. DATE OF SUBMISSION (YYMMDD)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (DO NOT STAPLE), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation was ambiguous, restrictive, unenforceable, too rigid, inconsistent with commercial practices, or which specified processes and procedures. Enter proposed wording changes which would alleviate the problems. Entering data in blocks 1, 2, 3, and 4 will facilitate processing of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

Fold along this line

Fold along this line

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 1966 WASHINGTON D.C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DSCP-RT
2800 South 20th Street

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.