A-A-53547A
16 September 1991
SUPERSEDING
A-A-53547
6 July 1988

COMMERCIAL ITEM DESCRIPTION

ELEVATOR, PERIOSTEAL (Key)

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-E-37995(DM), which is cancelled.

This commercial item description covers a key style periosteal elevator suitable for use in the elevation of periosteum and muscle from bone.

Salient characteristics:

General. The elevator shall be a Key style periosteal elevator, with a slightly curved blade and a straight shank and handle. The lateral blade edge shall be beveled and the distal blade edge shall be beveled and rounded. The distal blade edge shall be slightly sharp.

Material. The material of the elevator shall be Class 4 corrosion-resistant steel in accordance with ASTM Designation F899.

Hardness. The elevator shall have a Rockwell Hardness C Scale (HRC) of 45 to 52. The hardness shall be tested in accordance with ASTM Designation E18 or ASTM Designations E92 and E140.

Style, design and dimensions. The elevator shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 22327, for style, design and dimensions.

Finish. The finish of the elevator shall be mirror, satin or matte.

Passivation. The elevator shall be properly passivated.

Resistance to corrosion. The elevator shall show no signs of corrosion when tested in accordance to ASTM Designation F1089.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Identification marking. The elevator shall bear the contractor's name or registered trademark and the word "stainless". Elevators of foreign manufacture shall bear the name of the country of manufacture.

Workmanship. The elevator shall be free from defects which detract which detract from its appearance or impair its serviceability.

Unit. Each (EA). One elevator, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination(such as visual or auditory investigation without the use of special laboratoryappliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with and inspection level of II and an AQL of 1.0 (percent defective). Unit of product for examination purposes shall be one elevator. Defects shall be those listed below:

Defects

Elevator no free from cracks, pits, burs, sharp edges, scale, rough areas or corrosion.

Finish of the elevator not as specified.

Style, design and dimensions not as specified.

Sampling for tests. Sampling for tests shall be in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one elevator. AQL shall be 1.0 (percent defective) with an inspection level of S-2.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-139-8928	Elevator, Periosteal, Key, 6-7/8 inch (17 cm) by 1/4 inch (6 cm) blade width
6515-01-139-8996	Elevator, Periosteal, Key, 7-1/4 inch (18.4 cm) by 1/2 inch (1.27 cm) blade width
6515-01-139-8997	Elevator, Periosteal, Key, 7-5/8 inch (19.4 cm) by 3/4 inch (1.9 cm) blade width
6515-01-168-8107	Elevator, Periosteal, Key, 8 inch (20.3 cm) by 1 inch (2.54 cm) blade width

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-5297
Location: C:\word\finals\S-1
Brown Disk #4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.