MIL-R-6018C (ASG)
7 OCTOBER 1966
Superseding
MIL-R-6018B(ASG)
23 October 1961

MILITARY SPECIFICATION

REGULATOR, OXYGEN, DILUTER DEMAND

This specification has been approved by the Department of the Air Force and by the Naval Air Systems Command.

1. SCOPE

1.1 This specification covers a diluter demand oxygen regulator.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

O-T-236	Tetrachloroethylene, Perchloroethylene, Technical Grade
O-T-634	Trichloroethylene, Technical
QQ-P-416	Plating, Cadmium (Electrodeposited)
PPP-B-591	Box, Fiberboard, Wood-Cleated
PPP-B-636	Box, Fiberboard
PPP-B-640	Box, Fiberboard, Corrugated, Triple-Wall
PPP-T-60	Tape; Pressure-Sensitive Adhesive, Waterproof, for Packaging

Military

MIL-P-116	Preservation, Methods of
MIL-A-8625	Anodic Coatings, for Aluminum and Aluminum Alloys
MIL-F-22191	Films, Transparent, Flexible, Heat Sealable, for Packaging Applications
MIL-O-27210	Oxygen, Aviator's Breathing, Liquid and Gas

STANDARDS

Military

MIL-STD-100	Engineering Drawing Practices
MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes

FSC 1660

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T06:02Z
Check the source to verify that this is the current version before use.

MIL-STD-129	Marking for Shipment and Storage
MIL-STD-130	Identification of US Military Property
MS20913	Plug, Square Head Pipe Thread
MS33586	Metals, Definition of Dissimilar
AN6004	Regulator - Diluter Demand Oxygen

(Copies of specifications, standards, drawings, and publications required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Qualification.- The regulator furnished under this specification shall be a product which has been tested and has passed the qualification tests specified herein, and has been listed on or approved for listing on the applicable Qualified Products List.

3.2 Materials.-

3.2.1 Nonferrous materials.- Nonferrous materials shall be used for all parts of the regulator except where ferrous materials are essential.

3.2.2 Nonmetallic materials.- Any nonmetallic material that deteriorates or is adversely affected by continued use with oxygen shall not be used.

3.2.3 Dissimilar metals.- Unless suitably protected against electrolytic corrosion, dissimilar metals as defined in MS33586 shall not be used in intimate contact with each other.

3.2.4 Elastomers.- All elastomers, except silicone, shall have been manufactured not more than 12 months prior to the date of regulator manufacture. The cure date (quarter and year) shall be stamped on the elastomer component, provided there is sufficient area. Cure dates and age limitations shall not apply to the silicone types. The demand diaphragm shall be fabricated of an approved silicone compound or silicone-coated cloth. All elastomers shall be free from foreign agents which may cause objectionable or intolerable odors.

3.2.5 Standard parts.- Standard parts (MS or AN) shall be used wherever they are suitable for the purpose, and shall be identified on the drawing by their part numbers. Commercial utility parts such as screws, bolts, nuts, and cotter pins may be used, provided they possess suitable properties and are replaceable by the standard parts without alteration, and provided the corresponding standard part numbers are referenced in the parts list and, if practicable, on the contractor's drawings. In the event there is no suitable corresponding standard part in effect on date of invitation for bids, commercial parts may be used provided they conform to all requirements of this specification.

3.2.6 Interchangeability.- All parts having the same manufacturing part number shall be functionally and dimensionally interchangeable. Changes in manufacturer's part numbers shall be in accordance with drawing number requirements of MIL-STD-100.

Source: https://assist.dla.mil -- Downloaded: 20260-17T06:02Z
Check the source to verify that this is the current version before use.

3.3 Design.- The regulator shall be a diluter demand type that automatically mixes ambient air with supplied oxygen or provides continuous flow of 100 percent oxygen when so adjusted. The regulator shall operate on pressures ranging from 50 pounds per square inch gage (psig) to 1,800 psig. The configuration and dimensions shall be in accordance with AN6004. The relative position of the inlet, outlet, flow indicator connection, manual controls, mounting plate and mounting holes shall be as shown on AN6004.

3.3.1 Inlet screens.-

3.3.1.1 Oxygen inlet.- A 200-mesh screen or an equivalent filter shall be installed in the oxygen inlet so as to filter all supplied oxygen entering the regulator.

3.3.1.2 Ambient air inlet screen.- A screen not finer than 100 mesh nor coarser then 30 mesh shall be installed in the ambient air inlet so as to filter all ambient air that mixes with oxygen.

3.3.2 Flow indicator connection.- The regulator shall contain a 1/8-inch, female pipe thread, flow indicator connection. The connection shall be connected to a chamber in the regulator in which the pressure varies at least 0.5 psig in the pressure range of 0 to 10 psig as a direct result of a flow in comparison to no flow through the regulator outlet. The flow through the outlet causing the variation in the pressure shall be as a result of applying a negative pressure (see 6.3.5) to the outlet with the diluter control in the normal oxygen position and in the 100-percent oxygen position. The pressure in the chamber shall not exceed 25 psig. The chamber shall be so located that the oxygen entrapped in the flow indicator line will not be lost during the exhalation cycle.

3.3.3 Emergency continuous flow bypass.- The regulator shall contain an independent bypass around the regulating and mixing mechanism that will provide a 14 ±1.5 liters per minute(lpm)flow of 100-percent oxygen through the regulator outlet. The flow through the passage shall be controlled by an emergency control as shown on AN6004. A counterclockwise rotation of the control shall permit flow and a clockwise rotation shall stop flow. The rotational movement of the emergency control shall be 180 ±20 degrees.

3.3.4 Diluter control.- The regulator shall contain a diluter control as shown on AN6004 which provides a mixture of supplied oxygen and ambient air or 100 percent supplied oxygen through the outlet when a negative pressure is applied to the outlet. Stops shall be provided to limit the motion of the diluter control within 5 degrees of the minimum cited on AN6004. Tension of the diluter control shall be great enough so there will be no possibility of a change of setting during the endurance and vibration tests specified herein. The length of the diluter control lever, when placed in either setting, shall be within the profile of the regulator.

3.4 Construction.- The regulator shall be so constructed that no parts will work loose in service and to withstand the strains, jars, vibrations and other conditions incident to shipping, storage, installation, and service. The regulator shall be thoroughly cleaned and free of all foreign particles or matter. The regulator shall emit no odor.

Source: https://assist.dla.mil -- Downloaded: 202640-17T06:02Z
Check the source to verify that this is the currentversion before use.

3.4.1 The regulator shall be cleaned by using a vapor phase degreasing cleaner in accordance with O-T-236 or O-T-634. Ultrasonics may be used in conjunction with vapor phase cleaning. After cleaning, a General Electric, type H leak detector, or an equivalent halide testing apparatus, shall be used to determine the presence of cleaning compound.

3.4.2 Opening closure.- The inlet and outlet shall be closed with caps or plugs to prevent dust and foreign matter from entering the regulator during shipping and storage. The threaded plugs shall conform to MS20913, except that the material may be a plastic compound suitable to the procuring activity. The material and design of all other caps or plugs shall be subject to the approval of the procuring activity.

3.5 Performance.- The performance of the regulator shall be as determined by the design requirements specified herein and as required to pass all the tests specified herein. The regulator shall pass the flow characteristics and oxygen air ratio leak tests when the regulator diaphragm is horizontal or vertical.

3.6 Weights.- The weight of the regulator shall not exceed 2 pounds.

3.7 Finish.-

3.7.1 Protective treatment.- When materials are used in the construction of the regulator that are subject to corrosion in oxygen, salt air, or other atmospheric conditions likely to occur during service usage, they shall be protected against such corrosion in a manner that will in no way prevent compliance with the performance requirements of this specification. The use of any protective coating that will crack, chip, or scale with age or extremes of atmospheric conditions shall be avoided.

3.7.2 Regulator exterior.- The exterior of the regulator shall be finished in a durable dull black.

3.7.3 Aluminum and aluminum alloys.- Aluminum and aluminum alloy parts shall be anodized in accordance with MIL-A-8625.

3.7.4 Cadmium plating.- Cadmium plating, if used, shall be in accordance with QQ-P-416.

3.8 Marking.-

3.8.1 Special.- The regulators shall be marked as required by AN6004.

3.8.2 Identification.- The regulator shall be marked for identification as required by MIL-STD-130 except the Federal Stock Number shall be omitted.

3.8.3 Date.- The date of manufacture, consisting of the month and year, shall be legibly and permanently stamped on a protected section of the case. A rubber stamp and inspector's ink shall be used. The stamp shall be in the form of a 1/2-inch square with a horizontal line dividing the square into two segments. The upper segment shall contain the number of the month, and the lowest segment shall contain the last two numbers of the year.

Source: https://assist.dla.mil -- Downloaded: 20261O-17T06:02Z
Check the source to verify that this is the current version before use.

3.9 Workmanship.- The regulator shall be uniform in quality and shall be free from irregularities, defects, or foreign matter which could adversely affect safety, performance, reliability, or durability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsiblity for inspection.- Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.2 Classification of inspection.- The inspections to be performed are:

(a) Qualification inspection (see 4.3 and 6.3.1)
(b) Quality conformance inspection (see 4.4 and 6.3.2)

4.3 Qualification inspection.- Qualification inspection shall consist of performing all examinations and tests described in this specification, except examination of preparation for delivery (4.6.15).

4.3.1 Qualification samples.- Qualification samples shall consist of three regulators complete with plugs and caps. The samples shall be identified by securely attached durable tags and shipped to the Supply Officer, Naval Air Engineering Center, Attention: Director, Aerospace Crew Equipment Laboratory, Philadelphia, Pennsylvania 19112. The attached tags shall contain the following information:

Samples submitted by (name) (date) for qualification inspection in accordance with MIL-R-6018C under authorization (reference authorizing letter and number). (See 6.5.)

4.4 Quality conformance inspection.- Quality conformance inspection shall consist of performing the examinations and tests listed in table I.

4.4.1 Quality conformance inspection lot.- For the purpose of selecting production sample regulators to be subjected to quality conformance inspection, a lot shall be in accordance with MIL-STD-105.

4.4.2 Quality conformance samples.- Quality conformance samples shall be selected in accordance with MIL-STD-105. The sample unit shall be one regulator assembly. The number of samples shall be as specified in table I.

4.4.3 Quality conformance acceptance or rejection.- Quality conformance acceptance or rejection shall be in accordance with MIL-STD-105 and the acceptance criteria specified in table I.

4.5 Inspection conditions.-

4.5.1 Gas.- The gas used in performing the inspection specified herein shall be type I oxygen conforming to MIL-O-27210.

Source: https://assist.dla.mil -- Downloaded: 202640-17T06:02Z
Check the source to verify that this is the current version before use.

TABLE I. Quality conformance inspections, sample size, and acceptance criteria

Inspection	Method para.No.	Sample size	Acceptance criteria
Examination of regulator	4.6.1	Each regulator	Reject all defective regulators
Examination of configuration, weight, and dimension	4.6.2	1/ Inspection level S-2	Acceptance number zero, rejection number one
Flow characteristic	4.6.3	Each regulator	Reject all defective regulators
Oxygen air ratio	4.6.4	Each regulator	Reject all defective regulators
Emergency continuous flow	4.6.5	Each regulator	Reject all defective regulators
Flow indicator provision	4.6.6	One regulator of every 25 regulators	Acceptance number zero, rejection number one
Odor	4.6.7	1/ Inspection level S-2	Acceptance number zero, rejection number one
Pressure relief	4.6.8	Each regulator	Reject all defective regulators
Outlet leakage, negative pressure	4.6.9	Each regulator	Reject all defective regulators
Outlet leakage, positive	4.6.10	Each regulator	Reject all defective regulators
Leakage through the regulator	4.6.11	Each regulator	Reject all defective regulators
High temperature exposure (160°)	4.6.13	1/ Inspection level S-2	Acceptance number zero, rejection number one
Examination of preparation for delivery	4.6.15	Inspection level S-2	Acceptance quality level of 4.0 percent defective

1/ Sample size shall be based only on the applicable sample code letter corresponding to the specified inspection level of MIL-STD-105.

4.5.2 Temperature and pressure:- Unless otherwise specified, tests shall be conducted at local ambient temperature and barometric pressure. Test instruments shall be calibrated or adjusted according to their required usage in conducting individual tests. Any test results obtained at conditions other than normal temperature and pressure (NTP) conditions shall be mathematically corrected to NTP conditions. NTP conditions are 29.92 inches of mercury and 70° F. In case of pressure drop measurements, the flow setting for conducting the test shall be corrected to NTP conditions before the test is conducted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T06:02Z
Check the source to verify that this is the current version before use.

4.5.3 Vibration or tapping.- Except where specified, inspections shall be performed without subjecting the regulator to vibration or tapping.

4.5.4 Flow measurement.- The flow shall be measured by either a piezometer tap or ring or by a pressure tap designed to measure static pressures in a straight outlet connection assembly. The outlet connection assembly shall consist of a connector (specified by the procuring activity) joined by a short piece of rubber tubing to a straight metal tube 2 to 3 inches long with a 0.75- to 0.775-inch smooth inside diameter (ID). The tube shall be fitted with a pressure connection which communicates with the inside of the metal tube by a vent or vents whose axes are perpendicular to the direction of flow. The total distance to the pressure vent or vents from the edge of the regulator outlet shall not exceed 3 inches. At least 1 inch of smooth tube surface shall be allowed on either side of the piezometer vent or vents.

4.5.5 Orientation.- When the regulators are subjected to tests specified in 4.6.3 and 4.6.4 regulators shall be placed with the mounting plate on the bottom and the diaphragm horizontal. When conducting the tests specified in 4.6.3 and 4.6.4 during qualification inspection, alternate regulators shall be placed with the diaphragm in the vertical position.

4.6 Inspection methods.-

4.6.1 Examination of regulator.- The regulator shall be examined to determine conformance to this specification with respect to materials, proper assembly, required closure of inlet and outlet ports (see 3.4.2), operability of controls, color, marking, workmanship, and other requirements not covered by specific inspection.

4.6.2 Examination of configuration, weight, and dimensions.- The regulator shall be examined to determine conformance of configuration and dimension to AN6004. The regulator shall be weighed on springless scales graduated to one ounce. The weight shall be as specified in 3.6.

4.6.3 Flow characteristics.- With the emergency flow control off, a supply of oxygen at the pressures specified in table II shall be applied to the regulator oxygen inlet. A negative pressure shall be applied to the regulator outlet to obtain the three increments of flow specified in table II. All measurements of the outlet pressure shall be made directly at the regulator outlet. The performance of the flow characteristics test at the three increments of outlet flow shall be conducted with the diluter control in the normal oxygen position and repeated with the diluter control in the 100 percent oxygen position. The flow test shall be performed with the diluter control in both positions at an ambient equivalent to each altitude specified in table III. The negative pressure measured at the regulator shall not exceed the maximum values specified in table II for the applicable outlet flow for each altitude. After applying the negative pressure required to obtain the three increments of flow and reducing this suction to zero, the flow through the outlet shall not exceed 0.01 lpm.

Source: https://assist.dla.mil -- Downloaded: 2026-10-17T06:02Z
Check the source to verify that this is the current version before use.

4.6.4 Oxygen air ratio.- With an oxygen supply of 150 psig connected to the regulator oxygen inlet, the regulator shall be installed in an approved test apparatus and the ratio of supplied oxygen to ambient air that flows through the outlet determined. The ratio shall be determined with the emergency flow control in the "off" position and the diluter control in the normal oxygen position and in the 100-percent oxygen position at an ambient pressure equal to each altitude specified in table III. This test shall be conducted at 10, 40, 75, and 100 lpm. The ratio of supplied oxygen to ambient air shall be as specified in table III. The apparatus used in determining the ratio shall be as approved by the procuring activity. The oxygen air ratio test may be conducted concurrently with flow characteristics test (4.6.3).

TABLE II. Flow characteristics

Supply oxygen pressure	Flow of oxygen or oxygen-air mixture	Outlet negative pressure inches of water
PSI	Liters per minute ambient pressure	Maximum
50 to 1,800	0 to 30	0.40
50 to 1,800	1/ 50	0.97
150 to 1,800	2/ 85	1.50

1/ 40 below 15,000 feet.
2/ 70 below 15,000 feet.

TABLE III. Ratio of cylinder oxygen to total gas delivered

Altitude feet	Total volume gas flow of 10 ambient liters per minute	Total volume gas flow at 20 to 50 ambient liters per minute	Total volume gas flow at 51 to 85 ambient liters per minute	Total volume gas flow above 85 ambient liters per minute
5,000	0.01 to 1.00	0.01 to 0.30	0.01 to 0.40	0.01 to 1.00
10,000	.06 to 1.00	.06 to .30	.06 to .40	.06 to 1.00
15,000	.14 to 1.00	.14 to .32	.14 to .40	.14 to 1.00
20,000	.24 to 1.00	.24 to .46	.24 to .65	.24 to 1.00
25,000	.40 to 1.00	.40 to .80	.40 to .90	.40 to 1.00
28,000	.50 to 1.00	.50 to .90	.50 to .95	.50 to 1.00
30,000	.60 to 1.00	.60 to 1.00	.60 to 1.00	.60 to 1.00
32,000	.98 to 1.00	.98 to 1.00	.98 to 1.00	.98 to 1.00
With diluter at 100 percent oxygen				
All altitudes	0.98 to 1.00	0.98 to 1.00	0.98 to 1.00	0.98 to 1.00

4.6.5 Emergency continuous flow.- An oxygen supply of 100 psig shall be connected to the regulator oxygen inlet and the emergency flow control turned on. With the emergency flow control on, flow through the outlet shall be measured with the diluter control in the normal oxygen position and in the 100-percent oxygen position. The flow under these conditions shall be 14 ±1.5 lpm.

Source: https://assist.dla.mil -- Downloaded: 202610-17T06:02Z
Check the source to verify that this is the current version before use.

4.6.6 Flow indicator provision.- During the performance of tests specified in 4.6.3 and 4.6.5, the pressure and the variation in the pressure at the indicator connection shall be measured. The pressure and the variation in pressure shall conform to 3.3.2.

4.6.7 Odor.- An oxygen supply of at least 50 psig shall be connected to the regulator oxygen inlet and an oxygen mask of the type designed for use with a diluter demand regulator shall be connected to the regulator. The oxygen mask and the mask-to-regulator hose shall be thoroughly examined to assure they have no objectionable odor. Six subjects shall breathe through the mask and regulator for 5 minutes per each subject. One-half the breathing time for each subject shall be with the diluter control in the normal oxygen position and one-half with the diluter control in the 100-percent oxygen position. Any objectionable odor reported by any subject shall be cause for rejection.

4.6.8 Pressure relief.- The emergency flow control shall be placed in the "off" position and the diluter control placed in the 100-percent oxygen position. A source of pressure shall be connected to the regulator outlet. Mercury manometer shall be connected to the outlet to indicate pressure in the outlet. The pressure shall be increased until a flow of 10 lpm is established through the relief valve vent. The outlet pressure at this flow shall be measured and shall be at least 3 inches of mercury. The pressure in the outlet shall be reduced to a minimum of 1.1 inches of mercury and the flow through the relief valve vent measured. The flow shall not exceed 0.01 lpm.

4.6.9 Outlet leakage negative pressure.- The oxygen inlet port shall be closed to prevent the gas from entering the regulator and the emergency flow control shall be closed. With the diluter control in the 100-percent oxygen position, a negative pressure of at least 1 inch of water shall be applied to the outlet and the flow from within the regulator through the outlet shall be measured. The flow shall be less than 0.5 lpm.

4.6.10 Outlet leakage positive pressure.- With the emergency flow control closed and the diluter control in either position, a source of positive pressure shall be applied to the regulator outlet. The pressure shall be increased to 2 inches of water and the internal leakage of the regulator measured. The leakage shall not exceed 0.1 lpm. The positive pressure at the outlet shall then be increased to 10 inches of water and the internal leakage of the regulator measured. The leakage shall not exceed 0.5 lpm.

4.6.11 Leakage through regulator.- With the emergency flow control closed, a pressure of 500 psig shall be applied to the regulator oxygen inlet. Leakage through the regulator shall be measured and shall not exceed 0.01 lpm.

4.6.12 Low temperature operation.- The regulator shall be subjected to a temperature of -80° F for 48 hours. After 48 hours' exposure, the regulator shall be subjected to a temperature of -65° F for 3 hours. After 3 hours' exposure and with an ambient temperature of -65° F maintained, the regulator shall be cyclic operated to simulate normal respiration for 2 minutes. After the cyclic operation and with the ambient temperature of -65° F continuously maintained, the regulator shall be subjected to and pass the inspections specified in 4.6.3, 4.6.4, 4.6.5, 4.6.6, 4.6.8, 4.6.9, 4.6.10, and 4.6.11.

Source: https://assist.dla.mil -- Downloaded: 20260-17T06:02Z
Check the source to verify that this is the current version before use.

4.6.13 High temperature exposure.- With the emergency flow control closed, and a source of oxygen under 500 psig pressure connected to the regulator oxygen inlet,the regulator shall be exposed to a temperature of 160° ±2° F for 7 hours. After this exposure and while still at a temperature of 160° F, the regulator shall be cycled for 15 simulated breathing cycles and then subjected to and pass the leakage test specified in 4.6.11. After passing the test specified in 4.6.11, the temperature of the regulator shall be reduced to local normal temperature and subjected to and pass the tests specified in 4.6.3, 4.6.4, 4.6.5, 4.6.6, 4.6.7, 4.6.8, 4.6.9, 4.6.10, and 4.6.11.

4.6.14 Endurance and vibration.- With a supply of oxygen connected to the regulator oxygen inlet and the oxygen maintained at a minimum pressure of 50 psig, the regulator shall be subjected to an endurance-vibration test. The emergency flow control shall be closed and the regulator shall be operated for 50,000 cycles with the diluter control in the 100-percent oxygen position and for 50,000 cycles with the diluter control in the normal oxygen position. A cycle shall consist of applying a negative pressure to the regulator outlet that will cause a flow through the regulator, and then applying a positive pressure to the regulator outlet. The negative pressure shall cause a peak flow of 30 lpm for 90 percent of the cycles and a peak flow of 90 lpm for 10 percnet of the cycles. The positive pressure shall be 1/2 inch of water for 90 percent of the cycles and 1 inch of water for 10 percent of the cycles. The regulator shall be vibrated during the cycling for 3 hours at a frequency range from 8 to 42 cycles per second with a double amplitude range from 0.018 to 0.020 inch. After completing 90 percent of the vibration and with the vibration continuing, the regulator shall be subjected to and pass the tests specified in 4.6.3 and 4.6.4. After completing the cycling, the regulator shall be subjected to and pass the tests specified in 4.6.1, 4.6.3, 4.6.4, 4.6.5, 4.6.8, 4.6.9, 4.6.10, and 4.6.11.

4.6.15 Examination of preparation for delivery.- The packaging and packing of the regulators shall be examined to determine conformance or nonconformance to the preservation, packaging, and packing specified in the contract or order.

5. PREPARATION FOR DELIVERY

5.1 Packaging.- Packaging shall be level A or C, as specified (see 6.2).

5.1.1 Level A.- Each breathing regulator shall be preserved in accordance with Method 1C-4 of MIL-P-116. No contact preservation is required. Each breathing regulator shall be sealed within a snug-fitting bag of 0.0015 inch gage material conforming to type III of MIL-F-22191. The sealed breathing regulator shall be cushioned and packaged within a waterproof fiber can with sealed metal ends.The breathing regulator shall be cushioned in such a manner as to prevent damage to itself. Ten unit packages shall be packaged within a fiberboard container conforming to type I, class 2, grade 3 of PPP-B-636. The manufacturer's seam shall be adhesive sealed. All seams and joints shall be sealed with tape conforming to PPP-T-60.

5.1.2 Level C.- The breathing regulator shall be packaged to afford the minimum degree of protection necessary to prevent deterioration or damage during shipment under normal environmental conditions and commercial modes of transportation.

5.2 Packing.- Packing shall be levels A, B, or C, as specified (see 6.2).

Source: https://assist.dla.mil -- Downloaded: 202640-17T06:02Z
Check the source to verify that this is the currentversion before use.

5.2.1 Level A.- One hundred breathing regulators, packaged as specified in 5.1.1, shall be packed in overseas type exterior containers conforming to PPP-B-591, PPP-B-636, or class 2 of PPP-B-640. Closure and strapping of the containers shall be in accordance with the applicable container specification or appendix thereto. The gross weight of the packed shipping container shall not exceed the weight limitations of the applicable container specification.

5.2.2 Level B.- One hundred breathing regulators, packaged as specified in 5.1.1, shall be packed in domestic exterior containers conforming to PPP-B-591, PPP-B-636, or class 1 of PPP-B-640. Closure and strapping of the containers shall be in accordance with the applicable container specification or appendix thereto. The gross weight of the packed shipping container shall not exceed the weight limitations of the applicable container specification.

5.2.3 Level C.- The packaged breathing regulators, which the carrier requires to be packed, shall be packed within exterior type shipping containers in a manner that will insure safe transportation at the lowest rate to the point of delivery. The shipment shall conform to the minimum requirements of the rules and regulations applicable to the mode of transportation selected.

5.3 Marking.- Nomenclature tags shall be inserted in each bag showing title, part number or other data as required by the procuring activity. In addition, interior and exterior containers shall be marked in accordance with MIL-STD-129.

5.3.1 Precautionary marking.- The following precautionary marking shall appear on each package:

"CAUTION: DO NOT ALLOW CONTAMINANTS OF ANY KIND TO BE USED ON OR ABOUT THE OXYGEN REGULATORS."

6. NOTES

6.1 Intended use.- The regulator is intended for use at altitudes where supplemental oxygen is necessary for human respiration. The regulator is intended to furnish breathing air in quantity as demanded by inhalation.

6.2 Ordering data.- Procurement documents should specify:

(a) Title, number, and date of this specification.
(b) The levels of preservation, packaging, and packing required (see 5.1 and 5.2).
(c) Data requirements (see 6.4).

6.3 Definitions.-

6.3.1 Qualification inspection.- Qualification inspections, as used in this specification, is the process of examining and testing, prior to award of contract or order, a manufacturer's product to determine if the product meets the requirements of this specification and is acceptable for listing on the Qualified Products List.

Source: https://assist.dla.mil -- Downloaded: 202640-17T06:02Z
Check the source to verify that this is the current version before use.

6.3.2 Quality conformance inspection.- Quality conformance inspection, as used in this specification, is the process of examining and testing individual production line items or samples of production line items to determine if, according to the requirements of this specification, the items are acceptable for delivery to the Government.

6.3.3 Flow measurements.- All reference in this specification to gas flow at altitudes will be understood to refer to volume expanded gas at altitudes at room temperature, and not as corrected to standard sea level conditions.

6.3.4 Positive pressure.- Positive pressure is pressure greater than the pressure surrounding the outside of the regulator.

6.3.5 Negative pressure.- Negative pressure is a pressure less than the pressure surrounding the outside of the regulator.

6.4 Data.- For the information of contractors and contracting officers, any of the data specified in applicable documents listed in section 2 of this specification or referenced lower-tier documents need not be prepared for the Government unless specified in the contract or order. The data to be furnished should be listed on DD Form 1423 (Contractor Data Requirements List), which should be attached to and made a part of the contract or order. NavWeps Form 4200/15 (Drawings, Lists, and Specifications Required) should be attached where applicable.

6.5 Qualification.- With respect to products requiring qualification, awards will be made only for such products as have, prior to the time set for opening of bids, been tested and approved for inclusion in the applicable Qualified Products List, whether or not such products have actually been so listed by that date. The attention of the suppliers is called to this requirement, and manufacturers are urged to arrange to have the products that they propose to offer to the Federal Goverernment tested for qualification in order that they may be eligible to be awarded contracts or orders for the products covered by this specification. The activity responsible for the Qualified Products List is the Commander, Naval Air Systems Command, Dept.of the Navy, Washington, D. C. 20360; however, information pertaining to qualification of products may be obtained from the Director, Aerospace Crew Equipment Laboratory, Naval Air Engineeing Center, Philadelphia, Pennsylvania 19112.

Custodians:
Navy - AS
Air Force - 71

Reviewer activities:
Navy - AS
Air Force - 11, 71

Preparing activity:
Navy - AS

Source: https://assist.dla.mil -- Downloaded: 20261O-17T06:02Z
Check the source to verify that this is the current version before use.