A-A-58047B
07 December 2004
SUPERSEDING
A-A-58047A
30 July 2002

COMMERCIAL ITEM DESCRIPTION

GENERATOR, SIGNAL
(10 MHz to 26.5 GHz)

The General Services Administration has authorized the use of this commercial item description for all federal agencies.

1. SCOPE. This Commercial Item Description (CID) describes a solid state, general purpose, 10 MHz to 26.5 GHz, signal generator. This CID is meant as a minimum requirement for the signal generator in which only those manufacturers that meet or surpass the following requirements are qualified per this CID.

2. SALIENT CHARACTERISTICS. The equipment shall be capable of operation and product sample testing, if so requested, within the accuracies, limits, and specifications herein.

2.1 Classification. Equipment covered by this CID may be commercially available equipment modified to the extent necessary to meet the following description. The equipment shall be Class 3 in accordance with MIL-PRF-28800, except as specified herein.

2.2 Safety and Environmental. The equipment shall meet all safety and environmental requirements as specified in MIL-PRF-28800 for the classification as stated herein.

2.2.1 Temperature Operating. The equipment shall conform to its specified performance and accuracy while being operated as a volt meter in a temperature range of 0 °C to 50 °C.

2.2.2 Temperature Non-Operating. The equipment shall conform to its specified performance and accuracy after being stored in an environment in the temperature range of -40 °C to 71 °C.

Comments, suggestions, or questions on this document should be addressed to: WR-ALC/LEEC 295 Byron Street, Robins AFB, GA 31098-1611. Since contact information can change; you may want to verify the currency of this address information using the ASSIST Online database at http://assist.daps.dla.mil

FSC 6625

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:59Z
Check the source to verify that this is the current version before use.

2.3 Electrical Power Sources. The equipment shall operate from nominal single-phase commercial, military, and shipboard power sources of 110 and 220 volts AC (± 10%) at line frequencies of 48 Hz to 63 Hz. Transient state conditions shall meet MIL-PRF-28800 requirements for the classification specified herein.

2.4 Reliability. The design of the equipment shall be such that under normal use and operation the equipment does not fail within 4,000 hours of operation with a statistical certainty of 95%.

2.5 Calibration and Maintenance Adjustments. The design of the system shall provide for readily accessible calibration and maintenance adjustments. These adjustments shall be provided by variable value components, which are adjustable, by the use of simple means. The calibration by substitution of selected components or parts is unacceptable unless specifically approved. The calibration adjustments, wherever possible, shall be accessible without removal of the instrument case or modules. The calibration interval shall be a period of one year or greater, based on an operating time of 2000 hours per year.

2.6 Performance Requirements. The instrument shall meet all requirements specified herein, after the warm-up period specified in MIL-PRF-28800, under all combinations of input power conditions, output signal conditions, and operating service condition as specified in MIL-PRF-28800.

2.6.1 Frequency Range. The minimum frequency range of the signal generator shall be 10 MHz to 26.5 GHz.

2.6.1.1 Resolution. The generator shall have a frequency resolution of ≤ 0.1 Hz.

2.6.1.2 Accuracy and Aging Rate. The generator accuracy shall be based on the internal reference oscillator aging rate which shall be at least ± 1 x 10^{-9} per day after 10 days of continuous operation.

2.6.2 Spectral Purity.

2.6.2.1 Harmonics. The generator harmonics shall be ≤ -35dBc over the entire frequency range.

2.6.2.2 Sub-Harmonics. The sub-harmonics shall be as follows:

10 MHz to 18 GHz: ≤ -25 dBc
18 GHz to 26.5 GHz: ≤ -20 dBc

2.6.2.3 Spurious Non-Harmonics. The spurious non-harmonics shall be at least -50 dBc over the entire frequency spectrum.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:59Z
Check the source to verify that this is the current version before use.

2.6.2.4 SSB Phase Noise. The generator shall have the following single sideband phase noise characteristics, at a 100 kHz offset from the carrier:

10 MHz to 2 GHz: ≤ –100 dBc/Hz
2 GHz to 10 GHz: ≤ –104 dBc/Hz
10 GHz to 26.5 GHz: ≤ –97 dBc/Hz

2.6.3 RF Output. The minimum output level range into 50 ohm nominal shall be as follows:

10 MHz to 20 GHz: +8 dBm to -99 dBm
20 GHz to 26.5 GHz: +1 dBm to -99 dBm

2.6.3.1 Output Resolution. The generator shall have an output resolution of ≥ 0.1 dB.

2.6.3.2 Output Protection. The generator shall have a reverse power protection to +20 dBm.

2.6.3.3 Output Connector. The RF output connector shall be APC -3.5 female.

2.6.3.4 Output Accuracy. The output accuracy shall be ± 2dB.

2.6.4 Amplitude Modulation (AM).

2.6.4.1 AM Depth. The generator's amplitude modulation depth shall be at least 0 to 90%.

2.6.4.2 AM Accuracy. At a 1 kHz rate and 30% depth, the accuracy shall be ≤ (± 5% of the setting +1%).

2.6.4.3 AM Bandwidth. At 30% depth with pulse off, the 3 dB point shall be 10 kHz or better.

2.6.4.4 AM Distortion. At a 1 kHz rate and 50 % depth, the distortion shall be < 5%.

2.6.5 Frequency Modulation (FM).

2.6.5.1 FM Rate. The generator shall have a frequency modulation rate of 50 kHz to 1 MHz.

2.6.5.2 FM Deviation. The generator's frequency modulation maximum peak deviation shall be as follows:

10 MHz to 1 GHz: ≤ 2 MHz
1 GHz to 2 GHz: ≤ 4 MHz
2 GHz to 3.2 GHz: ≤ 8 MHz
3.2 GHz to 10 GHz: ≤ 16 MHz
10 GHz to 20 GHz: ≤ 32 MHz

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:59Z
Check the source to verify that this is the current version before use.

2.6.5.3 FM Accuracy. The generator shall have a deviation accuracy of < (± 5% of deviation + 20 Hz).

2.6.5.4 FM Sensitivity. The generator shall have a frequency modulation sensitivity of ± $1V_{peak}$.

2.6.6 Pulse Modulation.

2.6.6.1 On/Off Ratio. The on/off ratio for the pulse modulation shall be 80 dB or greater.

2.6.6.2 Rise and Fall Time. The rise and fall time between 10% and 90% of the pulse shall be 25 nanoseconds or less.

2.6.6.3 Repetition Rate. The generator shall provide externally generated pulse modulation with a repetition rate of ≥ 100 Hz to ≤ 1 MHz with a leveled output.

2.6.6.4 Pulse Width. The minimum pulse width shall be at least 100 nanoseconds.

2.6.7 Sweep Mode. The generator shall have a sweep mode, analog and/or digital, that has a step sweep of frequency or amplitude, or both, from start to stop.

2.6.7.1 Sweep Time. The minimum sweep time shall be at least 45 milliseconds.

2.6.7.2 Sweep Range. The sweep shall include a selectable sweep range between any two power levels at a single continuous wave (CW) frequency.

2.6.7.3 Switching Speed. The generator shall switch frequencies in ≤ 50 milliseconds.

2.6.7.4 Triggering. The generator shall have the ability to trigger externally, automatically, in a single mode, or as a result of a GPIB command.

2.6.7.5 Markers. The signal generator shall have at least 5 independent markers.

2.6.8 Remote interface. The generator shall come with an IEEE 488.2 standard interface.

2.7 Physical Parameters. The generator's component dimensions and weight shall be consistent with current commercial offerings. The equipment shall have the following physical characteristics.

2.7.1 Dimensions. The width and depth of the equipment shall be compatible with mounting in an ANSI/EIA-310-D standard equipment rack. The height shall be compatible with four standard height units, 4U. One unit or "U" is defined by ANSI/EIA-310-D to be 1.75 inches or 44.45 mm.

2.7.2 Weight. The total weight of the equipment, excluding accessories and manuals, shall not exceed a two-person lift of 74 pounds, (33.6 kilograms).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:59Z
Check the source to verify that this is the current version before use.

2.7.3 Rack Mount. The analyzer shall be capable of being rack-mounted in a standard ANSI/EIA-310-D equipment rack and include a rack-mount conversion kit if needed.

2.8 Manuals. The equipment shall be delivered with operator, maintenance/service, illustrated parts, and calibration manuals. Format and quantity shall be as specified in the contract or order. Level of maintenance philosophy, as defined in MIL-PRF-28800, shall be as specified in the contract or order.

If evidence substantiating that the Mean Time Between Failures (MTBF) of the instrument exceeds 20,000 hours is submitted, the maintenance/service manual shall not be required.

3. REGULATORY REQUIREMENTS. The offeror/contractor is encouraged to use recovered materials to the maximum extent practicable, in accordance with paragraph 23.403 of the Federal Acquisition Regulation (FAR).

4. PRODUCT CONFORMANCE PROVISIONS.

4.1 Product Conformance. The products provided shall meet the salient characteristics of this CID, conform to the producer's own drawings, specifications, standards and quality assurance practices, and be the same product offered for sale in the commercial market. The government reserves the right to require proof of such conformance.

5. PACKAGING.

5.1 Preservation, Packing, and Marking. Preservation, packing, and marking shall be as specified in the contract or order.

6. NOTES.

6.1 Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, provided they fall within the specified tolerances using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this CID are met. If a product manufactured to metric dimensions exceeds the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

6.2 Sources of documents.

6.2.1 Military Specifications, Standards, and Handbooks referenced herein may be obtained at http://assist.daps.dla.mil or available from the Standardization Documents Order Desk, 700 Robbins Ave, Bldg 4, Section D, Philadelphia, PA 19111-5094.

6.2.2 The Code of Federal Regulations (CFR) may be obtained at http://www.gpoaccess.gov/cfr/index.html or available from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC, 20402.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:59Z
Check the source to verify that this is the current version before use.

6.2.3 Institute of Electrical and Electronics Engineers (IEEE) standards products may be obtained at http://www.ieee.org or available from IEEE Operations Center. Phone: (800)678-IEEE. Mail: IEEE Operations Center, Sales Office, 445 Hoes Lane, PO Box 1331, Piscataway, NJ, 08855-1331, USA.

Military Interest

Custodian:
Air Force 99

Preparing Activity:
Air Force - 84

Agent:
Air Force - 99

(Project 6625-F101)

Note: The activities listed above were interested in this document as of the date of this document. Since organizations and responsibilities can change, you should verify the currency of the information above using the ASSIST Online database at http://assist.daps.dla.mil

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:59Z
Check the source to verify that this is the current version before use.