INCH POUNDS

A-A-58057A
04 April 2008
SUPERSEDING
A-A-58057
22 March 1996

COMMERCIAL ITEM DESCRIPTION

Processing Machine, Radiographic Film, Automatic, Stationary

The General Services Administration has authorized the use of this Commercial Item Description (CID) for all federal agencies.

1. SCOPE. This Commercial Item Description (CID) is for a high capacity, processor machine, radiographic film, automatic, stationary, developing, fixing, washing and drying of all standard brands of nondestructive testing x-ray film in an industrial environment.

2. SALIENT CHARACTERISTICS

2.1 Power Requirements. The processor shall operate in accordance with requirements of this CID when supplied with 220 ± 10% Volts, 3-wire single phase, 50/60 Hz, 40 amp maximum current.

2.2 Standby Mode. The processor shall have an automatic standby mode to conserve energy, water and wear by automatically shutting down the transport, wash and dryer systems when no film is being processed.

2.3 Display. The processor shall have a digital temperature display.

2.4 Feed Indicator. The processor shall provide an indicator to alert the operator that it is ready for additional film.

2.5 Weight. The maximum weight shall not exceed 700 lbs (empty).

2.6 Size. The processor, including feed tray, shall not occupy more than 12 sq. feet.

2.7 Warm-up. The processor warm-up period shall not exceed 25 minutes.

Beneficial comments, recommendations, additions, deletions, clarifications, etc., and any data which may improve this document should be sent to: WR-ALC/GBZO-ENV, 380 Richard Ray Blvd Ste 104, Robins AFB GA 31098-1640

AMSC N/A FSC 6635

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:59Z
Check the source to verify that this is the current version before use.

2.8 Capacity. The processor shall be capable of processing at least 55 sheets, 14 in x 17 in, industrial radiographic film per hour.

2.9 Auto Replenishment System. The processor shall incorporate an automatic solution replenishment system. The replenishment system shall be adjustable.

2.10 Film. The processor shall be able to accommodate all industrial types and brands of film in sheet and roll formats up to 17 inches wide.

2.11 Chemicals. The processor shall be able to use all commercial processing chemicals.

3. REGULATORY REQUIREMENTS. The offeror/contractor in encouraged to use recovered materials to the maximum extent practicable, in accordance with paragraph 23.403 of the Federal acquisition regulation (FAR).

4. PRODUCT CONFORMANCE

4.1 Product conformance. The products provided shall meet the salient characteristics of this Commercial Item Description, conform to the producer's own drawings, specifications, standards and quality assurance practices, and be the same product offered for sale in the commercial market. The government reserves the right to require proof of such conformance.

4.2 Market Acceptability. The following market acceptability criteria are necessary to document the quality of the product to be provided under this CID.

4.2.1 The company producing the item must have been producing a processing machine, radiographic film, automatic, stationary, meeting the requirements of this CID for at least 2 years.

4.2.2 The company must have sold at least 30 units meeting this CID in the commercial market place over the past two years.

5. PACKAGING. Preservation, packing and marking shall be specified in the contract or order.

6. NOTES.

6.1 Ordering data.

a) Title, number, and date of this CID.
b) Specify special packaging requirements.
c) Contracting officer may request proof of certification of commercial item prior to first contract delivery.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:59Z
Check the source to verify that this is the current version before use.

6.2 Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, provided they fall within the tolerances specified in conversion tables contained in the latest version of FED-STD-376 and all other requirements of this CID are met. If the manufacturer elects to use metric dimensions as the basis upon which to derive or base a mechanical feature, a request shall be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

6.3 Subject term (key word) listing.

1. Film
2. Nondestructive
3. Radiographic
4. X-ray

Custodians:
Air Force - 99
DLA - GS

Preparing Activity:
Air Force - 84

Agent:
Air Force - 99

(Project 6635-2008-006)

NOTE: The activities listed above were interested in this document as of the date of this document. Since organizations and responsibilities can change, you should verify the currency of the information above using the ASSIST Online database at http://assist.daps.dla.mil .

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:59Z
Check the source to verify that this is the current version before use.