INCH-POUND
A-A-58064
21 MAY 96

COMMERCIAL ITEM DESCRIPTION

RESISTOR, DECADE
(0.01-111,111.1 ohm)

The General Services Administration has authorized the use of this commercial item description by all federal agencies.

1. SCOPE.This commercial item description (CID)covers a 7 decade,Decade Resistor and is meant to be the minimum requirements for the item described.

2. SALIENT CHARACTERISTICS.

2.1 Unit shall have 7 dials, with a nominal total resistance of 111,111.1 ohms.

2.1 Resolution shall be 0.01 ohms.

2.2 The accuracy at zero setting shall be at least ± 0.01 % + 2 milliohm.

2.3 The incremental accuracy shall be at least ± (2.0% for the first decade, 0.4% for the second decade, 0.1% for the third decade, 0.4% for the fourth decade, 0.01% for the fifth - seventh decades.

2.4 The temperature coefficients shall be less than ±20 ppm per degree C for 100 ohms and below, less than ±10 ppm per degree C for values above 100 ohms.

2.5 The voltage breakdown shall be a minimum of 2000 volt peak.

2.6 The weight including all accessories shall not exceed 11 pounds.

2.7 The nominal dimensions of the unit shall not exceed (viewed as if rack mounted)4 inches high by 19 inches wide by 5 inches deep.

2.8 The decade resistor unit shall be capable of being rack mounted with optional hardware if specified in the contract.

Beneficial comments, recommendation, additions, deletions, clarifications, etc. and any data which may improve this document should be sent to: SA-ALC/TILDD; Tinker Drive, Bldg 207; Kelly AFB, TX 78241-5916

FSC 6625

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:00Z
Check the source to verify that this is the current version before use.

2.9 The following current and power specifications are specified as the minimum specifications per step:

Decade Number	Power (mW/step)	Current (mW)
1	160	4000
2	250	1580
3	600	770
4	600	240
5	600	77
6	500	22
7	500	7

4. QUALITY ASSURANCE PROVISIONS.

4.1 Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of the CID, and that the product conforms to the producer's own drawings, specifications, standards, and quality assurance practices, and is the same product offered for sale in the commercial marketplace. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

4.2 Bid samples may be required to assure product quality and conformance with specifications as specified in this CID.

5. PACKAGING.

Standard commercial packaging is acceptable unless otherwise specified in the contract.

6. NOTES.

KNOWN ACCEPTABLE SOURCE

The following products are known to have previously met the requirements of this CID and need not resubmit bid samples or certification.

MANUFACTURER	MODEL	NSN
IET LABS	HARS-X-7-.01-RM	6625-01-356-8709
PPM INC	R7-111,111.1K	6625-01-280-2062
GENERAL RESISTANCE	RDS 74M	6625-01-268-4444

POSSIBLE SOURCES

The following manufacturers are potential sources. However, they must meet requirements of this CID, and submit to bid sample testing and certification.

1. GUILDLINE INSTRUMENTS INC. CAGE CODE 3G310
2. OHMITE MANUFACTURING CO. CAGE CODE 44655
3. QUADTECH INC CAGE CODE 0PK96

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:00Z
Check the source to verify that this is the current version before use.

Custodian:
Air Force -99

Review Activity:
Army-MI
DLA-ES

Preparing Activity:
Air Force - 82

Agent:
Air Force - 99

Project No. 6625-0742

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:00Z
Check the source to verify that this is the current version before use.