A-A-30036B
November 27,1991
SUPERSEDING
A-A-30036A
October 13, 1982

COMMERCIAL ITEM DESCRIPTION

PUMP UNIT, ALTERNATING PRESSURE POINT PAD, PHYSICOTHERAPEUTIC

The General Services Administration has authorized the use of this commercial item description.

This commercial item description describes an air filled pad (used on top of the mattress) with a pump supplying air for the pad. The pad provides alternating pressure points intended to increase circulation, aiding in avoiding decubiti.

Salient characteristics:

Pump Unit, Alternating Pressure Point Pad, 110 Volts, 60 Hz., AC, with carrying handle, producing at least 60 mm, Hg. air pressure, pump approved under UL 544, or equal, 7-1/2 feet (229 cm) usable cord +1/2 foot (15 cm) - minus 0 feet (0 cm), with grounding type plug, individually boxed.

Pad Alternating Pressure Point, Physicotherapeutic, plastic, disposable, minimum 28 inches (71 cm) wide to 76 inches (193 cm) long +1 inch [(2.54 cm) - minus 0 inch (0 cm)] inflated, with at least 7 foot (213 cm) + 1 foot (30 cm) of connector tubing and adapter to fit all nationally sold APP pumps.

Unit. The unit of issue shall appear in solicitation.

Workmanship. The pump unit, alternating pressure point pad, physicotherapeutic, shall be free from defects which detract from its appearance or impair its serviceability.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:35Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105D (Normal), with an AQL of 1.0 and 2.5 (percent defective) and an inspection level of S-3.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T08:35Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification: The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements

Federal Food, Drug and Cosmetic Act - If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials - The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking - Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:35Z
Check the source to verify that this is the current version before use.

NOTE. The following National Stock Numbers are covered by this document:

NSN	NOMENCLATURE
6530-00-402-8200	PUMP UNIT, ALTERNATING PRESSURE POINT PAD
6530-00-412-1773	PAD, ALTERNATING PRESSURE POINT, PHYSICOTHERAPEUTIC

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government, may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6530-2291

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:35Z
Check the source to verify that this is the current version before use.