A-A-30043C
February 28, 1992
SUPERSEDING
A-A-30043B
May 13, 1983

COMMERCIAL ITEM DESCRIPTION

URINAL, MALE PATIENT

The General Services Administration has authorized the use of this commercial item description as a replacement for Commercial Item Description A-A-50863, which is cancelled.

This commercial item description describes a disposable and a reusable plastic urinal for the collection of male patient urine.

Salient characteristics. The male patient urinal shall consist of a container, an integral handle, and an attached tight-fitting, odor shield lid. Urinals shall be fabricated from a durable, inert, plastic material which may be a blend of virgin material and reground material. The reground material may not exceed 30 percent of the blend. The urinal shall be capable of standing in an upright position and to rest on its side with the opening angled upwards. Urinals shall be finished smooth on all edges and surfaces and shall be free of flaws and excess flashing. The bottom and side-sitting surfaces shall be formed in such a manner that will not allow the urinal to rock or wobble when resting on a smooth surface. Urinals shall not leak when filled to capacity and in the upright position.

Type I	Disposable, (polyethylene), translucent: beige, tan, mint green or white.
Type II	Reusable and sterilizable, (polypropylene), aqua, blue or turquoise.
Type III	Disposable, (polyethylene), translucent: white.

The handle shall be designed to permit the urinal to be hung on a bedrail (in an upright position) and may be of open finger grip or open contoured design.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:35Z
Check the source to verify that this is the current version before use.

Urinals shall have a capacity of 1000 ml ± 50 ml when placed in the upright position. Graduations are required for the Type I and Type II urinals. Graduations are not required for the Type III urinal; however, urinals having graduations are acceptable. Graduations must be easily read. These urinals are not recommended for use where accurate urinary output measurement is required.

The odor shield lid for the Type I and Type II urinals shall be tight-fitting to minimize accidental spills that may occur if the urinal is placed in other than the upright position. The lid for the Type III urinal shall be tight-fitting enough to prevent leaking when filled to capacity and when resting on its side-sitting surface.

The reusable urinal shall be steam sterilizable using steam under pressure at 250 degrees F. (121 degrees C.).

The urinal shall be free of grease, oil, burrs, cracks, pits, scratches, and foreign material.

Workmanship. The urinals shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Type I. Each (EA).

Type II. Each (EA).

Type III. Package (PG). One package containing 50 urinals, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:35Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Capacity test. The urinal shall be placed on a flat surface in an upright position and filled with 1000 ml of water from a known measured source. A capacity not within + or - 50 ml shall constitute a defect.

Leak test. The Type III urinal shall be filled to capacity and the lid shall be securely fixed in place. The filled urinal will then be slowly lowered to its side resting position. Any leakage of contents shall constitute a defect. This test is not applicable to the Type I and Type II urinals.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for examination purposes shall be one male patient urinal as specified in this description.

Contractor certification: The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:35Z
Check the source to verify that this is the current version before use.

Regulatory requirements

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

Type I - NSN 6530-00-494-8124
Type II - NSN 6530-00-494-8125
Type III - NSN 6530-01-042-2485

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Type of urinal

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:35Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6530-2293

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:35Z
Check the source to verify that this is the current version before use.