A-A-30051C
December 1, 1989
SUPERSEDING
A-A-30051B
July 1, 1986

COMMERCIAL ITEM DESCRIPTION

IRRIGATION KIT, PATIENT

The General Services Administration has authorized the use of this commercial item description as a replacement for VA Specification X-1377f, which is cancelled, and in lieu of Military Specification MIL-I-36969A.

This commercial item description describes a kit suitable for one time use in bladder, gastrointestinal, and wound irrigation.

Salient characteristics. Irrigation Kit, Patient shall be a sterile, disposable, prepackaged assembly, suitable for use in both surgery and nursing procedures. The kit shall be suitable for bladder, gastrointestinal, and wound irrigation and for specimen collection and transport. A waterproof hospital overwrap of sufficient size to permit wrapping of tray, with components inside, for aseptic handling. The hospital overwrap shall provide a sterile field for the manipulation and utilization of the components during surgery and shall provide protection for bed sheets.

The syringe shall be inserted into the solution container and both placed in the tray. The tray shall then be completely wrapped in the hospital overwrap in a manner that will facilitate aseptic handling and provide a sterile field for use in the operating room. The entire assembly shall then be packed and sealed in a plastic peel type bag capable of maintaining sterility until manually opened. The outer package shall be of a design which facilitates aseptic handling. The entire kit shall then be sterilized in a manner acceptable to the Food and Drug Administration.

Type I - Irrigation kit shall consist of the following components.

Plastic BULB type syringe (50/60 ml/cc) with protection tip cover; bulb and syringe securely assembled for an air tight seal. Catheter tip syringe opening shall be a minimum of 3/16 inch ID; must have capability of removing solution with no more than 20 ml remaining in container.
Minimum 1000 ml/cc capacity graduated plastic tray or basin.
Minimum 500 ml/cc capacity graduated plastic solution container.
A waterproof underpad, minimum size 16 by 18 inches

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Type II - Irrigation kit shall consist of the following components:

Plastic PLUNGER type syringe (60 ml/cc), with protection tip cover; catheter tip syringe opening shall be 3/16 inch minimum ID; finger grip on barrel of syringe. Plunger may include control ring.
Minimum 1000 ml/cc capacity graduated plastic tray or basin.
Minimum 500 ml/cc capacity graduated plastic solution container.
A waterproof underpad, 16 by 18 inches, minimum size.

NOTE: For Department of Defense (DoD) procurements, the following additional component is required:

An iodophor prep pad.

Workmanship. The irrigation kit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 20 irrigation kits, each kit individually packaged, sterile in a sealed peel open type container, constitutes one unit. The unit for VA requirements shall appear on the solicitation.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.The contracting officer has the option of accepting or rejecting the product.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification: The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Regulatory requirements

Federal Food, Drug and Cosmetic Act - If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its contractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials - The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation subpart 23.4, to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking - Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For Department of Defense (DoD) procurements.

Quality Assurance Provisions

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample size shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one package of twenty (20) individually sealed units.

NOTE. The following National Stock Numbers are covered by this document:

NSN	TYPE	NOMENCLATURE
6530-00-852-4751	Type I	IRRIGATION KIT, PATIENT
6530-00-865-1950	Type II	IRRIGATION KIT, PATIENT

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government, may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents.

(a) Title, number, and date of this commercial item description.
(b) Type

MILITARY INTERESTS.

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A):

AGENT:

DLA-DH

Custodians:

Army-MD
Navy-MS
Air Force-03

Project Number 6530-0017

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.