INCH-POUND

A-A-30058F
July 15, 1992
SUPERSEDING
A-A-30058E
March 24, 1987

COMMERCIAL ITEM DESCRIPTION

BEDPAN (Plastic)

The General Services Administration has authorized the use of this commercial item description as a replacement for A-A-51354, which is cancelled.

This commercial item description describes a bedpan for nonambulatory patients.

Salient characteristics:

Size 1 - Bedpan, durable strength plastic, oval shaped with side pontoons, permits cleaning in all commercial washers. Without handle, cover, frayed or sharp edges, front or rear pontoons. Bedpan shall not crack or deform when steam washed or when supporting patients up to 250 pounds (113.40 kg). Nominal volume shall not be less than 2000 ml. Shall be of the design which provides relative comfort, stability and balance during use. All measurements shall be taken at the widest points.

Type I - Disposable (one patient use); beige, tan, yellow or white. Length 14-1/2 inches plus or minus 1/4 inch (37 cm plus or minus .6 cm), width 11 inches plus or minus 1/2 inch (28 cm plus or minus 1.27 cm), front depth 3-1/2 inches to 4-9/16 inches (9 cm to 11.6 cm), rear depth 2-1/2 inches plus or minus 1/4 inch (6 cm plus or minus .6 cm), side depth 2-1/2 inches plus or minus 1/4 inch (6 cm plus or minus .6 cm).

Type II - Reusable and autoclavable; aqua, blue or turquoise. Length 14-1/2 inches plus or minus 1/4 inch (37 cm plus or minus .6 cm), width 11 inches plus or minus 1/2 inch (28 cm plus or minus 1.27 cm), front depth 3-1/2 inches plus or minus 1/4 inch (9 cm plus or minus .6 cm), rear depth 2-1/2 inches plus or minus 1/4 inch, (6 cm plus or minus .6 cm), side depth 2-1/2 inches plus or minus 1/4 inch (6 cm plus or minus .6 cm).

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Size 2 - Bedpan, durable strength plastic, oval shaped with side pontoons, permits cleaning in all commercial washers. Without handle, cover, frayed or sharp edges, front or rear pontoons. Length 14-1/2 inches plus or minus 1/2 inch (37 cm plus or minus 1.27 cm), width 11 inches plus or minus 1/2 inch (28 cm plus or minus 1.27 cm), front depth 2-3/4 inches to 4-1/2 inches (7 cm to 11 cm), rear depth 2-1/2 inches plus 1/4 inch, minus zero inch (6 cm plus .6 cm, minus zero cm), side depth 2-1/2 inches plus or minus 1/4 inch (6 cm plus or minus .6 cm). Bedpan shall not crack or deform when steam washed or when supporting patients up to 250 pounds (113.40 kg). The bedpan shall be capable of holding 2000 ml when inclined at any angle in any direction up to 10 degrees without spillage. Shall be of a design which provides relative comfort, stability and balance during use.

Workmanship. The bedpans shall be free from defects which detract from their appearance or impair their serviceability. The bedpans shall be smooth (no sharp edges) and material must be of uniform thickness (no thin areas).

Unit. Package (PG). One package containing twenty bedpans constitutes one unit. The UNIT for the VA items will appear in the solicitation.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

NSN	SIZE	TYPE	NOMENCLATURE
6530-00-014-1018	1	I	BEDPAN (Plastic)
6530-00-494-8121	1	II	BEDPAN (Plastic)
6530-01-049-0428	2		BEDPAN (Plastic)

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Type and size

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:36Z
Check the source to verify that this is the currentversion before use.

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6530-2355

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.