INCH-POUND

A-A-30092C
September 18, 1992

SUPERSEDING
A-A-30092B
December 16, 1985

COMMERCIAL ITEM DESCRIPTION

CRUTCH (Cane Type, Forearm and Elbow)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a cane type aluminum crutch which is used to support disabled persons in walking. The crutch has an adjustable arm support and forearm cuff.

Salient characteristics. CRUTCH, Cane Type, Forearm and Elbow, made of 7/8 inch or 1 inch (2 cm or 2.54 cm) anodized aluminum tubing and/or triple chrome plated nickel steel tubing; front or side opening cuff. Cuff to be stainless steel or plastic/vinyl coated spring steel. Cuff designed to withstand normal use without cracking or breaking. Cuff assembly to be held in place with lock nut and bolt or rivets (if rivets are supplied, must be chrome plated, solid or semi-tubular steel), designed to avoid loosening, twisting or wear. Crutch shall have adjustment capabilities from floor to handle and from handle to cuff. Minimum adjustment range from floor to handgrip 30 to 36 inches. Crutch shall be furnished with precise hole alignments to allow adjusting and locking at varying heights. Crutch shall be furnished with a nonmarring, slip resistant, rubber tip having a concave bottom surface, a metal disc on inside surface of the base section, internally molded reinforcement rings on inside wall and vertically ribbed outside wall to maintain rigidity. Handgrips shall be nonmovable, nonslip, rubber/vinyl and a color closely matching tip and cuff. Manufacturer's name or trademark must be affixed to the product so as to remain intact and identifiable under normal usage. Each crutch must be capable of supporting 300 lbs. weight. Crutches not to weigh more than 5 lbs. per pair.

Workmanship. Crutch shall be free of burrs and sharp edges or other defects which detract from its appearance or impair its serviceability.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Unit. Pair. Twelve pair per shipping container.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-4.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

NOTE. The following National Stock Number is covered by this document:

NSN 6515-00-810-9092

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VAMKC (904)
Agent VA-OSS VACO (92A)

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6515-5310

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.