COMMERCIAL ITEM DESCRIPTION

PADS BED LINEN PROTECTIVE

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description describes an absorbent, disposable bed linen protective pad which provides protection from excretions encountered in patient care.

Salient Characteristics Pad, Bed Linen Protective, disposable

Type I 23 x 24 inches ± 1/2 inch (or approx. 57.50 cm x 60 cm ± 1.25 cm), flame retardant (tested in accordance with DOC Standard FF-4-72 - Mattress and Cigarette Test) nonwoven top layer, absorbent filler, embossed waterproof blue backing, pad capable of absorbing 200 ml. of water when tested with verticle test, sealed on 4 edges, 200's

Type II Size 1 - 18 by 24 inches (45.7 by 61.0 cm), 300's
Size 2 - 23 by 24 inches (58.4 by 61.0 cm), 200's
Shall consist of a pad facing over absorbent layers, backed by a polyethylene film
Material Facing material shall be highly permeable, high wet strength nonwoven fabric Facing material shall have a wet bursting strength of not less than 5 psi Facing material shall be fire retardant in accordance with Flammable Fabrics Acts CS191-53 and shall have a minimum weight of 12 gms/sq yd. Absorbent layers shall consist of two layers of absorbent cellulose wadding with a layer of absorbent cellulose fluff between, or other suitable construction of comparable performance such as multiple layers of cellulose wadding Absorbent layers shall be a minimum 16 by 21 inches (40.6 by 53.3 cm) with a minimum weight of 29 grams for Size 1 pads, and shall be a minimum of 21 by 22 inches (53.3 by 55.9 cm) with a minimum weight of 37 grams for Size 2 pads Backing material shall be suitable polyethylene film 0.75 mil + 5%. The backing shall be suitably sealed to all sides of the facing, providing a continuous seal around the periphery of the pad, and shall also be sealed to the bottom absorbent layer Sealed sides shall be soft and pliable, so as not to irritate sensitive areas of a patient's body
Design and construction Each pad shall consist of layers of absorbent material between a polyethylene backing and a pad facing The dimensions of the finished pad shall be a minimum of 17 by 23 inches (43.1 by 58.4 cm) for Size 1 pads and shall be a minimum of 22 by 23 inches (55.9 by 58.4 cm) for Size 2 pads The finished pad shall weigh a minimum of 35 grams for Size 1 pads, and a minimum of 47 grams for Size 2 pads.
Absorbency The finished pad shall have an absorbent capacity of not less than 350 grams for Size 1 and 500 grams for Size 2 when tested as follows. Ten specimens of the finished pad, each 6 inches square, shall be cut from an area of least absorbent density, (but excluding the peripheral sealed area), and as free of adhesive as possible Each specimen shall be placed on a supporting screen and weighed to the nearest 0.1 gram Both the specimen and the screen shall be submerged in water for 10 seconds After removal from the water, the specimen and screen shall be allowed to drain at an angle of 30 degrees from the horizontal for one minute The specimen and the screen shall then be weighted to determine the weight of water absorbed by the specimen The average absorption of the ten specimens shall be calculated The total absorbency of the finished pad shall be determined by dividing the average specimen absorbency by 36 square inches and multiplying the total surface area, (square inches), of the absorbent portion of the finished pad.

Workmanship The finished pads shall be clean and free from any defects which do not affect their appearance or materially reduce their usefulness Pads shall [illegible] and be absorbent layer [illegible]

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Contractor Certification The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards and quality assurance practices, and is the same as the product sold in the commercial marketplace The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory Requirements

Federal Food, Drug and Cosmetic Act - If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations

Recovered materials - The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking - Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

MILITARY INTERESTS

Military Coordinating Activity

DoD-MB

Agent

DLA-DM

Custodians

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY

VA-OSS VAMKC (904A)

AGENT

VA-OSS (91X)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.