A-A-53007
26 June 1987

COMMERCIAL ITEM DESCRIPTION

ANALYZER, DEFIBRILLATOR

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a defibrillator analyzer to be used in testing of defibrillators and accompanying monitors when present.

Salient characteristics:

The unit shall be a portable unitized defibrillator/monitor analyzer.

All displays shall be digital.

The unit shall measure defibrillator output energy with a range maximum of at least 700 Joules, ±2 percent.

The defibrillator analyzer shall be compatible with Lown, Edmark and Trapezoid waveforms of up to 4kV amplitude.

The unit shall have a defibrillator pulse waveform memory.

The defibrillator analyzer shall have a real-time expanded time scale oscilloscope output, with calibrated 1000:1 attenuation.

The defibrillator analyzer shall have an expanded time scale ECG recorder output with a time expansion factor of 200:1.

The defibrillator analyzer shall measure and display actual cardioversion discharge delay time in milliseconds.

The unit shall measure defibrillator output energy via 50 ohm, ±1 percent test load.

The unit shall have a defibrillator output energy test pulse.

The unit shall generate ECG performance waveforms to test monitor gain and damping adjustment, low frequency response, bandpass, band reject, high frequency response, linearity, and ECG rate meter/alarms.

AMSC N/A FSC 6625

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

The defibrillator analyzer shall operate on 120 Vac, 50/60 Hz. power, and shall weigh less than 15 pounds.

Workmanship. The defibrillator analyzer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One defibrillator analyzer, as specified, constitutes one unit.

Metric products.

Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right perform any of the inspections set forth in the specification when such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

Sampling for testing and examination. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one defibrillator analyzer.

Table I. Sampling for Testing and Examination.

	Inspection Level	AQL (Defects Per Hundred Units)
Major Defects	III	0.4
Minor Defects	III	0.4

Tests. All tests shall be performed on a first article sample in accordance with table II.

Table II. Tests.

1) Safety

Electrical leakage current shall comply with those limits specified in ANSI/AAMI SCL 12/78. Tests shall be in accordance with above standard. Equipment shall be a leakage current tester/safety analyzer with capability to perform test described in above standard.

Unit shall incorporate a three-conductor, hospital grade plug on the power cord.

2) Performance

The manufacturer shall demonstrate to the Government via mutally acceptable test methods, compliance with the following specifications and performance criteria:

Weight: < 15 lbs.
Energy level meter accuracy: ±2 percent
Output load resistance: 50 ohms, ±1 percent
Cardioversion delay time display: ±2mS
Gain and damping adjustment waveform: Bipolar squarewave, 2 Hz, 1mVpp
Low frequency response: DC pulse, 4 seconds, ±1mV peak
Bandpass, band reject, AHA high frequency 3 dB point: Bipolar sine wave, 10, 40, 60, 100, 200, 1000 Hz burst frequencies, 1 second on - 1/4 second off duty cycle, 3mVpp.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:14Z
Check the source to verify that this is the current version before use.

Table II. Tests (Continued).

Linearity: Bipolar triangle, 2 Hz, 4mVpp ECG rate meter/alarm test: 60-30-60-120-240-60 Hz automatic cycle sequence, 20 second cycle duration each.

Examination. Each defibrillator analyzer shall be examined to determine compliance with the classification of defects in table III.

Table III. Classification of Defects.

Categories	Defects
Major	
101	Power and control switch(s) non-functional
102	Digital indicator non-functional
103	Inputs and outputs non-functional
Minor	
201	Identification markings not complete, permanent, and legible.
202	Finish not free of chips, cracks, scratches and other appearance defects.
203	Any component part not secured in place.

Examination shall not be limited to the classified possible defects listed herein.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:14Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6625-01-204-8164

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

A-A-53007

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6625-0674

Location: CID/ANALYZER/H17-5

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements All users of military standardization documents are invited to provide suggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn DPSC-RSTC
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53007	Analyzer, Defibrillator

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last First MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil – Downloaded: 2026-04-26T09:14Z
Check the source to verify that this is the current version before use.