A-A-0053079(DM)
17 July 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

INFUSION SET, SCALP VEIN, (21 Gage by 3/4 Inch Needle, with 3-1/2 Inch Tubing, Disposable, 40s)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a single type of pediatric scalp vein infusion set used in administering medication.

Salient characteristics:

Materials. The cannulae shall be made of Type 304 stainless steel in accordance with ASTM F899. The tubing shall be made of plasticized polyvinyl chloride and conform to Class IV Plastic requirements of the USP. The tubing shall be clear or clear with a frosted surface. The wings shall be plasticized polyvinylchloride. The hub shall be made of polyvinylchloride, polypropylene or nylon. The protective cap on the hub shall be made of polyethylene or polypropylene. The needle shall be protected with a polyethylene sleeve or cap.

Construction. The needle shall be a 21 gage thin wall needle 3/4 inch long with a short bevel. The bore of the needle shall be 0.0250 to 0.0300 inch. The tubing length shall be 3-1/2 inches and have an ID of 0.040 inch. The tubing volume shall be 0.17 ± 0.03 ml. The set shall have two wings permanently attached to the proximal end of the needle. The needle gage number shall be molded on a wing. The wings shall have an irregular surface to facilitate gripping. The needle bevel shall be up relative to the held wings. The wing spread shall be 1-1/8 inches by 3/8 inch by 0.040 inch thick. Tolerances unless otherwise stated are ± 10%. The hub shall have a Luer taper in accordance with ANSI/HIMA MD70.1. The hub shall be provided with a fully secured natural rubber injection plug. The needle shall be protected with a length of tubing extending over and beyond the length of the needle.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

The set with the needle stoppered by putting the tip into a piece of rubber shall withstand 20 psig air pressure for one minute without leakage of air. This test shall be conducted with the hub from which the injection plug has been removed attached to a Luer fitting having a lumen for the air entrance. The entire set shall be submerged in water during the test.

The tubing shall have a minimum ultimate tensile strength of 2800 psi and ultimate elongation of 250%.

The tubing shall have a 100% modulus of 1000 to 1900 psi determined by the tensile reading at the moment the tubing has been stretched to twice its original length.

A. Sterility test. The infusion set shall be tested for sterility in accordance with the method specified in the USP using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it is clearly differentiated from product intended for distribution. The incubation medium for the paper spore strip (carrier) shall be soybean-casein digest. The incubation medium for the inoculated product, or inoculated simulated product shall be fluid thioglycollate or alternative thioglycollate at 30° to 35° C for one half of the specimens and soybean-casein digest at 20° to 25° C for the remaining half of the specimens. The minimum incubation time shall be ten days. Sampling shall be as follows:

Medium	Paper spore strip	Inoc. Prod. or Inoc. Simulated Prod.
Soybean-casein digest	S-3*	S-3*
Fluid thioglycollate or Alternative thioglycollate		S-3*

*Use a minimum of 20 specimens. Acceptance number is zero.

B. Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Type of Sterilization	Organism
Steam	B. stearothermophilus
Dry Heat	B. stearothermophilus
	B. subtilis
Ethylene oxide	B. subtilis
Gamma radiation	B. pumilus

The number of viable spores shall be sufficient to provide assurance that the product has been sterilized.

C. Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586A, Sterilization Test Strip Set, Bacterial Spore. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in A. Sterility test and B. Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in B. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP and paragraphs A and B.

D. Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

E. Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP. The inspection level shall be as stated in paragraph A above.

Workmanship. The set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing forty individually packaged sterile infusion sets, as specified, constitutes one unit. The individual package shall be a sealed, peel open type container capable of maintaining sterility of contents unless package is damaged or opened.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one scalp vein infusion set.

Sampling criteria for tests.

Characteristic	Inspection level	AQL
Pressure test	S-2	1.0
Tensile strength & elongation	S-2	1.0
Elastic modulus	S-1	1.5
Sterility	(See sterility clauses)	
Class IV plastic	S-1	0

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-250-1141	INFUSION SET, SCALP VEIN Pediatric, 21 Gage by 3/4 Inch Needle, with 3-1/2 Inch Plastic Tubing, Disposable, 40s

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515- 4155

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s), or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D.C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATIS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions — Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* — Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* — Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* — Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.