20 July 1987

COMMERCIAL ITEM DESCRIPTION

OVEN, LABORATORY

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a bench mounted gravity convection oven with a +40 to +200 degrees centigrade (C) operating temperature range.

Salient characteristics.

Shall be a single chamber type gravity convection oven for preheating, drying, sterilizing, baking, and aging procedures.

Shall have an operating temperature range that includes +40 to +200 degrees (C).

Shall have nominal chamber dimensions of 16 1/2W X 11 1/2D X 16H (inches). Shall have nominal overall dimensions of 18W X 14D X 23H (inches).

Shall be suitable for bench mounting.

Shall be suitable for operation from 120 volts, 50/60 Hz, AC supply lines. Shall be rated for a power consumption of 1600 watts or VA.

Shall consist of a rectangular steel cabinet with a base. The cabinet shall enclose a thermostat controlled electric heating unit that shall be located below the chamber. The base shall include suitably mounted controls that include.

a. An on-off switch for the electric power.
b. A temperature setting control dial with reference graduations.
c. A red color neon lamp to indicate when on-off switch is in the "on" position.
d. An amber color neon lamp to indicate when the thermostat is activated.

AMSC/NA FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

The cabinet shall have suitable openings for a fresh air intake and an adjustable exhaust part. The top of the cabinet shall have a suitable opening for the insertion of a glass thermometer.

Shall include a hinged steel door with a handle. The door shall be designed and constructed to make a tight and secure fit with the cabinet housing during closure.

Shall have door and cabinet walls designed and constructed or reinforced so that no "oil canning" effect is present. The inner walls of the chamber shall be insulated with non-asbestos and non-combustible material. The insulating material shall be installed in either of the following methods.

a. Installed between a double wall chamber housing that may be supplied.
b. Installed in the form of sheeting material that is secured to the interior walls of the chamber housing.

The insulating material that may be supplied within a double wall type construction shall be of sufficient density to prevent settling and powdering. The construction of the chamber walls (for double wall construction) shall prevent infiltration of any moisture from the chamber to the insulation, or the infiltration of insulation into the chamber. The insulation sheeting, that may be supplied for single wall housing construction, shall be durable without evidence of or tendency to rot, rust, crack or chip.

Shall be supplied with at least two (2) removable metal shelves that are suitably protected against corrosion. The height of a shelf within the chamber shall be adjustable in at least five (5) positions. The shelves shall be supported by brackets or runners that are securely attached to the two (2) opposite walls in the chamber.

Shall be supplied with a thermometer for measuring the temperature within the chamber. The thermometer shall be top mounted and partially inserted into the chamber. The thermometer shall be capable of being secured to the oven housing to a depth where it can be easily read without having to be removed, or having to open the cabinet door. The thermometer shall have a range that includes 0 to +250 degrees (C) with at least 2 degree (C) increments.

Shall have all metal surfaces and hardware suitably protected against corrosion.

Shall be supplied with a terminal box for use on 120 volt supply lines.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Shall be supplied with service data that contains complete instructions for installation, operation, maintenance, and a list of component parts. Wiring diagrams shall be included for electrical items or components. Each part of the oven, other than common fasteners, shall be identified in the service data and parts list by the manufacturer's part number.

Workmanship. The laboratory oven shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One laboratory oven, as specified, constitutes one unit.

Ouality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with commercial practice.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the currenversion before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6640-01-052-8326

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - EA
Navy - MS
Air Force - 99

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PREPARING ACTIVITY:

DLA-DM

Project No. 6640-1432

Location: NCR DISK # 3

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:15Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, allowed too much latitude, was loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4986 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-LS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER | 2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION

b. ADDRESS *(Street, City, State, ZIP Code)*

4. TYPE OF ORGANIZATION *(Mark one)*

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional

b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional

c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional

8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.