A-A-0053073DM)
15 July 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

GLOVES, DENTAL
(Rubber, Non-Sterile, Disposable, 50 PR)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers non-sterile rubber latex gloves suitable for non-surgical dental procedures. Their purpose in use is to prevent infection of dental personnel.

Salient characteristics:

The dental gloves shall be of the following sizes:

National Stock No.	Size
6515-01-255-6315	5-1/2
6515-01-255-6316	6-1/2
6515-01-255-6317	7
6515-01-255-6318	9
6515-01-255-6319	6
6515-01-255-6320	7-1/2
6515-01-256-4991	8
6515-01-257-4665	8-1/2

The gloves shall be in accordance with ANSI/ASTM D3577 except as follows:

5.1 Materials. Delete entire paragraph and substitute: Natural rubber latex shall be used to fabricate the gloves. Compounds and additives present in or on the rubber latex shall be biologically safe. Absorbable Dusting Powder, USP shall be used to facilitate donning. The gloves are not required to be sterile. They must, however, be clean and have been made under conditions of good housekeeping as required for medical devices.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the current version before use.

8.3 Airtight test. Gloves shall be inflated with air to a pressure of 20 mm ± 2 mm Hg while immersed in water at 70° to 75°F (21° to 24°C) to observe pinholes. The gloves shall be manipulated so that the hand of the tester does not remain on one spot of the glove and thus obscure possible leakage.

8.3.1 Alternate airtight test. Gloves shall be inflated with air to a pressure of 12 to 14 mm Hg. The glove is then sealed at the cuff. Examination shall be made after 15 minutes to determine the presence of a leak as shown by deflation.

8.4.1 Thickness of the glove shall be as follows:

Finger tip	.13 to .23 mm
Palm	.13 to .25 mm
Wrist	.13 to .30 mm

Workmanship. The gloves shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 50 pairs of gloves suitably labeled and designed for dispensing the right hand or left hand glove, or a joined pair, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the currentversion before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one pair of gloves.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-255-6315	GLOVES, DENTAL, Rubber, Size 5-1/2, Non-Sterile, Disposable, 50 PR
6515-01-255-6316	GLOVES, DENTAL, Rubber, Size 6-1/2, Non-Sterile, Disposable, 50 PR
6515-01-255-6317	GLOVES, DENTAL, Rubber, Size 7, Non-Sterile, Disposable, 50 PR
6515-01-255-6318	GLOVES, DENTAL, Rubber, Size 9, Non-Sterile, Disposable, 50 PR
6515-01-255-6319	GLOVES, DENTAL, Rubber, Size 6, Non-Sterile, Disposable, 50 PR
6515-01-255-6320	GLOVES, DENTAL, Rubber, Size 7-1/2, Non-Sterile, Disposable, 50 PR
6515-01-256-4991	GLOVES, DENTAL, Rubber, Size 8, Non-Sterile, Disposable, 50 PR
6515-01-257-4665	GLOVES, DENTAL, Rubber, Size 8-1/2, Non-Sterile, Disposable, 50 PR

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC SF045

Project No. 6515-4151

☆U.S. GOVERNMENT PRINTING OFFICE: 1987 - 505-033-75503

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the currentversion before use.