A-A-53073A
11 July 1989
SUPERSEDING
A-A-0053073(DM)
15 July 1987

COMMERCIAL ITEM DESCRIPTION

GLOVES, DENTAL
(Rubber, Non-Sterile, Disposable, 50 Pr)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers non-sterile rubber latex gloves suitable for non-surgical dental procedures. Their purpose in use is to prevent infection of dental personnel.

Salient characteristics:

The dental gloves shall be of the following sizes:

National Stock No.	Size
6520-01-255-6315	5-1/2
6520-01-255-6316	6-1/2
6520-01-255-6317	7
6520-01-255-6318	9
6520-01-255-6319	6
6520-01-255-6320	7-1/2
6520-01-256-4991	8
6520-01-257-4665	8-1/2

The gloves shall be in accordance with ANSI/ASTM Designation D3578 for Rubber Examination Gloves except as follows:

4.1 Materials. Delete entire paragraph and substitute: "Natural rubber latex shall be used to fabricate the gloves. Absorbable Dusting Powder, USP, shall be used to facilitate donning. The gloves are not required to be sterile. They must, however, be clean and have been made under conditions of good housekeeping as required for medical devices."

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

7.3 Watertight test. Delete "filling with 300 ± 25 ml of water" and replace with "filling with 400 ± 25 ml of water".

7.4.1 Thickness of the glove shall be as follows:

Finger tip	.13 to .23 mm
Palm	.13 to .25 mm
Wrist	.13 to .30 mm

9.2.1 Add to paragraph: "Each pair of gloves shall consist of one contoured right hand glove and one contoured left hand glove".

Workmanship. The gloves shall be free from defects which detract from their appearance or impair their serviceability. The gloves shall be constructed in one piece, free of seams and welts, and shall have a smooth, bisque or textured finish. The gloves shall be prepowdered with a minimum of dusting powder on the interior only. Gloves shall be free of foreign matter, bloom, bleedings, strings or other imperfections.

Unit. Package (PG). One package containing 50 pairs of gloves suitably labeled to include size and designed for dispensing the right hand or left hand glove, or a joined pair, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one glove.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number are covered by this document:

NSN	ITEM IDENTIFICATION
6520-01-255-6315	GLOVES, DENTAL, Rubber, Size 5-1/2, Non-Sterile, Disposable, 50 Pr.
6520-01-255-6316	GLOVES, DENTAL, Rubber, Size 6-1/2, Non-Sterile, Disposable, 50 Pr.
6520-01-255-6317	GLOVES, DENTAL, Rubber, Size 7, Non-Sterile, Disposable, 50 Pr.
6520-01-255-6318	GLOVES, DENTAL, Rubber, Size 9, Non-Sterile, Disposable, 50 Pr.
6520-01-255-6319	GLOVES, DENTAL, Rubber, Size 6, Non-Sterile, Disposable, 50 Pr.
6520-01-255-6320	GLOVES, DENTAL, Rubber, Size 7-1/2, Non-Sterile, Disposable, 50 Pr.
6520-01-256-4991	GLOVES, DENTAL, Rubber, Size 8, Non-Sterile, Disposable, 50 Pr.
6520-01-257-4665	GLOVES, DENTAL, Rubber, Size 8-1/2, Non-Sterile, Disposable, 50 Pr.

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC SF045

Project No. 6520-2189

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE GLOVES, DENTAL (Rubber, Non-Sterile, Disposable, 50 Pr)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

ISTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in bmitting comments and suggestions for improvements. All users of military standardization documents are invited to provide ggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and ailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was o rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the roblems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an knowledgement will be mailed to you within 30 days to let you know that your comments were received and are being nsidered

OTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of ecification requirements on current contracts Comments submitted on this form do not constitute or imply authorization waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

ICIAL BUSINESS
ALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RST
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.