A-A-53078
16 July 1987

COMMERCIAL ITEM DESCRIPTION

TUBE AND SYRINGE, STOMACH
(Adult, Lavage, Edlich, Disposable, 34 Fr)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers an adult tube and syringe used for flushing the stomach.

Salient characteristics.

General. Shall be a disposable, adult stomach tube and syringe consisting of a 34 Fr tube and a 5-ounce catheter tip syringe.

Material. The syringe shall be fabricated from polypropylene. The tube shall be fabricated from transparent polyvinyl chloride. The syringe adapter and slide clamp shall be fabricated from rigid plastic. The syringe plunger tip shall be fabricated from rubber. The plastic components coming in contact with the fluid shall meet the requirements of the Pharmacopeia of the United States for Class IV plastics (USP).

Those rubber and plastic components coming in contact with the fluid shall contain not more than 5 parts per million (ppm) total lead, barium and cadmium when tested in accordance with the Physico-Chemical Tests in the USP.

Characteristics. Shall consist of a 5-ounce syringe, a 34 Fr ± 2 Fr tomach tube with a length of 36 inches +4,-1 inch and a slide clamp. The proximal end of the tubing shall have an adapter designed to attach to the syringe in a leak-tight fit.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

The syringe shall have subgraduations every 5 milliliters (ml) up to the 140 ml mark. Each 10 ml shall be numbered and longer than the subgraduations. There shall also be a numbered graduation at each ounce with subgraduations at each 1/4 ounce. The ounce graduations shall be longer than the subgraduations.

The plunger tip shall be lubricated with pharmaceutical grade silicone lubricant. The plunger shall make a good seal between the barrel and the plunger such that there shall be no leakage past the plunger in either direction.

The plunger tip shall have a minimum of two rings making contact with the barrel. The distal tip ring shall accurately indicate the syringe fluid volume on the barrel. The plunger tip shall be of a color which clearly contrasts with the syringe barrel for maximum visibility.

The syringe shall have an integral catheter tip and an outside flange designed to provide a finger grip and also function as an anti-roll device. The catheter tip shall be one inch long and taper from a nominal 0.450 to 0.500 inch.

The tubing shall have 4 non-opposing eyes within 6 inches of the distal tip. The eyes shall measure nominally 1/4 inch on the short axis and 3/8 inch on the long axis. The tubing shall be a nominal 5/16 inch on the inside diameter. The tubing shall have insertion depth marks at 18, 22, 26 and 30 inches from the distal tip. Each eye shall have the same cross sectional area as the tubing.

Freedom from foreign matter. The fluid path shall be free from chips, dirt or other foreign matter.

Workmanship. The tube and syringe assemblies shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing twenty individually packaged tube and syringe assemblies, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for visual examination shall be in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one stomach tube and syringe. Sampling shall be in accordance with table I listed below.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one stomach tube and syringe. AQL shall be 1.0 percent defective. Inspection level shall be S-2.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

TABLE I Sampling for Examination

Type of Defect	Inspection Level	AQL (% defective)
Major	S-3	1.0

Classification of Defects

Categories	Defects
Major	
101	Sharp edges or burrs at eyes or closed end of tube.
102	Poor fit of syringe catheter tip to adapter or tube resulting in separation or leakage.
103	Lack of seal at plunger tip-barrel interface resulting in leakage past plunger tip.
104	Graduation marks on syringe not clear, or removable by rubbing.
105	Burrs at catheter tip of syringe.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-435-8959	TUBE AND SYRINGE, STOMACH, Adult, 34 Fr, Lavage, Edlich, Disposable, 20s

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES.

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3942

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53078	TUBE AND SYRINGE, STOMACH (Adult, Lavage, Edlich, Disposable, 34 FR)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.