INTERIM COMMERCIAL ITEM DESCRIPTION

HANGER, X-RAY FILM

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This Interim Commercial Item Description covers a hanger used for processing medical x-ray films in developing and drying procedures.

Salient characteristics:

The x-ray film hanger shall be designed for developing and drying sheet medical x-ray films. The hanger shall be supplied in one of the following nominal frame sizes:

Size 1 - 9-1/2 inches by 9-1/2 inches

Size 2 - 24 cm by 30 cm

All parts of the hanger, including clips, clip points, top bar pins, springs, frames, and any other parts, shall be constructed of one piece corrosion-resisting steel, conforming to American Iron and Steel Institute (AISI) series 300.

Each hanger shall consist of a top bar, a rectangular frame, spring clips, and, if required, a crossbar designed and constructed to hold film in a taut, vertical position for processing.

The top bar shall be 16 $\pm$ 1/16 inch long of 1/4 inch channel or solid bar stock. If channel, it shall be formed of stainless steel 0.031 inch thick with commercial tolerances. Each leg of the channel shall form an angle of 90° + 5° -0° with the base. If solid bar stock, the top bar shall be 0.250 $\pm$0.015 inch square. The sides of the top bar shall be parallel with the plane of the wire frame, with a maximum offset of 0.005 inch.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

There shall be a pin 5/8 ± 1/8 inch long attached to the under surface of the top bar 3/4 ± 1/8 inch from one end and 14-3/8 ± 1/8 inches from the opposite side of the frame, outside-to-outside dimensions. The other end of the top bar shall extend 3/4 ± 1/8 inch beyond the frame.

The frame shall be constructed from 3/16 inch diameter rod (commercial tolerances permitted), securely riveted or welded to the top bar, and shall be formed to hold film with its long axis vertical. To assure complete tank submersion of film secured in the hanger, the hanger design shall be such that the upper clips fall at least 3 inches below the top bar with a film mounted into position. Dimensional requirements shall be as indicated below:

Size	Overall vertical height	Inside width of frame
1	13-1/2 inches (minimum)	9-1/2 inches (minimum)
2	40 cm (minimium)	24 cm (minimum)

Four spring clips shall be riveted, clinched, or welded to the hanger frame to securely grip film with square or round (3/8 inch radius, typ.) corners, not less than 3/16 inch from the nearest film edge. The two bottom clips shall be attached to the frame. The two top clips shall be attached to springs of sufficient tension to hold the film uniformly taut without warping the film, or without pulling the film out of the clips in processing.

The design of the clips shall be such that trapping of processng solutions is minimized. The overall thickness of the clip shall be no greater than 1/2 inch. All joints shall be securely riveted, spot-welded, or clinched. All welds shall be ground smooth with an iron-free grinding wheel.

The hanger shall be passivated to restore corrosion-resisting properties after fabrication. As an alternate to passivation, the hangers may be electropolished.

The hanger shall be designed to withstand crazing, cracking, distortion or other deterioration during normal use for the entire life of the hanger.

Any welding shall be durable, sound in every detail, free of cracks, porosity, and other defects, and in accordance with the latest practices recommended by the American Welding Society.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the currenversion before use.

The hanger shall be free from burrs or sharp edges. It shall be free of oil, grease, dirt, or other foreign matter.

Workmanship. The x-ray film hanger shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One hanger, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one x-ray film hanger as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and minor combined)	II	4.0
For dimensional examination	S-1	2.5

Examination. The x-ray film hanger shall be examined to determine compliance with all requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects below.

TABLE II. Classification of Defects.

Classification	Defect
Major	
101	Any component part missing, broken, or otherwise unusable.
102	Hanger not fabricated from one piece corrosion-resisting steel, AISI 300 series.
103	Upper clips do not fall at least 3 inches below the top bar with a film mounted into position.
104	Hanger not passivated or electropolished.
105	Exposed edges of hanger not free of burrs or sharpness.
106	Riveted or clinched joints not securely and properly fastened.
107	Finish not free of pits, cracks, scratches, or corroded areas.
Minor	
201	Hanger not free of grease, oil, dirt, or other foreign matter.

Dimensional Examination. The x-ray film hanger shall be examined for defects in dimensions. Any dimension not within the specifications herein shall be classified as a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6525-01-113-0138 (Size 1)
6525-01-112-4053 (Size 2)

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the currenversion before use.

A-A-0053082(DM)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6525-1229

PGC 70042
70043

Location: CID/HANGER/H13-29

☆U.S. GOVERNMENT PRINTING OFFICE: 1987 - 505-033-75625

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the current version before use.