A-A-53085A
17 June 1988
SUPERSEDING
A-A-0053085(DM)
22 July 1987

COMMERCIAL ITEM DESCRIPTION

MOULAGE KIT, COLOSTOMY

The General Services Administraton has authorized the use of this Commerical Item Description.

This Commercial Item Description covers a full size colostomy moulage kit suitable for use in providing demonstrations or practice in colostomy irrigation.

Salient characteristics:

The kit shall closely simulate the front abdominal region and pertinent adjacent areas of the human anatomy. The kit shall consist of three colostomy moulages simulating different stages of colostomy development. The kit shall include the following components:

a. A complete moulage with a colostomy rosebud (single-barreled) without infection around the base of the rosebud.

b. A complete moulage with a colostomy rosebud (single-barreled) having infection around the base of the rosebud.

c. A complete moulage with a colostomy loop (double-barreled) having necrotic skin edges, and a sutured laparotomy incision.

d. Carrying case

e. Instruction sheet

AMSC N/A FSC 6910

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the current version before use.

Colostomy Moulages

Each moulage shall be adaptable for attachment to a nursing technique doll or to a human subject as a recipient of treatment. The moulage shall be a full size model with flesh shown in relief on the surface of the moulage and shall cover sufficient area, when attached, to present a realistic picture of the front abdominal region. Each complete colostomy moulage shall include the following components:

a. Rigid plastic base
b. Removable skin area
c. Foam plastic filler
d. Fastening straps (3 pair)
e. Plastic water bag
f. Drain tube, 18 inches long (minimum) with a connector approximately 1/2 inch long, 3/16 inch OD, and 3/32 inch ID at each end.
g. Rosebud simulating a stage in the development of colostomy.

Drop withstand. Each complete colostomy moulage, without case, shall be capable of being dropped from a height of 48 inches (121.92 cm) onto a concrete floor without evidence of damage to structure or finish.

Straps. Each moulage shall be provided with three strap sets. Side one shall be 8 inches (20.3 cm) by 3/4 inch (1.9 cm). Side two shall be 16 inches (40.64 cm) by 3/4 inch (1.9 cm). The straps shall be made of flexible plastics. Side one shall have a permanently affixed hook and side two shall have a metal slide buckle. The buckle shall be permanently affixed and adjustable for allowance of variations in subject size. The straps shall be distanced equally on vertical to effect a good fit. One end of each strap shall remain a permanent fixed part of the moulage.

Bag. The plastic water bag with sufficient flexible tubing shall be designed for fastening on the inner surface of the rigid plastic base on the right side of the navel. The bag shall be so placed that fluid may be introduced with the doll or human subject in either a sitting or supine position with no backflow resulting.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:15Z
Check the source to verify that this is the currentversion before use.

Skin. The exterior skin of the moulage shall be resilient vinyl plastisol or equal and shall retain its finish without chipping. It shall be able to stand up under repeated use in first-aid training procedures. The moulage skin shall have a minimum skin thickness at any point of 1/16 inch (1.58 mm).

Material. The material used directly beneath the skin shall be foam rubber or foam plastic. The material used shall be compatible with the material used for the skin. The colostomy moulage shall be mounted on a rigid plastic base made of either vacuum formed impact resistant plastic or fiberglass reinforced plastic. The plastic water bag shall be vinyl. The Government shall furnish an information sheet to be permanently affixed on the interior of the lid and covered with clear acetate or a clear plastic covering.

All metal parts or fastenings used in the moulages or components shall be corrosion-resistant. Brass, anodized aluminum, and 300 series stainless steel are acceptable materials.

The moulage shall be finished with a matte coating which shall best simulate human skin and which shall form a durable bond with the material used in making the moulage or the raw material shall be vinyl plastisol which when cured shall simulate human skin in color and texture. The finish shall be durable and shall not chip, soften, crack, or become sticky when used as intended.

All colors used shall be those shades which best simulate the particular flesh and tissue of the anatomy whether by coating or additive to raw material.

Carrying Case The carrying case shall be made of either vacuum formed impact resistant plastic or fiberglas reinforced plastic. Each case shall have a 1 inch (2.54 cm) high by 3 inch (7.62 cm) long by 0.012 inch (0.304 mm) thick aluminum nameplate. Nameplates shall be black lettering on an aluminum finish using the metal photo process or equal and shall be affixed by rivets to the center top of the exterior case. The nameplate shall read:

U.S. GOVERNMENT PROPERTY
TRAINING AID, COLOSTOMY MOULAGE KIT
DEVICE 08-10

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the currentversion before use.

Instruction sheet. One "Instruction Sheet" shall be furnished by the Government. The instruction sheet shall be made by the contractor and shall be 8-1/2 by 11 inches, with black printing on white paper, printed front to back, and placed between two (2) sheets of clear plastic or be covered with clear acetate. One instruction sheet shall be retained in each moulage case, but shall not be permanently affixed to the case in any way. The instruction sheet shall be as follows:

DRAFT INSTRUCTION SHEET

1. GENERAL DESCRIPTION

a. The kit contains:

(1). One "single-barreled" uncomplicated colostomy moulage.

(2). One "single-barreled" colostomy moulage with necrotic skin edges.

(3). One "double-barreled" colostomy with required skin approximation and a sutured laparotomy incision.

b. Each moulage comes complete with base, flexible tubing and reservior. There is a sheet of sponge rubber placed between the plastic skin and the base.

2. ASSEMBLY AND UTILIZATION

a. Fasten the particular moulage to be used to the base.

b. Blow into the plastic reservior to make sure that the sides are not sticking together.

c. Fasten the moulage to the appropriate area of the DCV 8-1 Nursing Technique Doll or of the human subject, by the straps provided.

d. Conceal the edges of the moulage with sheets or clothing, for more realism.

e. Fasten the plastic reservoir by slipping the strap at the lower right side of the moulage through the loop at the top of the reservior, and connect the tubing to the reservoir. The device is now ready for use.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the currentversion before use.

CAUTION: Use only clear water containing no drugs or solutions, as the irrigating fluid.

f. After the fluid has been introduced, the nursing doll or human subject can be turned on the left side, thus elevating the reservoir to provide gravity backflow of the instilled fluid. Gravity back-flow of the fluid can be induced also by placing the subject in a sitting position and elevating the reservoir above the colostomy.

3. PREVENTIVE MAINTENANCE

CAUTION: Maintenance procedures given below must be observed to avoid damage to the moulages.

a. Do not fold for storing, Make sure the sheet of sponge rubber is completely dry before storing.

b. Wash the plastic skin with a mild soap and warm water; do not rub, scrub, or use abrasive cleaners.

c. Drain reservoirs and moulage capillary lines immediately after using. Do not use boiling water under any circumstances.

Workmanship. The moulage kit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One moulage kit, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6910-01-234-2400

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the currentversion before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6910-0008

PGC 60778

Location: CID/MOULAGE/H14-19

☆U.S. GOVERNMENT PRINTING OFFICE: 1988 - 505-033/92065

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the current version before use.