A-A-53091
31 August 1987

COMMERCIAL ITEM DESCRIPTION

CHAIR, WHIRLPOOL BATH

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a mobile, adjustable high chair for use during hydrotherapy of foot and lower leg in whirlpool baths.

Salient characteristics:

The whirlpool bath chair shall be a mobile, adjustable high chair with heavy tubular aluminum or chrome-plated steel frame construction with four legs suitable for use with leg tank whirlpool baths. Each leg shall have a suitable circular bumper.

The chair shall have four institutional type, quiet, smooth rolling, ball bearing, swivel casters, two of which shall have foot-locking brakes. The minimum outside diameter of the casters shall be three inches.

The chair shall have a suitable seat with back and arm supports. The seat shall be adjustable thru a nominal height of 33 inches to 38 inches. The seat shall be furnished with a sturdy, long-wearing pad. A strong, long-wearing patient security belt shall be attached to the back support.

The chair shall be designed to permit the seat to protrude over the edge of the whirlpool tank to afford patient comfort. The chair shall have a full-width step located at a nominal 13 inches height. The step shall be furnished with a slip-resistant surface.

All welds shall be complete with no breaks or cracks. The welds shall be ground smooth to blend with adjacent surfaces, and shall be clean with no pitting, spatter, or discoloration.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:15Z
Check the source to verify that this is the current version before use.

The chair shall be so constructed that all corners shall be rounded and all edges smooth to prevent injury to patients.

The chair finish shall be free of chips, cracks, scratches, and other appearance defects.

Workmanship. The whirlpool bath chair shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One whirlpool bath chair, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

First article inspection. First article inspection shall be performed on the whirlpool bath chair when a first sample is required. This inspection shall include the examination.

Quality conformance inspection. Quality conformance inspection shall include the examination specified herein.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the current version before use.

Sampling.

For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one whirlpool bath chair as specified in this description

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and minor combined)	II	4.0
For dimensional defects	S-1	2.5
For all tests	S-2	1.5

Examination. The whirlpool bath chair shall be examined to determine compliance with the requirements of this specification in accordance with the classification of defects in table II.

TABLE II. Classification of Defects*

Classification	Defects
Major	
101	Chair does not have a heavy tubular aluminum or chrome-plated steel frame construction.
102	Chair does not have suitable circular leg bumpers.
103	Chair does not have three inch minimum diameter casters.
104	Chair does not have foot-locking brakes on at least two of its casters.
105	Seat does not adjust through a nominal height of 33 inches to 38 inches.
106	Chair does not have a strong, long-wearing patient security belt attached to the back support.
107	Chair does not have a full width step.
108	Step not furnished with a slip-resistant surface.
109	Seat not furnished with a sturdy, long-wearing pad.
110	Chair not free of burrs or sharpness.
Minor	
201	Chair finish not free of chips, cracks, scratches, and other appearance defects.
202	Any component parts of chair not secured in place.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

*Examination shall not be limited to the classified possible defects listed herein.

Dimensional Examination. The whirlpool bath chair shall be examined for defects in dimensions. Any dimension not within the specifications herein shall be classified as a defect.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-00-875-3714

NOTE: For use with whirlpool baths, NSN 6530-00-875-3646 and NSN 6530-00-710-4005.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1923

PGC 50458

Location: CID/CHAIR/H13-23

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:15Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53091	Chair, Whirlpool Bath

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:15Z
Check the source to verify that this is the current version before use.